

Australian Government

Department of Industry, Innovation and Science National Measurement Institute

Proficiency Test Report AQA 19-01 Methamphetamine

July 2019

ACKNOWLEDGMENTS

This study was conducted by the National Measurement Institute (NMI). Support funding was provided by the Australian Government Department of Industry, Innovation and Science.

I would like to thank the management and staff of the participating laboratories for supporting the study. It is only through widespread participation that we can provide an effective service to laboratories.

The assistance of the following NMI staff members in the planning, conduct and reporting of the study is acknowledged.

Lesley Johnston

Luminita Antin

Raluca Iavetz

A/g Manager, Chemical Reference Values 105 Delhi Rd, Riverside Corporate Park, North Ryde NSW 2113

Phone: 61-2-9449 0178

raluca.iavetz@measurement.gov.au



ATA Accredited for compliance with ISO/IEC 17043

| ТАВ | LE | OF | CO | NTE | ENTS |
|-----|----|----|----|-----|------|
| | | | | | |

| SUMN | <i>I</i> ARY | 1 |
|------|---|----|
| 1.1 | NMI Proficiency Testing Program | 2 |
| 1.2 | Study Aims | 2 |
| 1.3 | Study Conduct | 2 |
| 2 S | STUDY INFORMATION | 3 |
| 2.1 | Study Timetable | 3 |
| 2.2 | Participation | 3 |
| 2.3 | Test Material Specification | 3 |
| 2.4 | Laboratory Code | 3 |
| 2.5 | Test Sample Homogeneity | 3 |
| 2.6 | Sample Dispatch and Receipt | 4 |
| 2.7 | Instructions to Participants | 4 |
| 2.8 | Interim Report | 4 |
| 3 P | PARTICIPANT LABORATORY INFORMATION | 5 |
| 3.1 | Test Method Summaries | 5 |
| 3.2 | Reported Basis of Participants' Measurement Uncertainty Estimates | 6 |
| 3.3 | Details of Participant Calibration Standard | 8 |
| 3.4 | Participants' Comments | 9 |
| 4 P | PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS | 10 |
| 4.1 | Results Summary | 10 |
| 4.2 | Assigned Value | 10 |
| 4.3 | Performance Coefficient of Variation (PCV) | 10 |
| 4.4 | Target Standard Deviation | 11 |
| 4.5 | z-Score | 11 |
| 4.6 | En-Score | 11 |
| 4.7 | Traceability and Measurement Uncertainty | 11 |
| 5 T | ABLES AND FIGURES | 12 |
| 6 D | DISCUSSION OF RESULTS | 20 |
| 6.1 | Assigned Value | 20 |
| 6.2 | Measurement Uncertainty Reported by Participants | 20 |
| 6.3 | z-Score | 21 |
| 6.4 | Z-Score scatter plot | 22 |
| 6.5 | En-Score | 22 |
| 6.6 | Identification of Cutting Agent | 23 |
| 6.7 | Participants' Analytical Methods | 23 |
| 6.8 | Summary of Participation and Performance in Methamphetamine Studies | 26 |
| 7 R | REFERENCES | 27 |
| APPE | NDIX 1 - MEASUREMENT UNCERTAINTY OF THE ROBUST AVERAGE | 28 |
| APPE | NDIX 2 – REFERENCE VALUE AND HOMOGENEITY TESTING | 29 |
| APPE | NDIX 3 - ACRONYMS AND ABBREVIATIONS | 31 |

SUMMARY

AQA 19-01 was conducted in March 2019. Three test samples of methamphetamine hydrochloride were sent to twenty-four laboratories. Two laboratories requested two sets of the test samples. Twenty-six sets of results were submitted by the due date.

Test samples were prepared at the NMI laboratory in Sydney using methamphetamine hydrochloride approximately 80% base (m/m) supplied by the Australian Federal Police (AFP).

The assigned values for Samples S1 and S2 were the reference values determined by quantitative nuclear magnetic resonance spectrometry (QNMR) with maleic acid (NMI certified reference material QNMR010) as internal standard.

Traceability: The reference values are traceable to the SI through Australian Standards for mass via balance calibration certificates and the purity of the NMI maleic acid certified reference material QNMR010.2018.01.

The assigned value for Sample S3 was the robust average of participants' results.

Traceability: The consensus of participants' results is not traceable to any external reference, so although expressed in SI units, metrological traceability has not been established.

The outcomes of the study were assessed against the aims as follows:

• assess the proficiency of laboratories measuring methamphetamine in samples typical of a routine seizure;

Laboratory performance was assessed by z-score and E_n-score.

Laboratories 3, 4, 6, 7, 8, 9, 11 (only two results submitted), 13, 14, 15, 17, 18, 21, 22, 23, 24, 25 and 26 returned satisfactory z and E_n -scores for all results.

Laboratory **10** returned unsatisfactory z-scores and E_n-scores for all samples.

Of the 77 results for which z-scores were calculated, 70 (91%) returned $|z| \le 2$ indicating a satisfactory performance.

Of the 77 results for which $|E_n|$ -scores were calculated, 68 (88%) returned $|E_n| \le 1$ indicating agreement of the participants' results with the assigned value within their respective expanded uncertainties.

• *develop a practical application of traceability and measurement uncertainty and provide participants with information that will assist uncertainty estimates; and*

71 results (92%) were reported with an associated expanded uncertainty. Laboratories **16** and **22** did not report an uncertainty. These laboratories were not accredited.

Laboratory **12** reported an identical uncertainty for samples which were of significantly different concentrations.

The magnitude of reported uncertainties was within the range 1% to 27% relative.

• *test the ability of participants to identify a cutting agent commonly found in controlled drug preparation*

Caffeine was used to prepare Samples S1 and S2. Twenty-five laboratories reported on the diluent and correctly identified it.

Dimethylsulfone (DMS or MMS) was used to prepare Sample S3. Eighteen laboratories reported the diluent and correctly identified it.

INTRODUCTION

1.1 NMI Proficiency Testing Program

The National Measurement Institute (NMI) is responsible for Australia's national measurement infrastructure, providing a range of services including a chemical proficiency testing program.

Proficiency testing (PT) is: 'evaluation of participant performance against pre-established criteria by means of interlaboratory comparison.'¹ NMI PT studies target chemical testing in areas of high public significance such as trade, environment, law enforcement and food safety. NMI offers studies in:

- pesticide residues in fruit and vegetables, soil and water;
- petroleum hydrocarbons in soil and water;
- PFAS in water, soil and biota;
- metals in soil, water, food and pharmaceuticals;
- controlled drug assay and clandestine laboratory;
- allergens in food; and
- folic acid in flour.

1.2 Study Aims

The aims of the study were to:

- assess the proficiency of laboratories measuring methamphetamine in samples typical of a routine seizure;
- develop a practical application of traceability and measurement uncertainty and provide participants with information that will assist uncertainty estimates; and
- test the ability of participants to identify a cutting agent commonly found in controlled drug preparation.

The choice of the test method was left to the participating laboratories.

1.3 Study Conduct

NMI is accredited by the National Association of Testing Authorities, Australia (NATA) to ISO 17043¹ as a provider of proficiency testing schemes. This controlled drug proficiency test is within the scope of NMI's accreditation.

The conduct of NMI proficiency tests is described in the NMI Chemical Proficiency Testing Study Protocol.² The statistical methods used are described in the NMI Chemical Proficiency Testing Statistical Manual.³ These documents have been prepared with reference to ISO 17043 and The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories.⁴

2 STUDY INFORMATION

2.1 Study Timetable

The timetable of the study was:

| Invitation issued: | 8 January 2019 |
|------------------------|----------------|
| Samples dispatched: | 25 March 2019 |
| Results due: | 17 June 2019 |
| Interim report issued: | 17 June 2019 |

The results due date was changed from 3 May 2019 to 17 June 2019 due to delays in dispatch to overseas laboratories.

2.2 Participation

A total of ninety-nine international, national, state government and private laboratories were invited to participate.

Twenty-four laboratories agreed to participate and submitted results. Two laboratories requested two sets of samples to be analysed independently by two different analysts.

2.3 Test Material Specification

Three test samples were prepared in February 2019. The starting material for samples S1, S2 and S3 was methamphetamine hydrochloride approximately 80% base (m/m), supplied by the Australian Federal Police. Caffeine and dimethylsulfone, purchased from Sigma Aldrich, were used as cutting agents. Sample S1 and S2 were blind duplicates, cut with caffeine, and Sample S3 was cut with dimethylsulfone.

The methamphetamine was ground and sieved through a 180 μ m sieve. The cutting agents were processed similarly to the methamphetamine powder.

Test samples were then prepared by mixing a known mass of sieved drug material with a known mass of sieved cutting agent in a tumbler overnight.

Portions of 150 mg of each of the test samples were weighed into labelled glass vials.

Sample S1 was prepared to contain 56.9% methamphetamine base (m/m).

Sample S2 (duplicate) was identical with Sample S1.

Sample S3 was prepared to contain 24.9% methamphetamine base (m/m).

2.4 Laboratory Code

Each participant was randomly assigned a confidential laboratory code.

2.5 Test Sample Homogeneity

The preparation of homogeneous test samples is an important part of a proficiency testing study. Given the small (<150 mg) test portions normally used for controlled substances analysis the particle size must be sufficiently small and uniformly distributed to ensure minimal influence on analytical precision.

The homogeneity testing of Samples S1 and S2 is described in Appendix 2. Samples were demonstrated to be sufficiently homogeneous for the purpose of this PT study.

Sample S3 was prepared using the same procedure and the results returned by participants gave no reason to question the homogeneity of the test samples.

2.6 Sample Dispatch and Receipt

A set of three test samples, each containing approximately 150 mg of test material, were dispatched on 25 March 2019.

The following items were packaged with the samples:

- a covering letter with instructions for participants; and
- a form for participants to confirm the receipt of the test samples.

An Excel spreadsheet for the electronic reporting of results was e-mailed to participants.

2.7 Instructions to Participants

Participants were asked to analyse the samples using their routine quantitative method and return the following information:

- one result for each sample as % m/m methamphetamine base;
- an estimate of the expanded uncertainty associated with the result as % m/m methamphetamine base at the 95% confidence level;
- brief detail on how the uncertainty was calculated e.g. uncertainty budget method;
- the identity of the cutting agents in all three samples, if part of routine analysis;
- origin and stated purity of the analytical reference standard used;
- brief summary of the quantitative method used;
- the completed results sheet by 3 May 2019. This date was later modified to 17 June 2019 due to delays in delivery of samples overseas; and
- any other comments.

2.8 Interim Report

An interim report was emailed to all participants on 17 June 2019.

3 PARTICIPANT LABORATORY INFORMATION

3.1 Test Method Summaries

Reported participant method summary is presented as received for information in Table 1.

| Lab | Extraction | Internal | Calib | | | |
|------|--|---------------------------|--------|-----------|--|---|
| Code | solvent | standard | points | Technique | Detector | Column |
| 1 | ACN/MeOH/H2O | Analog of methamphetamine | 7 | UPLC | MSMS | C-18 coloumn |
| 2 | Methanol | N/A | 6 | HPLC | UV 258nm | Phenomenex Luna C18, 0.5% DEA pH 8.5:CH3OH 40:60 |
| 3 | Methanol:KOH buffer (50:50) | Methoxyphenamine | 3 | UPLC | PDA | Acquity UPLC BEH C18 |
| 4 | Methanol | | 5 | HPLC | DAD | C18 |
| 5 | Deuterium oxide | maleic acid | | 1H QNMR | Bruker AVIII 600 with BBFO probe | N/A |
| 6 | Methanol | Strychnine | 6 | UPLC | FID | Waters Acquity UPLC BEH Phenyl 1.8um, 2.1 mm x 100 mm |
| 7 | Methanol | NO | 1 | HPLC | DAD | Zorbax XDB-C18 (4,6x150 mm) |
| 8 | Isooctane + Ammonium Hydroxide | Dodecane | 3 | GC | FID, MS | HP-1MS |
| 9 | Methanol | Procaine | 4 | HPLC | DAD | HP5 |
| 10 | ethyl acetate | diphenylamine | 5 | GC | FID | HP1 |
| 11 | Dissolution in acetonitrile/water | Methoxyphenamine HCl | 3 | HPLC | DAD | Alltima C-18 |
| 12 | Water/Acetonitril e (80:20) | None | 2 | HPLC | UV/Vis | C18 |
| 13 | Purified Water | Phentermine | 1 | LC | DAD | Agilent Zorbax SB-C8 |
| 14 | methanol | propylparaben | 3 | UPLC | PDA | ACQUITY C-18 |
| 15 | Purified Water | Phentermine | 1 | LC | DAD | Agilent Zorbax SB-C8 |
| 16 | Methanol | Methamphet-amine D11 | 6 | GC | MS | Rxi-5Sil MS 30mx0.25mmlDx0.25um |
| 17 | Water | None | 4 | HPLC | PDA | C18 |
| 18 | MilliQ water | None | 6 | UPLC | UV | Acquity UPLC® BEH C18 1.7 μm 2.1 x 100 mm |
| 19 | water | None | 5 | HPLC | UV DAD | Silica |
| 20 | Acetonitrile, ammonium acetate, diethylamine & water | None | 3 | HPLC | Diode Array | LiChrospher RP-18 (5µm) |
| 21 | Water | - | 4 | HPLC | DAD | Zorbax RX-sil |
| 22 | Ethanol | Propylparaben | 7 | UPLC | DAD | BEH Shield RP18 |
| 23 | Methanol | None | 5 | HPLC | DAD | Phenomenex C-18-XB |
| 24 | D2O | Maleic acid | | QNMR | | NA |
| 25 | Water | N/A | 6 | LC | DAD | UPLC BEH C18 1.7uM |
| 26 | acetic acid/ACN/water (4/20/76 V/V/V) | No ISTD | 5 | HPLC | UV DAD | POROSHELL 120 EC-C18 |

 Table 1 Participant Summary of Test Methods

3.2 Reported Basis of Participants' Measurement Uncertainty Estimates

Participant returns are listed in Table 2.

| | | | 5 | |
|--------------------------------|---|--|---|-------------------------------------|
| Lab. Approach to Estimating MU | | Information S | Guide Document | |
| Coue | | Precision | Method Bias | for Estimating WO |
| 1 | Top Down - precision and estimates of the method and laboratory bias | | | |
| 2 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Reference Material Duplicate Analysis | | Eurachem/CITAC Guide |
| 3 | Top Down - reproducibility (standard deviation) from PT studies used directly | Control Samples – Reference Material Duplicate Analysis | Standard Purity Masses and volumes | Eurachem/CITAC Guide |
| 4 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | Control Samples – Reference Material | Laboratory bias from PT studies Instrument Calibration Masses and volumes | ISO/GUM |
| 5 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | | | |
| 6 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | Control Samples – Duplicate Analysis | Laboratory bias from PT studies Standard Purity Instrument Calibration Masses and volumes Homogeneity of sample | Nata Technical Note 33 |
| 7 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Certified Reference Material Duplicate Analysis | Laboratory bias from PT studies Standard Purity Instrument Calibration | Nordtest Report TR537 |
| 8 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | Control Samples – Certified Reference Material Duplicate Analysis | Recoveries of SS Standard Purity Matrix Effects Instrument Calibration Masses and volumes Homogeneity of sample | Eurachem/CITAC Guide |
| 9 | Top Down - precision and estimates of the method and laboratory bias | Control Samples | Laboratory bias from PT studies Recoveries of SS Standard Purity Matrix Effects Instrument Calibration Masses and volumes Homogeneity of sample | EA-4/16: 2003 and ILAG G-17:2002 |
| 10 | Standard deviation of replicate analyses multiplied by 2 or 3 | Control Samples – reference material | Standard Purity | |
| 11 | Bottom up (ISO/GUM, Fish bone/ Cause and effect diagram) | Duplicate Analysis | Recoveries of SS Standard Purity Matrix Effects Instrument Calibration Masses and volumes | ISO/GUM |
| 12 | Standard deviation of replicate analyses multiplied by 2 or 3 | Control Samples – Reference Material Duplicate Analysis | Standard Purity Masses and volumes | Eurachem/CITAC Guide |
| 13 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Spiked sample Duplicate Analysis | Laboratory bias from PT studies | Eurachem/CITAC Guide |

Table 2 Reported Basis of Uncertainty Estimate

| Lab. | Approach to Estimating MU | Information S | Guide Document for Estimating MU | |
|------|--|--|--|--------------------------------|
| Couc | | Precision | Method Bias | for Estimating WO |
| 14 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Certified Reference Material Duplicate Analysis | Standard Purity Instrument Calibration Masses and volumes Homogeneity of sample | Nata Technical Note 33 |
| 15 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Spiked sample Duplicate Analysis | Laboratory bias from PT studies | Eurachem/CITAC Guide |
| 16 | Standard deviation of replicate analyses multiplied by 2 or 3 | Duplicate Analysis | Masses and volumes | ISO/GUM |
| 17 | | | | |
| 18 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Reference Material Duplicate Analysis | Instrument Calibration Masses and volumes Homogeneity of sample | Nata Technical Note 33 |
| 19 | repeatability, sample heterogeneity (ENFSI quantitative sampling guideline) | Control Samples – Reference Material Duplicate Analysis | Homogeneity of sample | Eurachem/CITAC Guide |
| 20 | Uncertainty Budget Method | Control Samples – Reference Material Duplicate Analysis | Standard Purity Instrument Calibration Masses and volumes | Internal document |
| 21 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Sample from police case. Duplicate Analysis | Laboratory bias from PT studies | Nordtest Report TR537 |
| 22 | | | | |
| 23 | | | | |
| 24 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – previously analysed real seizure samples Duplicate Analysis | Matrix Effects Instrument Calibration Masses and volumes Homogeneity of sample | Eurachem/CITAC Guide |
| 25 | Top Down - precision and estimates of the method and laboratory bias | Control Samples – Reference Material Duplicate Analysis | Standard Purity Matrix Effects Instrument Calibration Masses and volumes Homogeneity of sample | Nata Technical Note 33 |
| 26 | Accuracy Profile - based on intermediate precision and repeatability | Reference Material | Standard Purity | ISO 5725-2 and ISO/TS 21748 |

3.3 Details of Participant Calibration Standard

Participant returns as received are listed in Table 3.

| Lab. Code | Reference Standard* | Purity (%) |
|--------------|-------------------------------|--------------------|
| 1 | Sigma Aldrich | 100 |
| 2 | NMI | 99.8 |
| 3 | NMI | 99.8 |
| 4 | NMI | 99.8 |
| 5 | Sigma Aldrich Prod. no. 92816 | 99.98 ± 0.13 |
| 6 | NMI | 99.8 |
| 7 | Lipomed | 99.47 |
| 8 | NMI | 99.8 +/- 1.9 |
| 9 | Lipomed | 99.467 +/- 0.015 |
| 10 | LGC | 99.8 |
| 11 | NMI | 99.8 |
| 12 | Sigma Aldrich | >98 |
| 13 | Lipomed | 99.940 ± 0.006 |
| 14 | NMI | 99.8 |
| 15 | Lipomed | 99.940 ± 0.006 |
| 16 | Lipomed | 99.5788 |
| 17 | In house synthesis | 100.6 |
| 18 | NMI | 99.8 |
| 19 | Lipomed | 99,5 |
| 20 | NMI | 99.8 |
| 21 | Sigma | 100 |
| 22 | Euromedex | 99.5 |
| 23 | Sigma | 99.9 |
| 24 | - | - |
| 25 | NMI | 99.8 |
| 26 | LIPOMED | 99.47 |

| Table 3 | Participant | Calibration | Standard |
|---------|-------------|-------------|----------|
|---------|-------------|-------------|----------|

*Some data has been edited to preserve confidentiality

3.4 Participants' Comments

The study manager welcomes comments or suggestions from participants as it provides information which will improve future studies. All returns are listed as received in Table 4 along with the study manager's response, where appropriate.

| Lab. Code | Participant comments | Study co-ordinator response |
|--------------|---|---|
| 11 | Local illicit "ICE" seizures usually contain very high purity of methamphetamine hydrochloride (over 90% w/w). The determined purity of the test sample in this study is below those of our daily encountered samples. | The study coordinator provides a range of samples of different purity. In this study all samples were cut, while in the previous study Sample S1 was uncut (high purity). |
| 16 | It would appear the samples for this study have been delayed in transit - perhaps at customs. Would perhaps be useful to have a larger time window for completion and submission of test results | The turn-around-time (TAT) for Controlled drugs PT is set to 6 weeks from dispatch of samples which is two weeks longer than all the other PT programs. Even with a longer TAT, the study coordinator acknowledges extra delays in transit for some countries and that is why sometimes the TAT is modified after the start of the program. |
| 18 | I believe this is a fair and true study. | Thank you. |

| Table 4 Participant (| Comments |
|-----------------------|----------|
|-----------------------|----------|

4 PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS

4.1 Results Summary

Participant results are listed in Tables 5 to 7 with resultant summary statistics: mean, median, maximum, minimum, robust average, robust standard deviation (Robust SD) and robust coefficient of variation (Robust CV).



Figure 1 Guide to Presentation of Results

4.2 Assigned Value

Assigned value is defined as: 'the value attributed to a particular quantity and accepted, sometimes by convention, as having an uncertainty appropriate for a given purpose'.⁴

For a proficiency test, the assigned value is the best available measurement of the true concentration of an analyte in the test sample.

4.3 Performance Coefficient of Variation (PCV)

The performance coefficient of variation (PCV) is a measure of the between laboratory variation that in the judgement of the study organiser would be expected from participants given the sample concentration. It is important to note that this is a performance measure set by the study coordinator; it is not the coefficient of variation of participant results.

4.4 Target Standard Deviation

The target standard deviation (σ) is the product of the assigned value (*X*) and the performance laboratory coefficient of variation (PCV) as presented in Equation 1. This value is used for calculation of participant z-score.

$$\sigma = X * PCV$$
 Equation 1

4.5 z-Score

For each participant result a z-score is calculated according to Equation 2 below:

$$z = \frac{(\chi - X)}{\sigma} \qquad Equation 2$$

where:

z is z-score

- χ is participants' result
- X is the study assigned value
- σ is the target standard deviation from Equation 1

A z-score with absolute value (|z|):

- $|z| \le 2$ is satisfactory;
- 2 < |z| < 3 is questionable;
- $|z| \ge 3$ is unsatisfactory.

4.6 E_n-Score

The E_n -score is complementary to the z-score in assessment of laboratory performance. E_n -score includes measurement uncertainty and is calculated according to Equation 3 below:

$$E_n = \frac{(\chi - X)}{\sqrt{U_{\chi}^2 + U_X^2}} \qquad Equation 3$$

where:

 E_n is E_n-score

- χ is a participants' result
- X is the assigned value
- U_{χ} is the expanded uncertainty of the participants' result
- U_x is the expanded uncertainty of the assigned value

An E_n -score with absolute value ($|E_n|$):

- $|E_n| \le 1$ is satisfactory;
- $|E_n| > 1$ is unsatisfactory.

4.7 Traceability and Measurement Uncertainty

Laboratories accredited to ISO/IEC Standard 17025:2017⁵ must establish and demonstrate the traceability and measurement uncertainty associated with their test results. Guidelines for quantifying uncertainty in analytical measurement are described in the Eurachem /CITAC Guide.⁶

5 TABLES AND FIGURES

Table 5

Sample Details

| Sample No. | S1 |
|------------|-----------------|
| Matrix. | Powder |
| Analyte. | Methamphetamine |
| Units | % base (m/m) |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E _n -Score |
|-----------------|--------|-------------|---------|-----------------------|
| 1 | 57 | 8.6 | -0.40 | -0.08 |
| 2 | 56.7 | 4.8 | -0.58 | -0.20 |
| 3 | 58.4 | 3.3 | 0.40 | 0.20 |
| 4 | 56.8 | 1.8 | -0.52 | -0.42 |
| 5 | 57.4 | 0.6 | -0.17 | -0.22 |
| 6 | 56.6 | 3.5 | -0.64 | -0.30 |
| 7 | 58.6 | 2.93 | 0.52 | 0.28 |
| 8 | 57.3 | 2.3 | -0.23 | -0.15 |
| 9 | 57.7 | 2.9 | 0.00 | 0.00 |
| 10 | 49.9 | 3.2 | -4.51 | -2.28 |
| 11 | 57.0 | 3.6 | -0.40 | -0.18 |
| 12 | 58.6 | 5.519 | 0.52 | 0.16 |
| 13 | 57.1 | 3.5 | -0.35 | -0.16 |
| 14 | 57.5 | 1.9 | -0.12 | -0.09 |
| 15 | 57.7 | 3.5 | 0.00 | 0.00 |
| 16 | 60 | NR | 1.33 | 1.92 |
| 17 | 58.8 | 2.0 | 0.64 | 0.47 |
| 18 | 55 | 5.5 | -1.56 | -0.48 |
| 19 | 57.1 | 5.7 | -0.35 | -0.10 |
| 20 | 55.2 | 2.0 | -1.44 | -1.07 |
| 21 | 56.9 | 2.8 | -0.46 | -0.26 |
| 22 | 58.7 | NR | 0.58 | 0.83 |
| 23 | 58 | 6.4 | 0.17 | 0.05 |
| 24 | 56.9 | 1.88 | -0.46 | -0.36 |
| 25 | 56 | 5.6 | -0.98 | -0.30 |
| 26 | 59.7 | 2.7 | 1.16 | 0.68 |
| Statistics | | | | |
| Assigned Value | 57.7 | 1.2 | | |
| Reference Value | 57.7 | 1.2 | | |

| Assigned Value | 57.7 | 1.2 |
|-----------------|------|-----|
| Reference Value | 57.7 | 1.2 |
| Robust Average | 57.4 | 0.6 |
| Median | 57.2 | 0.3 |
| Mean | 57.2 | |
| Ν | 26 | |
| Max. | 60 | |
| Min. | 49.9 | |
| Robust SD | 1.2 | |
| Robust CV | 2.1% | |



z-Scores: S1 - Methamphetamine





En-Scores: S1 - Methamphetamine

Figure 2

| Т | ab | le | 6 |
|---|----|----|---|
| | | | |

| Sample Details | | | | |
|---------------------|--------------|-------------|---------|-----------------------|
| Sample No. | S2 | | | |
| Matrix. | Powder | | | |
| Analyte. | Methamphetan | nine | | |
| Units | % base (m/m) | | | |
| Participant Results | | | | |
| Lab Code | Result | Uncertainty | z-Score | E _n -Score |
| 1 | 54 | 8.1 | -2.14 | -0.45 |
| 2 | 56.8 | 4.8 | -0.52 | -0.18 |
| 3 | 58.9 | 3.2 | 0.69 | 0.35 |
| 4 | 56.8 | 1.9 | -0.52 | -0.40 |
| 5 | 57.2 | 0.6 | -0.29 | -0.37 |
| 6 | 56.7 | 3.5 | -0.58 | -0.27 |
| 7 | 56.9 | 2.84 | -0.46 | -0.26 |
| 8 | 57.8 | 2.4 | 0.06 | 0.04 |
| 9 | 58.5 | 2.9 | 0.46 | 0.25 |
| 10 | 51.2 | 3.3 | -3.76 | -1.85 |
| 11 | 56.5 | 3.5 | -0.69 | -0.32 |
| 12 | 58.5 | 5.519 | 0.46 | 0.14 |
| 13 | 57.6 | 3.5 | -0.06 | -0.03 |
| 14 | 57.0 | 1.9 | -0.40 | -0.31 |
| 15 | 58.0 | 3.5 | 0.17 | 0.08 |
| 16 | 59 | NR | 0.75 | 1.08 |
| 17 | 59.1 | 2.0 | 0.81 | 0.60 |
| 18 | 58 | 5.8 | 0.17 | 0.05 |
| 19 | 57.3 | 5.7 | -0.23 | -0.07 |
| 20 | 58.2 | 2.1 | 0.29 | 0.21 |
| 21 | 59.0 | 3.0 | 0.75 | 0.40 |
| 22 | 58.4 | NR | 0.40 | 0.58 |
| 23 | 59 | 6.5 | 0.75 | 0.20 |
| 24 | 56.5 | 1.86 | -0.69 | -0.54 |
| 25 | 56 | 5.6 | -0.98 | -0.30 |
| 26 | 59.8 | 2.7 | 1.21 | 0.71 |
| Statistics | 1 | | | |
| Assigned Value | 57.7 | 1.2 | | |
| Reference Value | 57.7 | 1.2 | | |
| Robust Average | 57.6 | 0.6 | | |
| Median | 57.7 | 0.5 | | |
| Mean | 57.4 | | | |
| Ν | 26 | | | |
| Max. | 59.8 | | | |
| Min. | 51.2 | | | |
| Robust SD | 1.3 | | | |
| Robust CV | 2.3% | | | |



z-Scores: S2 - Methamphetamine





En-Scores: S2 - Methamphetamine



| - | | | _ |
|----|----|---|---|
| Ia | bl | е | 1 |

| Sample Details | | | | |
|---------------------|--------------|-------------|---------|-----------------------|
| Sample No. | S3 | | | |
| Matrix. | Powder | | | |
| Analyte. | Methamphetar | nine | | |
| Units | % base (m/m) | | | |
| Participant Results | | | | |
| Lab Code | Result | Uncertainty | z-Score | E _n -Score |
| 1 | 24 | 3.6 | 1.15 | 0.22 |
| 2 | 25.5 | 2.1 | 3.30 | 1.07 |
| 3 | 22.5 | 1.2 | -1.01 | -0.54 |
| 4 | 22.9 | 0.7 | -0.43 | -0.35 |
| 5 | 24.0 | 0.3 | 1.15 | 1.37 |
| 6 | 23.7 | 1.5 | 0.72 | 0.32 |
| 7 | 23.0 | 1.15 | -0.29 | -0.16 |
| 8 | 22.9 | 1.0 | -0.43 | -0.27 |
| 9 | 22.1 | 1.1 | -1.58 | -0.91 |
| 10 | 19.9 | 1.3 | -4.74 | -2.37 |
| 11 | NR | NR | | |
| 12 | 20.7 | 5.519 | -3.59 | -0.45 |
| 13 | 23.5 | 1.5 | 0.43 | 0.19 |
| 14 | 23.2 | 1.1 | 0.00 | 0.00 |
| 15 | 22.7 | 1.4 | -0.72 | -0.34 |
| 16 | 24 | NR | 1.15 | 1.60 |
| 17 | 22.3 | 0.8 | -1.29 | -0.95 |
| 18 | 24 | 2.4 | 1.15 | 0.33 |
| 19 | 25.6 | 2.6 | 3.45 | 0.91 |
| 20 | 24.2 | 0.9 | 1.44 | 0.97 |
| 21 | 23.6 | 1.2 | 0.57 | 0.31 |
| 22 | 23.6 | NR | 0.57 | 0.80 |
| 23 | 22 | 2.4 | -1.72 | -0.49 |
| 24 | 23.4 | 0.77 | 0.29 | 0.22 |
| 25 | 22 | 2.2 | -1.72 | -0.53 |
| 26 | 24.2 | 3.0 | 1.44 | 0.33 |
| Statistics | 1 | | | |
| Assigned Value | 23.2 | 0.5 | | |
| Robust Average | 23.2 | 0.5 | | |
| Median | 23.4 | 0.4 | | |
| Mean | 23.2 | | | |
| N | 25 | | | |
| Max. | 25.6 | | | |
| Min. | 19.9 | | | |
| Robust SD | 1.1 | | | |
| Robust CV | 4.7% | | | |



z-Scores: S3 - Methamphetamine





En-Scores: S3 - Methamphetamine

Figure 4



Duplicate Results S1 and S2 Methamphetamine

Figure 5 Results for duplicate samples S1 and S2 Horizontal lines (solid) are the upper and lower 95% confidence interval of the assigned value.

| Lab | Cutting agents | | |
|------|----------------------|----------------------|---|
| Code | S1 | S2 | S3 |
| 1 | Caffeine | Caffeine | |
| 2 | | | |
| 3 | Caffeine | Caffeine | Methylsulfonylmethane |
| 4 | Caffeine | Caffeine | Dimethylsulfone |
| 5 | Caffeine (28.3% m/m) | Caffeine (28.8% m/m) | Dimethylsulfone (70.1% m/m) |
| 6 | Caffeine | Caffeine | Dimethylsulfone |
| 7 | Caffeine | Caffeine | Dimethylsulfone |
| 8 | Caffeine | Caffeine | Dimethylsulfone |
| 9 | Caffeine | Caffeine | Dimethylsulfone |
| 10 | Caffeine 25.1% | Caffeine 25.5% | |
| 11 | Caffeine | Caffeine | |
| 12 | Caffeine | Caffeine | Dimethylsulfone (DMSO ₂) |
| 13 | Caffeine | Caffeine | Dimethylsulfone |
| 14 | caffeine | caffeine | |
| 15 | Caffeine | Caffeine | Dimethylsulfone |
| 16 | Caffeine | Caffeine | Dimethylsulfone |
| 17 | Caffeine | Caffeine | Dimethylsulfone |
| 18 | Caffeine | Caffeine | Dimethylsulfone |
| 19 | caffeine | caffeine | |
| 20 | Caffeine | Caffeine | Dimethylsulphone |
| 21 | Caffeine | Caffeine | |
| 22 | Caffeine | Caffeine | Dimethylsulfone |
| 23 | Caffeine | Caffeine | Methylsulfonylmethane |
| 24 | Caffeine | Caffeine | Dimethylsulfone |
| 25 | Caffeine | Caffeine | Dimethylsulfone |
| 26 | Caffeine | Caffeine | |

Table 8 Participants' identification of cutting agents

6 DISCUSSION OF RESULTS

6.1 Assigned Value

A reference value was obtained for identical Samples S1 and S2 using the quantitative nuclear magnetic resonance spectrometry (QNMR) measurement method described in Appendix 2. Maleic acid (NMI certified reference material QNMR010) was used as internal standard. The measured reference value was in agreement with the gravimetric preparation value and the robust average of participants' results, within their respective associated uncertainties. The uncertainty of the reference value was estimated in accordance with the ISO GUM⁷ by combining standard uncertainty terms for method precision, sample homogeneity, weighing of sample, preparation and addition of standard solution, the very small uncertainties in molecular weights and an estimate of potential bias made by comparing the results from different NMR signals.

The reference value was used as the assigned value for these samples.

Traceability: The measurements of the reference values for Samples S1 and S2 were made using QNMR and are traceable to the SI through Australian Standards for mass via balance calibration certificates and the purity of the NMI maleic acid certified reference material QNMR010.2018.01.

The assigned value for Sample S3 is the robust average of the results reported participants. The robust average and associated expanded uncertainties were calculated using the procedure described in 'ISO 13528:2015, Statistical methods for use in proficiency testing by interlaboratory comparisons'.⁸ The calculation procedure for the expanded uncertainty for Sample S3 is presented in Appendix 1.

Traceability: The consensus of participants' results is not traceable to any external reference, so although expressed in SI units, metrological traceability has not been established.

6.2 Measurement Uncertainty Reported by Participants

Participants were asked to report an estimate of the expanded measurement uncertainty associated with their results and the basis of this uncertainty estimate (Table 2).

It is a requirement of the ISO Standard 17025⁵ that laboratories have procedures to estimate the uncertainty of chemical measurements and to report this uncertainty in specific circumstances, including 'when the client's instruction so requires.' From 1 July 2012 this is also a requirement of ASCLD/Lab-International accreditation program.

71 results (92%) were reported with an associated expanded uncertainty. Laboratories **16** and **22** did not report an uncertainty. These laboratories were not accredited.

Laboratory **12** reported an identical uncertainty for samples which were of significantly different concentrations.

The magnitude of reported uncertainties was within the range 1% to 27% relative.

Sixty-one of 71 (86%) expanded uncertainties were between 3% and 10% relative to the result. Laboratories reporting uncertainties smaller than 3% or larger than 10% relative may wish to consider whether these estimates are realistic or fit for purpose.

Laboratories having a satisfactory z-score and an unsatisfactory E_n -score are likely to have underestimated the expanded uncertainty associated with the result.

In some cases the results were reported with an inappropriate number of significant figures. The recommended format is to write the uncertainty to no more than two significant figures and then to write the result with the corresponding number of decimal places (for example instead of $58.5 \pm 5.519\%$ the recommended format is $58.5 \pm 5.5\%$).⁶

6.3 z-Score

A target standard deviation equivalent to 3% CV was used to calculate z-scores. Target SDs, the between-laboratory coefficient of variation predicted by the Thomson - Horwitz equation⁹ and the between-laboratories coefficient of variation obtained in this study are presented in Table 9.

Table 9 Target standard deviations, coefficient of variations from predictive model and between laboratories

| Sample | Analyte | Assigned value (% base m/m) | Target SD (as PCV) | Thompson Horwitz CV | Between laboratories CV |
|------------|-----------------|--------------------------------------|--------------------------|---------------------------|-------------------------------|
| S1/S2 | Methamphetamine | 57.7 | 3% | 2.2% | 2.1%, 2.3% |
| S 3 | Methamphetamine | 23.2 | 3% | 2.5% | 4.7% |

A summary of z-scores by laboratory is presented in Figure 6.



70 of 77 numeric results (91%) returned a satisfactory z-score with $|z| \le 2$.

- Twenty-one participants (81%) 3, 4, 5, 6, 7, 8, 9, 11(only two results submitted), 13, 14, 15, 16, 17, 18, 20, 21, 22, 23, 24, 25 and 26 returned satisfactory scores for all three samples;
- Four participants returned at least one questionable or unsatisfactory z-score;

• Laboratory **10** returned unsatisfactory z-scores for all test samples demonstrating an unsatisfactory performance. Results for all test samples were lower than the assigned value (negative bias).

6.4 Z-Score scatter plot

Samples S1 and S2 were blind duplicates. A scatter plot was used for the evaluation of the participants' within-laboratory repeatability (Figure 7).

Most laboratories are plotted in the upper-right or lower-left quadrants. Points close to the diagonal axis demonstrate excellent repeatability while points close to zero demonstrate excellent repeatability and accuracy.



S1 and S2 - Methamphetamine

Figure 7 z-scores scatter plot

6.5 E_n-Score

The dispersal of participants' E_n -scores is graphically presented in Figure 8. Where a laboratory did not report an expanded uncertainty with a result, an expanded uncertainty of zero (0) was used to calculate the E_n -score.



Figure 8 Summary of participants' En-Score

68 of 77 numeric results (88%) returned a satisfactory $E_n\mbox{-score with }|E_n| \leq 1$.

- Twenty-one participants (81%) 1, 3, 4, 6, 7, 8, 9, 11 (only reported two results), 12, 13, 14, 15, 17, 18, 19, 21, 22, 23, 24, 25 and 26 returned satisfactory scores for all three samples;
- Three laboratories returned at least one questionable E_n-score; and
- Laboratories 10 and 16 returned $|E_n| > 1$ for all samples.

6.6 Identification of Cutting Agent

Twenty-five laboratories (96%) reported on the identity of the cutting agents in Samples S1 and S2 and eighteen (72%) in Sample S3 and the findings are presented in Table 8.

Samples were methamphetamine hydrochloride approximately 80% base (m/m) supplied by the Australian Federal Police (AFP).

Caffeine was used to prepare Samples S1 and S2. Twenty-five laboratories reported on the diluent and correctly identified it.

Dimethylsulfone (DMS or MMS) was used to prepare Sample S3. Eighteen laboratories reported the diluent and correctly identified it.

6.7 Participants' Analytical Methods

Participants were requested to analyse the samples using their normal test methods and to report a single result for each sample as they would normally report to a client. Results reported in this way reflect the true variability of results reported to laboratory clients. The method descriptions provided by participants are presented in Table 1.

| Accredited | Laboratory Code |
|---|--|
| Yes to ISO 17025 | 1 2 3 4 6 8 9 11 12 14 17 18 19 20 21 24 25 26 (Qualitative only) |
| Other (Unspecified) | 13 15 20 |
| No | 5 7 10 16 22 23 |
| Sample Mass Used (mg) | Laboratory Code |
| 4-10 | 24 |
| 11-30 | 2 3 5 8 10 11 12 13 14 15 18 20 23 25 26 |
| 31-50 | 1 6 7 9 13 15 16 19 21 22 |
| 51-100 | 4 |
| 101-150 | 17 20 |
| Instrument Used for quantification | Laboratory Code |
| GC-FID | 8 10 |
| GC-MS | 8 16 |
| | |
| UPLC-MS(MS) | 1 |
| UPLC-MS(MS) HPLC(UPLC)-DAD | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR Sources of Calibration Standard | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 Laboratory Code |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR Sources of Calibration Standard NMI Australia | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 Laboratory Code 2 3 4 6 8 11 14 18 20 25 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR Sources of Calibration Standard NMI Australia Lipomed | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 Laboratory Code 2 3 4 6 8 11 14 18 20 25 7 9 13 15 16 19 26 |
| UPLC-MS(MS) HPLC(UPLC)-DAD HPLC-FLD QNMR Sources of Calibration Standard NMI Australia Lipomed Sigma Aldrich | 1 2 3 4 7 9 11 12 13 14 15 17 18 19 20 21 22 23 25 26 6 5 24 Laboratory Code 2 3 4 6 8 11 14 18 20 25 7 9 13 15 16 19 26 1 5 12 21 23 |

A summary of accreditation status, participants' methods and reference standards is presented below.

Plots of extraction solvent vs z-score, measurement instrument vs z-score and calibration standard vs z-score are presented in Figures 9 to 11. A variety of solvents and calibration standards were used. HPLC (UPLC)-DAD was the most common measurement technique with the participants. No trends were identified.



Figure 9 Extraction solvent vs z-score



Figure 10 Measurement instrument vs z-score



Figure 11 Calibration standard vs z-score

6.8 Summary of Participation and Performance in Methamphetamine Studies

Overall percentages of satisfactory z-scores and E_n -scores obtained by laboratories since 2009 are presented in Figure 12. The proportion of satisfactory z-scores and E_n -scores over 10 years has increased with an average of 82% and 77% respectively.



Figure 12 Summary of participants' performance since 2009

7 REFERENCES

- [1] ISO/IEC 17043 2010, Conformity assessment General requirements for proficiency testing.
- [2] NMI 2016, *Chemical Proficiency Testing Study Protocol*, viewed 18 June 2019, <<u>http://www.measurement.gov.au></u>.
- [3] NMI 2016, *Chemical Proficiency Testing Statistical Manual*, viewed 18 June 2019, <<u>http://www.measurement.gov.au</u>>.
- [4] Thompson, M., Ellison, S.L.R. & Wood, R. 2006. 'The international harmonized protocol for proficiency testing of (chemical) analytical laboratories', *Pure Appl. Chem*, vol 78, pp 145-196.
- [5] ISO/IEC 17025 2017, General requirements for the competence of testing and calibration laboratories.
- [6] Eurachem 2012, Quantifying Uncertainty in Analytical Measurement, 3rd edition, viewed 10 June 2019,
 http://www.eurachem.org/images/stories/Guides/pdf/QUAM2012 P1.pdf>.
- [7] ISO/IEC 98-3 2008, Uncertainty of measurement Part 3 Guide to the expression of uncertainty in measurement.
- [8] ISO/IEC 13528 2015, Statistical methods for use in proficiency testing by interlaboratory comparisons.
- [9] Thompson M. and Lowthian, P.J. 1995, 'A Horwitz-like function describes precision in a proficiency test', *Analyst*, vol 120, pp 271-272
- [10] Thompson, M. and Fearn, T.,2001 A new test for sufficient homogeneity, *Analyst*, vol 126, pp 1414-1417.

APPENDIX 1 - MEASUREMENT UNCERTAINTY OF THE ROBUST AVERAGE

When the robust average is calculated using the procedure described in 'ISO13528:2015, Statistical methods for use in proficiency testing by interlaboratory comparisons – Annex C'⁸, the uncertainty is estimated as:

| $u_{rob average} = 1$ | $1.25*S_{rob\ average}/\sqrt{p}$ | Equation 4 |
|--|---|------------|
| where: | | |
| u _{rob average} Srob average | robust average standard uncertainty robust average standard deviation | |
| р | number of results | |

The expanded uncertainty $(U_{rob\ average})$ is the standard uncertainty multiplied by a coverage factor of 2 at approximately 95% confidence level.

A worked example is set out below in Table 10.

Table 10 Uncertainty of assigned value for Sample S3 as % base (m/m)

| No. results (p) | 25 |
|--------------------|-------|
| Robust average | 23.23 |
| $S_{rob\ average}$ | 1.07 |
| Urob average | 0.268 |
| k | 2 |
| $U_{rob\ average}$ | 0.535 |

The robust average for Sample S3 is $23.2 \pm 0.5\%$ Methamphetamine base (m/m).

APPENDIX 2 – REFERENCE VALUE AND HOMOGENEITY TESTING

Five sample bottles from each of Samples S1 and S2 (duplicate) were selected at random for the purpose of assigning a reference value. Samples were analysed in duplicate.

Measurements were made using quantitative nuclear magnetic resonance spectrometry (QNMR) with maleic acid as internal standard. A Certified Reference Material of maleic acid was obtained from NMI Chemical Reference Materials. The purity data supplied with the material is shown in Table 11 and is traceable to the SI unit for mass, the kilogram (kg), through QNMR. Internal standard solutions were prepared gravimetrically in D₂O.

| | Supplier | Catalogue / Lot No | Purity (95% confidence) |
|-------------|----------------------------------|--------------------|-------------------------|
| Maleic acid | NMI Chemical Reference Materials | QNMR010.2018.01 | 98.8 ± 0.12 % |

Samples were prepared gravimetrically by accurately weighing 20 mg of sample, dissolving in 650 μ L of internal standard solution (12.9 mg/g for S1/S2) and accurately weighing the final solution. Samples were analysed on a Bruker 500 MHz Ascend NMR spectrometer, using a QNMR relaxation time of 25 s. The mass fraction of methamphetamine was determined from the NMR response at 1.15 ppm.

The average of mass fractions determined for Samples S1 and S2 (Table 12), was used as the reference value and the assigned value for the PT study. The standard uncertainty on the mass fraction reference value was estimated in accordance with the ISO GUM by combining standard uncertainty terms for method precision, sample homogeneity, weighing of sample, preparation and addition of standard solution, the very small uncertainties in molecular weights and an estimate of potential bias made by comparing the results from different NMR signals.

| Bottle Fill | Methamphetamine (% base m/m) | | |
|-------------|------------------------------|-------------|--|
| No. | Replicate 1 | Replicate 2 | |
| 115 | 57.0 | 56.6 | |
| 116 | 58.2 | 57.5 | |
| 127 | 57.5 | 57.4 | |
| 129 | 57.4 | 58.0 | |
| 136 | 58.0 | 57.8 | |
| 203 | 57.9 | 57.6 | |
| 219 | 57.9 | 57.8 | |
| 227 | 57.9 | 58.1 | |
| 239 | 58.5 | 57.8 | |
| 248 | 57.8 | 57.6 | |
| Mean | 57.7 | | |
| CV | 0.73% | | |

Table 12 Reference value for Sample S1/S2

Reference value 57.7 \pm 1.2% methamphetamine base $(m\!/\!m)^a$

^a The uncertainty is an expanded uncertainty at 95% confidence level. A coverage factor k was calculated using the effective degrees of freedom derived from the Welch-Satterthwaite⁹ equation (k = 2.1).

The measured reference value was in agreement with the gravimetric preparation value and the robust average of participants' results, within their respective associated uncertainties.

Homogeneity check was based on that described by Thompson and Fearn¹⁰ which is also the procedure described in the International Protocol.⁴

| Test | Test Value | Critical Value | Result |
|-------------------------------|------------|-------------------|--------|
| Cochran | 0.28 | 0.60 | Pass |
| S_a/σ | 0.17 | 0.5 | Pass |
| s ² _{sam} | 0.096 | 0.59 | Pass |

Thompson and Fearn Homogeneity Tests for duplicate pair Samples S1 and S2

Samples were found to be sufficiently homogeneous for use in a proficiency test with a target standard deviation of 3%.

APPENDIX 3 - ACRONYMS AND ABBREVIATIONS

| ASCLD | American Society of Crime Laboratory Directors | |
|-----------------------------|---|--|
| CITAC | Cooperation on International Traceability in Analytical Chemistry | |
| CRM | Certified Reference Material | |
| CV | Coefficient of Variation | |
| DAD | Diode Array Detector | |
| DMS | Dimethyl sulfone | |
| $ \mathbf{E}_{\mathbf{n}} $ | Absolute value of an E _n -score | |
| FID | Flame Ionization Detector | |
| FLD | Fluorescence Detector | |
| GC | Gas Chromatography | |
| GC-MS | Gas Chromatography Mass Spectrometry | |
| GUM | Guide to the expression of uncertainty in measurement | |
| HPLC | High Performance Liquid Chromatography | |
| ISO | International Standards Organisation | |
| LC | Liquid Chromatography | |
| Max | Maximum value in a set of results | |
| Md | Median | |
| Min | Minimum value in a set of results | |
| NATA | National Association of Testing Authorities | |
| NMI | National Measurement Institute Australia | |
| NR | Not Reported | |
| NT | Not Tested | |
| PCV | Performance Coefficient of Variation | |
| PDA | Photodiode Array | |
| PT | Proficiency Test | |
| QNMR | Quantitative Nuclear Magnetic Resonance | |
| Robust CV | Robust Coefficient of Variation | |
| Robust SD | Robust Standard Deviation | |
| SI | International System of Units | |
| Target SD (σ) | Target standard deviation | |
| UPLC | Ultra Performance Liquid Chromatography | |
| UV | Ultraviolet | |
| z | Absolute value of a z-score | |

END OF REPORT