



Australian Government
Department of Industry,
Innovation and Science

**National
Measurement
Institute**

Proficiency Test Report AQA 19-16 Trace Elements in Sea Water

January 2020

ACKNOWLEDGMENTS

This study was conducted by the National Measurement Institute (NMI). Support funding was provided by the Australian Government Department of Industry, Innovation and Science.

I would like to thank the management and staff of the participating laboratories for supporting the study. It is only through widespread participation that we can provide an effective service to laboratories.

The assistance of the following NMI staff members in the planning, conduct and reporting of the study is acknowledged.

Luminita Antin

Andrew Evans

Luke Viskovic

Hamish Lenton

Raluca Iavetz

Manager, Chemical Proficiency Testing

Phone: 61-2-9449 0111

proficiency@measurement.gov.au



Accredited for compliance with ISO/IEC 17043

TABLE OF CONTENTS

| | | |
|-----|---|-----|
| 1 | SUMMARY | 4 |
| 2 | INTRODUCTION | 5 |
| 2.1 | NMI Proficiency Testing Program | 5 |
| 2.2 | Study Aims | 5 |
| 2.3 | Study Conduct | 5 |
| 3 | STUDY INFORMATION | 5 |
| 3.1 | Selection of Matrices and Inorganic Analytes | 5 |
| 3.2 | Participation | 6 |
| 3.3 | Test Material Specification | 6 |
| 3.4 | Laboratory Code | 6 |
| 3.5 | Sample Preparation, Analysis and Homogeneity Testing | 6 |
| 3.6 | Stability of Analytes | 6 |
| 3.7 | Sample Storage, Dispatch and Receipt | 6 |
| 3.8 | Instructions to Participants | 6 |
| 3.9 | Interim Report | 7 |
| 4 | PARTICIPANT LABORATORY INFORMATION | 8 |
| 4.1 | Methodology for Total and Dissolved Elements | 8 |
| 4.2 | Additional Information | 8 |
| 4.3 | Basis of Participants' Measurement Uncertainty Estimates | 8 |
| 4.4 | Participant Comments on this PT Study or Suggestions for Future Studies | 9 |
| 5 | PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS | 10 |
| 5.1 | Results Summary | 10 |
| 6 | TABLES AND FIGURES | 12 |
| 7 | DISCUSSION OF RESULTS | 92 |
| 7.1 | Assigned Value | 92 |
| 7.2 | Measurement Uncertainty Reported by Participants | 92 |
| 7.3 | E_n -score | 93 |
| 7.4 | z-Score | 93 |
| 7.5 | Participants' Results and Analytical Methods for Total and Dissolved Elements | 100 |
| 7.6 | Comparison with Previous NMI Proficiency Tests of Metals in Water | 110 |
| 7.7 | Reference Materials and Certified Reference Materials | 111 |
| 8 | REFERENCES | 113 |
| | APPENDIX 1 – SAMPLE PREPARATION, ANALYSIS AND HOMOGENEITY TESTING | 114 |
| | Sample Preparation | 114 |
| | Sample Analysis and Homogeneity Testing | 114 |
| | APPENDIX 2 – ASSIGNED VALUE, Z-SCORE AND E_N SCORE CALCULATION | 115 |
| | APPENDIX 3 - USING PT DATA FOR UNCERTAINTY ESTIMATION | 116 |
| | APPENDIX 4 - ACRONYMS AND ABBREVIATIONS | 118 |
| | APPENDIX 5 - INSTRUMENT DETAILS FOR DISSOLVED ELEMENTS | 119 |

1 SUMMARY

This report presents the results of the proficiency test AQA 19-16, Trace Elements in Sea Water. The study focused on the measurement of dissolved: Al, Ag, As, Be, Cd, Co, Cr, Cu, Fe, Hg, Mn, Ni, P, Pb, Se, Sn, Tl, U, V and Zn and total: Al, Ag, As, Ba, Cd, Cr, Cu, Fe, Mn, Mo, Ni, P, Pb, Sb, Se, Sn, Tl, U, V and Zn.

The assigned values were the robust average of participants' results. The associated uncertainties were estimated from the robust standard deviation of the participants' results.

The outcomes of the study were assessed against the aims as follows, to:

- i. *compare the performance of participant laboratories and assess their accuracy;*

Laboratory performance was assessed using both z-scores and E_n -scores.

Of 347 numeric results, 328 (95%) returned a satisfactory score of $|z| \leq 2.0$.

Of 347 results, 315 (91%) returned a satisfactory score of $|E_n| \leq 1.0$.

- ii. *evaluate the laboratories' methods used in determination of inorganic analytes in seawater;*

Aluminium, nickel and phosphorus at low levels in seawater were the analytes which presented the most analytical difficulty to participating laboratories.

- iii. *compare the performance of participant laboratories with their past performance;*

Overall, participants performed better in this study than in the previous NMI PT studies in seawater AQA 18-16 and AQA 17-16.

- iv. *develop the practical application of traceability and measurement uncertainty and provide participants with information that will be useful in assessing their uncertainty estimates;*

Of 347 numerical results, 330 (95%) were reported with an expanded measurement uncertainty. An example of estimating measurement uncertainty using only the proficiency testing data is given in Appendix 3.

- v. *produce materials that can be used in method validation and as control samples.*

Surplus test samples from the present study are available for sale. The samples were assessed for homogeneity and stability and are well characterised, both by in-house testing and from the results of the proficiency round.

2 INTRODUCTION

2.1 NMI Proficiency Testing Program

The National Measurement Institute (NMI) is responsible for Australia's national measurement infrastructure providing a wide range of services, including a chemical proficiency testing program.

Proficiency testing (PT) "is evaluation of participant performance against pre-established criteria by means of interlaboratory comparison."¹ NMI PT studies target chemical testing in areas of high public significance such as trade, environment and food safety. NMI offers studies in:

- inorganic analytes in soil, water, food and pharmaceuticals;
- pesticide residues in fruit and vegetables, soil and water;
- petroleum hydrocarbons in soil and water;
- PFAS in water, soil, biota and food;
- allergens in food;
- controlled drug assay; and
- folic acid in flour.

AQA 19-16 is the 24th NMI proficiency study of metals in water.

2.2 Study Aims

The aims of the study were to:

- compare the performance of participant laboratories and assess their accuracy;
- evaluate the laboratories methods used in determination of inorganic analytes in seawater;
- compare the performance of participant laboratories with their past performance;
- develop the practical application of traceability and measurement uncertainty; and
- produce materials that can be used in method validation and as control samples.

2.3 Study Conduct

The conduct of NMI proficiency tests is described in the NMI Chemical Proficiency Testing Study Protocol.² The statistical methods used are described in the NMI Chemical Proficiency Statistical Manual.³ These documents have been prepared with reference to ISO Standard 17043¹ and The International Harmonized Protocol for Proficiency Testing of (Chemical) Analytical Laboratories.⁴

NMI is accredited by National Association of Testing Authorities, Australia (NATA) to ISO/IEC 17043 as a provider of proficiency testing schemes. This proficiency test is within the scope of NMI's accreditation.

The choice of the test method was left to the participating laboratories.

3 STUDY INFORMATION

3.1 Selection of Matrices and Inorganic Analytes

The 40 tests were selected from those for which an investigation level is published in Australian and New Zealand Guidelines for Fresh and Marine Water Quality⁵ and are commonly measured by water testing laboratories.

3.2 Participation

Thirteen laboratories participated and twelve submitted results.

The timetable of the study was:

Invitation issued: 01 October 2019
Samples dispatched: 28 October 2019
Results due: 15 November 2019
Interim report issued: 22 November 2019

3.3 Test Material Specification

Two samples were provided for analysis:

Sample S1 was 100 mL of filtered seawater preserved by adding 2% (v/w) nitric acid; and

Sample S2 was 100 mL of unfiltered seawater preserved by adding 2% (v/w) nitric acid.

3.4 Laboratory Code

All participant laboratories were assigned a confidential code number.

3.5 Sample Preparation, Analysis and Homogeneity Testing

A partial homogeneity testing was conducted in this study. The same sample preparation procedure was followed as in previous studies.² The test samples from the previous studies were demonstrated to be sufficiently homogeneous for evaluation of participants' performance. The results from the partial homogeneity testing are reported in this study as homogeneity values. No homogeneity testing was conducted for U in S1 and S2.

The preparation and analysis are described in Appendix 1.

3.6 Stability of Analytes

No stability study was carried out for samples S1 and S2. Stability studies conducted for similar previous proficiency studies of metals in seawater found no significant changes in any of the analytes' concentration.

3.7 Sample Storage, Dispatch and Receipt

Samples S1 and S2 were refrigerated before dispatch.

The samples were dispatched by courier on 28 October 2019.

A description of the test samples, instructions for participants, and a form for participants to confirm the receipt of the test samples, were sent with the samples.

An Excel spreadsheet for the electronic reporting of results was e-mailed to participants.

3.8 Instructions to Participants

Participants were instructed as follows:

- Quantitatively analyse the samples using your normal test method.
- Participants are asked to report results in units of $\mu\text{g/L}$ for S1 and S2

| SAMPLE S1 filtered, acidified sea water | | SAMPLE S2 unfiltered, acidified sea water | |
|---|---|---|---|
| Test Dissolved | Approximate Conc. Range $\mu\text{g/L}$ | Test Total | Approximate Conc. Range $\mu\text{g/L}$ |
| Al | <50 | Al | >100 |

| | | | |
|----|------|----|------|
| Ag | <10 | Ag | >100 |
| As | <10 | As | >100 |
| Be | <10 | Ba | >100 |
| Cd | <10 | Cd | >100 |
| Co | <10 | Cr | >100 |
| Cr | <10 | Cu | >100 |
| Cu | <50 | Fe | >100 |
| Fe | <50 | Mn | >100 |
| Hg | <10 | Mo | >100 |
| Mn | <10 | Ni | >100 |
| Ni | <10 | P | >100 |
| P | >100 | Pb | >100 |
| Pb | <10 | Sb | >100 |
| Se | <10 | Se | >100 |
| Sn | <10 | Sn | >100 |
| Tl | <10 | Tl | >100 |
| U | <10 | U | >100 |
| V | <10 | V | >100 |
| Zn | <50 | Zn | >100 |

- Report results as you would report to a client. For each analyte in each sample, report the expanded measurement uncertainty associated with your analytical result (e.g. 5.23 \pm 0.51 μ g/L).
- Please send us the requested details regarding the test method and the basis of your uncertainty estimate.
- Return the completed results sheet by e-mail (proficiency@measurement.gov.au) by 15 November 2019. Late results cannot be included in the report.

3.9 Interim Report

An interim report was emailed to participants on 22 November 2019.

4 PARTICIPANT LABORATORY INFORMATION

4.1 Methodology for Total and Dissolved Elements

Summaries of test methods are transcribed in Tables 1 and 2. The instruments and settings reported by participants are presented in Appendix 5.

Table 1 Methodology for Total Elements

| Lab. Code | Method Reference | Sample Volume (mL) | Digestion Temp. (°C) | Digestion Time (min) | Vol. HNO ₃ (mL) | Vol. HCl (mL) |
|-----------|--|--|----------------------|----------------------|----------------------------|---------------|
| 1* | | 10 | 170 | 30 | 1.25 | 0.63 |
| 3 | 3051A | 20 | 170 | 15 | 1 | 1 |
| 6* | 200.7-8 | | | | | |
| 7 | | 30 | 95-100 | 90 | 2 | |
| 8 | ICP-MS:APHA (2017),3125 / ICP-OES:US EPA (1994), Method 200.7 Revision 4.4 / Hg:US EPA,Method 1631 Revision.E | 10 | 105 | 2 | 0.5 | 0.25 |
| 9* | EPA 200.8 | 1mL of HCl was added to 50mL of the 2% HNO ₃ acidified sample received. | | | | |
| 10 | USEPA6020 | 10 | 95 | 90 | 2 | 3 |
| 11 | APHA Method 3030 E 23 rd ed. 2017 (modified) APHA Method 3030 F 23 rd ed. 2017 (modified) | 10 | 100 | 60 | 0.5 | 0.1 (Ag Only) |
| 12 | In House W32a | | | | | |
| 13 | USEPA METHOD 3050B | 50 | 85 | 120 | 1 | 1 |

*Additional Information in Table 2.

4.2 Additional Information

Participants had the option to report additional information for each sample analysed. These are transcribed in Table 2.

Table 2 Additional information

| Lab. Code | Additional Information |
|-----------|--|
| 1 | Samples were digested 10 ml and made to 50 ml. An additional dilution of 3 ml into 15ml was made making a total dilution offline of 25x. |
| 6 | Samples not digested |
| 9 | Where O2 has been selected as the gas it is a H ₂ +N ₂ O mix. |

4.3 Basis of Participants' Measurement Uncertainty Estimates

Participants were requested to provide information about the basis of their uncertainty estimates (Table 3).

Table 3 Basis of Uncertainty Estimate

| Lab. Code | Approach to Estimating MU | Information Sources for MU Estimation | | Guide Document for Estimating MU |
|-----------|--|---|---|----------------------------------|
| | | Precision ^a | Method Bias ^a | |
| 1 | Method still in development-uncertainty not determined to date. | | | |
| 2 | Professional Judgement | | | |
| 3 | Top Down - precision and estimates of the method and laboratory bias | Control Samples-CRM Duplicate Analysis | CRM | NMI Uncertainty Course |
| 4 | Top Down - precision and estimates of the method and laboratory bias | Control Samples Duplicate Analysis | CRM | |
| 6 | Top Down - precision and estimates of the method and laboratory bias | Control Samples Duplicate Analysis Instrumental Calibration | CRM Recoveries of SS | NATA Technical Note 33 |
| 7 | Top Down - precision and estimates of the method and laboratory bias | Control Samples-RM Duplicate Analysis | Recoveries of SS | Nordtest Report TR537 |
| 8 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | Instrument Calibration | CRM Recoveries of SS Instrument Calibration Standard Purity | Eurachem/CITAC Guide |
| 9 | Top Down - precision and estimates of the method and laboratory bias | Control Samples-SS | CRM Laboratory bias from PT studies Recoveries of SS | NATA Technical Note 33 |
| 10 | Bottom Up (ISO/GUM, fish bone/ cause and effect diagram) | Control Samples-SS Duplicate Analysis | CRM Recoveries of SS | Eurachem/CITAC Guide |
| 11 | Standard deviation of replicate analyses multiplied by 2 or 3 | Control Samples-CRM Duplicate Analysis Instrumental Calibration | CRM Recoveries of SS Instrument Calibration | Eurachem/CITAC Guide |
| 12 | Top Down - precision and estimates of the method and laboratory bias | Control Samples-CRM Duplicate Analysis | CRM | Nordtest Report TR537 |
| 13 | Standard deviation of replicate analyses multiplied by 2 or 3 | Control Samples-SS Duplicate Analysis | Recoveries of SS | NATA Technical Note 33 |

^aRM = Reference Material, CRM = Certified Reference Material, SS =Spiked samples.

4.4 Participant Comments on this PT Study or Suggestions for Future Studies

The study co-ordinator welcomes comments or suggestions from participants about this study or possible future studies. Such feedback may be useful in improving future studies.

Participants' comments are reproduced in Table 4.

Table 4 Participants' Comments

| Participants' Comments | Study Co-ordinator's Response |
|--|--|
| Other elements of interest are: Li, Si, Ti, Sr, La, Pt, Bi, & Th | These elements have been included in this year program for trace elements in water |

5 PRESENTATION OF RESULTS AND STATISTICAL ANALYSIS

5.1 Results Summary

Participant results are listed in Tables 5 to 44 with results' summary statistics: robust average, median, maximum, minimum, robust standard deviation (SD_{rob}) and robust coefficient of variation (CV_{rob}). Bar charts of results and performance scores are presented in Figures 2 to 41.

An example chart with an interpretation guide is shown in Figure 1.




Figure 1 Guide to Presentation of Results

5.2 Assigned Value

An example of an assigned value calculation using data from the present study is given in Appendix 2. The assigned value is defined as: 'the value attributed to a particular property of a proficiency test item.'¹ In this study, the property is the mass concentration of analyte. Assigned values were the robust average of participants' results; the expanded uncertainties were estimated from the associated robust standard deviations.

5.3 Robust Average

The robust averages and associated expanded measurement uncertainties were calculated using the procedure described in 'Statistical methods for use in proficiency testing by interlaboratory comparisons, ISO13528:2015(E)'.⁶

5.4 Robust Between-Laboratory Coefficient of Variation

The robust between-laboratory coefficient of variation (CV) is a measure of the variability of participants' results and was calculated using the procedure described in ISO13528:2015(E).⁶

5.5 Target Standard Deviation

The target standard deviation (σ) is the product of the assigned value (X) and the performance coefficient of variation (PCV) as presented in Equation 1. This value is used for calculation of participant z-score and provides scaling for laboratory deviation from the assigned value.

$$\sigma = (X) * PCV \quad \text{Equation 1}$$

It is important to note that the PCV is a fixed value and is not the standard deviation of participants' results. The fixed value set for PCV is based on the existing regulation, the

acceptance criteria indicated by the methods, the matrix, the concentration level of analyte and on experience from previous studies. It is backed up by mathematical models such as Thompson Horwitz equation (see Table 45).⁷ By setting a fixed and realistic value for the PCV, the participants' performance does not depend on other participants' performance and can be compared from study to study and against achievable performance.

5.6 z-Score

An example of z-score calculation using data from the present study is given in Appendix 2. For each participant's result a z-score is calculated according to Equation 2 below:

$$z = \frac{(\chi - X)}{\sigma} \quad \text{Equation 2}$$

where:

- z is z-score
- χ is participants' result
- X is the study assigned value
- σ is the target standard deviation

A z-score with absolute value ($|z|$):

- $|z| \leq 2.0$ is satisfactory;
- $2.0 < |z| < 3.0$ is questionable;
- $|z| \geq 3.0$ is unsatisfactory.

5.7 E_n-Score

An example of E_n-score calculation using data from the present study is given in Appendix 3. The E_n-score is complementary to the z-score in assessment of laboratory performance. E_n-score includes measurement uncertainty and is calculated according to Equation 3 below:

$$E_n = \frac{(\chi - X)}{\sqrt{U_\chi^2 + U_X^2}} \quad \text{Equation 3}$$

where:

- E_n is E_n-score
- χ is participants' result
- X is the assigned value
- U_χ is the expanded uncertainty of the participants' result
- U_X is the expanded uncertainty of the assigned value

An E_n-score with absolute value ($|E_n|$):

- $|E_n| \leq 1.0$ is satisfactory;
- $|E_n| > 1.0$ is unsatisfactory.

5.8 Traceability and Measurement Uncertainty

Laboratories accredited to ISO/IEC Standard 17025:2018⁸ must establish and demonstrate the traceability and measurement uncertainty associated with their test results. Guidelines for quantifying uncertainty in analytical measurement are described in the Eurachem/CITAC Guide.⁹

6 TABLES AND FIGURES

Table 5

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Ag |
| Units | µg/L |


Participant Results

| Lab Code | Result | Uncertainty | z-Score | E _n -Score |
|----------|--------|-------------|---------|-----------------------|
| 1 | NT | NT | | |
| 2 | 1.2 | 0.3 | -5.32 | -14.35 |
| 3 | 6.0 | 1.4 | 0.06 | 0.04 |
| 4 | 6.1 | 0.8 | 0.17 | 0.18 |
| 6 | 6.1 | 1.8 | 0.17 | 0.08 |
| 7 | 6.1 | 1.2 | 0.17 | 0.12 |
| 8 | 6.03 | 0.41 | 0.09 | 0.18 |
| 9 | 6 | 1 | 0.06 | 0.05 |
| 10 | 5.14 | 1.21 | -0.91 | -0.66 |
| 11 | 5.9 | NR | -0.06 | -0.36 |
| 12 | 4.7 | 0.8 | -1.40 | -1.54 |
| 13 | 5.92 | 0.89 | -0.03 | -0.03 |


Statistics

| | | |
|--------------------------|------|------|
| Assigned Value* | 5.95 | 0.14 |
| Spike | 6.08 | 0.17 |
| Homogeneity Value | 6.4 | 1.3 |
| Robust Average | 5.95 | 0.11 |
| Median | 6.00 | 0.10 |
| Mean | 5.38 | |
| N | 11 | |
| Max. | 6.1 | |
| Min. | 1.2 | |
| Robust SD | 0.15 | |
| Robust CV | 2.5% | |

*Robust Average excluding Laboratory 2.



z-Scores: S1 - Ag



En-Scores: S1 - Ag




Figure 2

Table 6

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Al |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 19 | 5 | 1.40 | 0.57 |
| 3 | 7.4 | 1.6 | -3.52 | -2.44 |
| 4 | 14 | 4 | -0.72 | -0.34 |
| 6 | NT | NT | | |
| 7 | 18 | 4 | 0.98 | 0.46 |
| 8 | 13.3 | 0.70 | -1.02 | -0.78 |
| 9 | 12.2 | 0.2 | -1.49 | -1.16 |
| 10 | 16.1 | 3.4 | 0.17 | 0.09 |
| 11 | 20 | NR | 1.83 | 1.43 |
| 12 | 13 | 2.5 | -1.15 | -0.69 |
| 13 | NT | NT | | |

Statistics

| | | |
|--------------------------|------|-----|
| Assigned Value* | 15.7 | 3.0 |
| Spike | 12.2 | 0.4 |
| Homogeneity Value | 15.5 | 3.1 |
| Robust Average | 14.9 | 3.4 |
| Median | 14.0 | 2.4 |
| Mean | 14.8 | |
| N | 9 | |
| Max. | 20 | |
| Min. | 7.4 | |
| Robust SD | 4.1 | |
| Robust CV | 28% | |

*Robust Average excluding Laboratory 3.






Figure 3

Table 7

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | As |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 2 | 1 | 0.21 | 0.06 |
| 3 | <10 | NR | | |
| 4 | 1.8 | 0.4 | -0.48 | -0.32 |
| 6 | NT | NT | | |
| 7 | 2.2 | 0.4 | 0.89 | 0.60 |
| 8 | 2.10 | 0.10 | 0.55 | 0.81 |
| 9 | 1.8 | 0.2 | -0.48 | -0.53 |
| 10 | 1.68 | 0.37 | -0.89 | -0.64 |
| 11 | <4 | 2.7 | | |
| 12 | 2.0 | 0.5 | 0.21 | 0.11 |
| 13 | 1.90 | 0.29 | -0.14 | -0.12 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 1.94 | 0.17 |
| Spike | 2.34 | 0.34 |
| Homogeneity Value | 2.25 | 0.45 |
| Robust Average | 1.94 | 0.17 |
| Median | 1.95 | 0.19 |
| Mean | 1.94 | |
| N | 8 | |
| Max. | 2.2 | |
| Min. | 1.68 | |
| Robust SD | 0.20 | |
| Robust CV | 10% | |






Figure 4

Table 8

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Be |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 1 | 1 | -1.58 | -0.30 |
| 3 | 1.1 | 1 | -1.07 | -0.21 |
| 4 | 1.6 | 0.3 | 1.48 | 0.80 |
| 6 | 1.6 | 0.5 | 1.48 | 0.54 |
| 7 | 1.4 | 0.3 | 0.46 | 0.25 |
| 8 | 1.52 | 0.14 | 1.07 | 0.86 |
| 9 | 1.3 | 0.1 | -0.05 | -0.04 |
| 10 | 1.30 | 0.41 | -0.05 | -0.02 |
| 11 | NT | NT | | |
| 12 | 1.3 | 0.3 | -0.05 | -0.03 |
| 13 | 1.02 | 0.15 | -1.48 | -1.16 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 1.31 | 0.20 |
| Spike | 1.30 | 0.06 |
| Homogeneity Value | 1.20 | 0.24 |
| Robust Average | 1.31 | 0.20 |
| Median | 1.30 | 0.22 |
| Mean | 1.31 | |
| N | 10 | |
| Max. | 1.6 | |
| Min. | 1 | |
| Robust SD | 0.25 | |
| Robust CV | 19% | |






Figure 5

Table 9

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Cd |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 1 | 1 | -0.55 | -0.09 |
| 3 | <1 | NR | | |
| 4 | 1.1 | 0.3 | 0.06 | 0.03 |
| 6 | 1.3 | 0.4 | 1.28 | 0.51 |
| 7 | 1.0 | 0.2 | -0.55 | -0.41 |
| 8 | 1.05 | 0.08 | -0.24 | -0.33 |
| 9 | 1.2 | 0.2 | 0.67 | 0.50 |
| 10 | 1.03 | 0.24 | -0.37 | -0.23 |
| 11 | 0.97 | 0.25 | -0.73 | -0.45 |
| 12 | 1.2 | 0.3 | 0.67 | 0.35 |
| 13 | 1.12 | 0.17 | 0.18 | 0.16 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 1.09 | 0.09 |
| Spike | 1.11 | 0.05 |
| Homogeneity Value | 1.08 | 0.22 |
| Robust Average | 1.09 | 0.09 |
| Median | 1.08 | 0.08 |
| Mean | 1.10 | |
| N | 10 | |
| Max. | 1.3 | |
| Min. | 0.97 | |
| Robust SD | 0.11 | |
| Robust CV | 10% | |






Figure 6

Table 10

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Co |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 1 | 1 | -1.16 | -0.21 |
| 3 | 1.4 | 1 | 1.05 | 0.19 |
| 4 | 1.2 | 0.3 | -0.06 | -0.03 |
| 6 | NT | NT | | |
| 7 | 1.4 | 0.3 | 1.05 | 0.59 |
| 8 | 1.24 | 0.09 | 0.17 | 0.20 |
| 9 | 1.2 | 0.2 | -0.06 | -0.04 |
| 10 | 1.09 | 0.27 | -0.66 | -0.41 |
| 11 | 1.2 | NR | -0.06 | -0.08 |
| 12 | 1.3 | 0.3 | 0.50 | 0.28 |
| 13 | 1.06 | 0.16 | -0.83 | -0.75 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 1.21 | 0.12 |
| Spike | 1.19 | 0.03 |
| Homogeneity Value | 1.21 | 0.24 |
| Robust Average | 1.21 | 0.12 |
| Median | 1.20 | 0.11 |
| Mean | 1.21 | |
| N | 10 | |
| Max. | 1.4 | |
| Min. | 1 | |
| Robust SD | 0.15 | |
| Robust CV | 12% | |






Figure 7

Table 11

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Cr |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 3 | 2 | 1.81 | 0.31 |
| 3 | <5 | NR | | |
| 4 | 2.0 | 0.4 | -1.02 | -0.63 |
| 6 | NT | 0.9 | | |
| 7 | 2.6 | 0.5 | 0.68 | 0.37 |
| 8 | 1.38 | 0.12 | -2.77 | -2.29 |
| 9 | 2.5 | 0.5 | 0.40 | 0.22 |
| 10 | 2.28 | 0.62 | -0.23 | -0.11 |
| 11 | 2.88 | 0.83 | 1.47 | 0.56 |
| 12 | 2.1 | 0.4 | -0.73 | -0.45 |
| 13 | 2.28 | 0.34 | -0.23 | -0.15 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 2.36 | 0.41 |
| Spike | 2.21 | 0.08 |
| Homogeneity Value | 2.69 | 0.54 |
| Robust Average | 2.36 | 0.41 |
| Median | 2.28 | 0.32 |
| Mean | 2.34 | |
| N | 9 | |
| Max. | 3 | |
| Min. | 1.38 | |
| Robust SD | 0.49 | |
| Robust CV | 21% | |






Figure 8

Table 12

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Cu |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 13 | 3 | 0.16 | 0.10 |
| 3 | 12 | 1.3 | -0.37 | -0.46 |
| 4 | 13 | 3 | 0.16 | 0.10 |
| 6 | NT | NT | | |
| 7 | 13 | 3 | 0.16 | 0.10 |
| 8 | 12.2 | 0.45 | -0.26 | -0.54 |
| 9 | 1.2 | 0.23 | -6.04 | -13.82 |
| 10 | 12.1 | 2.8 | -0.31 | -0.21 |
| 11 | 13.6 | 3.4 | 0.47 | 0.26 |
| 12 | 14 | 2.5 | 0.68 | 0.50 |
| 13 | 10.8 | 1.62 | -1.00 | -1.05 |

Statistics

| | | |
|--------------------------|------|-----|
| Assigned Value* | 12.7 | 0.8 |
| Spike | 10.9 | 1.8 |
| Homogeneity Value | 11.8 | 2.4 |
| Robust Average | 12.4 | 1.0 |
| Median | 12.6 | 0.6 |
| Mean | 11.5 | |
| N | 10 | |
| Max. | 14 | |
| Min. | 1.2 | |
| Robust SD | 1.3 | |
| Robust CV | 10% | |

*Robust Average excluding Laboratory 9.






Figure 9

Table 13

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Fe |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 16 | 2 | 0.04 | 0.03 |
| 3 | 14 | 1.5 | -0.80 | -0.74 |
| 4 | 14 | 3 | -0.80 | -0.52 |
| 6 | NT | NT | | |
| 7 | 18 | 4 | 0.88 | 0.46 |
| 8 | 11 | 0.46 | -2.05 | -2.28 |
| 9 | 15 | 2.3 | -0.38 | -0.29 |
| 10 | 22.6 | 8.0 | 2.81 | 0.81 |
| 11 | 15.4 | 3.3 | -0.21 | -0.13 |
| 12 | 18 | 3.0 | 0.88 | 0.57 |
| 13 | 17.1 | 2.56 | 0.50 | 0.36 |

Statistics

| | | |
|--------------------------|------|-----|
| Assigned Value | 15.9 | 2.1 |
| Spike | 15.1 | 0.4 |
| Homogeneity Value | 14.7 | 2.9 |
| Robust Average | 15.9 | 2.1 |
| Median | 15.7 | 1.8 |
| Mean | 16.1 | |
| N | 10 | |
| Max. | 22.6 | |
| Min. | 11 | |
| Robust SD | 2.7 | |
| Robust CV | 17% | |






Figure 10

Table 14

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Hg |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 0.34 | 0.09 | 6.29 | 1.73 |
| 3 | 0.14 | 0.5 | -1.33 | -0.07 |
| 4 | 0.2 | 0.1 | 0.95 | 0.24 |
| 6 | 0.16 | 0.05 | -0.57 | -0.25 |
| 7 | 0.22 | 0.04 | 1.71 | 0.89 |
| 8 | 0.13 | 0.008 | -1.71 | -1.41 |
| 9 | 0.17 | 0.05 | -0.19 | -0.08 |
| 10 | 0.20 | 0.05 | 0.95 | 0.42 |
| 11 | 0.209 | 0.059 | 1.30 | 0.51 |
| 12 | <0.5 | NR | | |
| 13 | 0.148 | 0.022 | -1.03 | -0.71 |

Statistics

| | | |
|--------------------------|-------|-------|
| Assigned Value* | 0.175 | 0.031 |
| Spike | 0.201 | 0.006 |
| Homogeneity Value | 0.192 | 0.029 |
| Robust Average | 0.183 | 0.035 |
| Median | 0.185 | 0.032 |
| Mean | 0.192 | |
| N | 10 | |
| Max. | 0.34 | |
| Min. | 0.13 | |
| Robust SD | 0.044 | |
| Robust CV | 24% | |

*Robust Average excluding Laboratory 2.






Figure 11

Table 15

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Mn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 3 | 2 | -1.25 | -0.34 |
| 3 | 4.9 | 1 | 2.19 | 1.17 |
| 4 | 3.5 | 0.6 | -0.34 | -0.29 |
| 6 | NT | NT | | |
| 7 | 3.9 | 0.8 | 0.38 | 0.25 |
| 8 | 3.6 | 0.26 | -0.16 | -0.24 |
| 9 | 4 | 0.6 | 0.56 | 0.47 |
| 10 | 3.78 | 1.04 | 0.16 | 0.08 |
| 11 | 3.72 | 0.69 | 0.05 | 0.04 |
| 12 | 3.4 | 0.7 | -0.52 | -0.39 |
| 13 | 3.65 | 0.55 | -0.07 | -0.07 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 3.69 | 0.26 |
| Spike | 3.44 | 0.16 |
| Homogeneity Value | 3.85 | 0.77 |
| Robust Average | 3.69 | 0.26 |
| Median | 3.69 | 0.21 |
| Mean | 3.75 | |
| N | 10 | |
| Max. | 4.9 | |
| Min. | 3 | |
| Robust SD | 0.33 | |
| Robust CV | 8.9% | |






Figure 12

Table 16

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Ni |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 2 | 1 | 1.27 | 0.30 |
| 3 | <10 | NR | | |
| 4 | 1.6 | 0.4 | -0.32 | -0.15 |
| 6 | NT | NT | | |
| 7 | 1.8 | 0.4 | 0.48 | 0.22 |
| 8 | 1.05 | 0.1 | -2.50 | -1.64 |
| 9 | 1.5 | 0.3 | -0.71 | -0.38 |
| 10 | 1.35 | 0.30 | -1.31 | -0.69 |
| 11 | <7 | NR | | |
| 12 | 2.1 | 0.4 | 1.67 | 0.77 |
| 13 | 2.02 | 0.30 | 1.35 | 0.71 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 1.68 | 0.37 |
| Spike | 1.68 | 0.08 |
| Homogeneity Value | 1.69 | 0.34 |
| Robust Average | 1.68 | 0.37 |
| Median | 1.70 | 0.38 |
| Mean | 1.68 | |
| N | 8 | |
| Max. | 2.1 | |
| Min. | 1.05 | |
| Robust SD | 0.42 | |
| Robust CV | 25% | |






Figure 13

Table 17

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | P |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 195 | 49 | 1.96 | 0.81 |
| 3 | 105 | 20 | -1.25 | -0.69 |
| 4 | NR | NR | | |
| 6 | NT | NT | | |
| 7 | 140 | 30 | 0.00 | 0.00 |
| 8 | 76.8 | 5.21 | -2.26 | -1.34 |
| 9 | 119 | 32 | -0.75 | -0.37 |
| 10 | 194 | 61 | 1.93 | 0.70 |
| 11 | NT | NT | | |
| 12 | NT | NT | | |
| 13 | 150 | 22.5 | 0.36 | 0.19 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 140 | 47 |
| Spike | 142 | 8 |
| Homogeneity Value | 140 | 28 |
| Robust Average | 140 | 47 |
| Median | 140 | 48 |
| Mean | 140 | |
| N | 7 | |
| Max. | 195 | |
| Min. | 76.8 | |
| Robust SD | 50 | |
| Robust CV | 36% | |






Figure 14

Table 18

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Pb |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | <1 | 1 | | |
| 3 | <1 | NR | | |
| 4 | 1.0 | 0.2 | 0.03 | 0.02 |
| 6 | <1 | NR | | |
| 7 | <1 | NR | | |
| 8 | 1.02 | 0.09 | 0.17 | 0.23 |
| 9 | 0.1 | 0.02 | -6.00 | -13.74 |
| 10 | 0.99 | 0.22 | -0.03 | -0.02 |
| 11 | 1.08 | 0.68 | 0.57 | 0.12 |
| 12 | 0.94 | 0.2 | -0.37 | -0.26 |
| 13 | 0.937 | 0.14 | -0.39 | -0.38 |

Statistics

| | | |
|--------------------------|-------|-------|
| Assigned Value* | 0.995 | 0.062 |
| Spike | 0.979 | 0.028 |
| Homogeneity Value | 1.04 | 0.21 |
| Robust Average | 0.974 | 0.079 |
| Median | 0.990 | 0.069 |
| Mean | 0.867 | |
| N | 7 | |
| Max. | 1.08 | |
| Min. | 0.1 | |
| Robust SD | 0.084 | |
| Robust CV | 8.6% | |

*Robust Average excluding Laboratory 9.






Figure 15

Table 19

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Se |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 4 | 2 | 1.80 | 0.42 |
| 3 | 3.2 | 1 | 0.11 | 0.05 |
| 4 | 3 | 1 | -0.32 | -0.14 |
| 6 | NT | NT | | |
| 7 | 3.7 | 0.7 | 1.16 | 0.68 |
| 8 | 3.17 | 0.21 | 0.04 | 0.04 |
| 9 | 3 | 0.7 | -0.32 | -0.18 |
| 10 | 1.85 | 0.46 | -2.75 | -2.11 |
| 11 | <4 | 2.7 | | |
| 12 | 2.9 | 0.6 | -0.53 | -0.34 |
| 13 | 3.11 | 0.47 | -0.08 | -0.06 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 3.15 | 0.41 |
| Spike | 3.01 | 0.10 |
| Homogeneity Value | 4.42 | 0.88 |
| Robust Average | 3.15 | 0.41 |
| Median | 3.11 | 0.13 |
| Mean | 3.10 | |
| N | 9 | |
| Max. | 4 | |
| Min. | 1.85 | |
| Robust SD | 0.49 | |
| Robust CV | 16% | |






Figure 16

Table 20

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Sn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 6 | 2 | 0.14 | 0.06 |
| 3 | 6.0 | 1 | 0.14 | 0.11 |
| 4 | 6 | 2 | 0.14 | 0.06 |
| 6 | 5.4 | 1.7 | -0.54 | -0.28 |
| 7 | 7.2 | 1.4 | 1.50 | 0.91 |
| 8 | 5.39 | 0.44 | -0.56 | -0.86 |
| 9 | 6 | 1.3 | 0.14 | 0.09 |
| 10 | 5.34 | 1.31 | -0.61 | -0.40 |
| 11 | 6.3 | NR | 0.48 | 1.17 |
| 12 | NT | NT | | |
| 13 | 5.76 | 0.86 | -0.14 | -0.13 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 5.88 | 0.36 |
| Spike | 5.98 | 0.17 |
| Homogeneity Value | 6.1 | 1.2 |
| Robust Average | 5.88 | 0.36 |
| Median | 6.00 | 0.29 |
| Mean | 5.94 | |
| N | 10 | |
| Max. | 7.2 | |
| Min. | 5.34 | |
| Robust SD | 0.46 | |
| Robust CV | 7.8% | |






Figure 17

Table 21

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | TI |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 3 | 2 | 0.11 | 0.02 |
| 3 | 2.5 | 1 | -1.02 | -0.44 |
| 4 | 3.2 | 0.6 | 0.56 | 0.40 |
| 6 | 3.1 | 1 | 0.34 | 0.15 |
| 7 | 3.0 | 0.6 | 0.11 | 0.08 |
| 8 | 3.05 | 0.22 | 0.23 | 0.36 |
| 9 | 2.9 | 0.5 | -0.11 | -0.09 |
| 10 | 3.13 | 0.73 | 0.41 | 0.24 |
| 11 | 2.8 | NR | -0.34 | -0.88 |
| 12 | 3.0 | 0.6 | 0.11 | 0.08 |
| 13 | 2.14 | 0.32 | -1.83 | -2.24 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 2.95 | 0.17 |
| Spike | 3.03 | 0.09 |
| Homogeneity Value | 2.94 | 0.59 |
| Robust Average | 2.95 | 0.17 |
| Median | 3.00 | 0.10 |
| Mean | 2.89 | |
| N | 11 | |
| Max. | 3.2 | |
| Min. | 2.14 | |
| Robust SD | 0.22 | |
| Robust CV | 7.5% | |






Figure 18

Table 22

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | U |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 5 | 2 | -0.69 | -0.29 |
| 3 | 5.3 | 1 | -0.33 | -0.27 |
| 4 | 6.1 | 1.0 | 0.62 | 0.50 |
| 6 | NT | NT | | |
| 7 | 5.6 | 1.1 | 0.02 | 0.02 |
| 8 | 5.65 | 0.38 | 0.08 | 0.14 |
| 9 | 5.3 | 0.8 | -0.33 | -0.33 |
| 10 | 6.05 | 1.36 | 0.56 | 0.34 |
| 11 | 5.4 | NR | -0.22 | -0.58 |
| 12 | 5.8 | 1.2 | 0.26 | 0.18 |
| 13 | 5.59 | 0.84 | 0.01 | 0.01 |

Statistics

| | | |
|-----------------------|------|------|
| Assigned Value | 5.58 | 0.31 |
| Spike* | 2.50 | 0.07 |
| Robust Average | 5.58 | 0.31 |
| Median | 5.60 | 0.27 |
| Mean | 5.58 | |
| N | 10 | |
| Max. | 6.1 | |
| Min. | 5 | |
| Robust SD | 0.39 | |
| Robust CV | 7% | |

*Incurred value not included






Figure 19

Table 23

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | V |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 5 | 2 | 0.50 | 0.17 |
| 3 | 2.5 | 1 | -3.08 | -2.04 |
| 4 | 4.7 | 0.9 | 0.07 | 0.05 |
| 6 | NT | NT | | |
| 7 | 4.8 | 1.0 | 0.22 | 0.14 |
| 8 | 4.49 | 0.34 | -0.23 | -0.34 |
| 9 | 4.5 | 0.8 | -0.22 | -0.17 |
| 10 | 4.14 | 1.18 | -0.73 | -0.42 |
| 11 | 5.0 | 2.7 | 0.50 | 0.13 |
| 12 | 4.7 | 0.8 | 0.07 | 0.06 |
| 13 | 5.12 | 0.77 | 0.67 | 0.56 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value | 4.65 | 0.33 |
| Spike | 5.21 | 0.51 |
| Homogeneity Value | 4.17 | 0.83 |
| Robust Average | 4.65 | 0.33 |
| Median | 4.70 | 0.27 |
| Mean | 4.50 | |
| N | 10 | |
| Max. | 5.12 | |
| Min. | 2.5 | |
| Robust SD | 0.41 | |
| Robust CV | 8.8% | |






Figure 20

Table 24

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S1 |
| Matrix. | Sea Water |
| Analyte. | Zn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NT | NT | | |
| 2 | 17 | 4 | 10.66 | 2.55 |
| 3 | 6.4 | 1 | -0.14 | -0.10 |
| 4 | 7 | 2 | 0.47 | 0.21 |
| 6 | NT | NT | | |
| 7 | 6.4 | NR | -0.14 | -0.16 |
| 8 | 5.07 | 0.34 | -1.50 | -1.54 |
| 9 | 7.4 | 2.1 | 0.88 | 0.38 |
| 10 | 5.40 | 1.31 | -1.16 | -0.72 |
| 11 | 6.9 | 2.9 | 0.37 | 0.12 |
| 12 | 8.1 | 2.0 | 1.59 | 0.71 |
| 13 | 6.18 | 0.93 | -0.37 | -0.28 |

Statistics

| | | |
|--------------------------|------|------|
| Assigned Value* | 6.54 | 0.89 |
| Spike | 7.59 | 0.49 |
| Homogeneity Value | 7.6 | 1.5 |
| Robust Average | 6.75 | 0.99 |
| Median | 6.65 | 0.65 |
| Mean | 7.59 | |
| N | 10 | |
| Max. | 17 | |
| Min. | 5.07 | |
| Robust SD | 1.26 | |
| Robust CV | 19% | |

*Robust Average excluding Laboratory 2.






Figure 21

Table 25

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Ag |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 110 | 12 | -0.35 | -0.32 |
| 4 | NT | NT | | |
| 6 | 120 | 20 | 0.53 | 0.29 |
| 7 | 120 | 24 | 0.53 | 0.25 |
| 8 | 108 | 3.83 | -0.53 | -1.08 |
| 9 | 113.8 | 19.3 | -0.02 | -0.01 |
| 10 | 116 | 27 | 0.18 | 0.07 |
| 11 | 115 | 51 | 0.09 | 0.02 |
| 12 | NT | NT | | |
| 13 | 112 | 17 | -0.18 | -0.11 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 114 | 4 |
| Spike | 122 | 3 |
| Homogeneity Value | 115 | 23 |
| Robust Average | 114 | 4 |
| Median | 114 | 4 |
| Mean | 114 | |
| N | 8 | |
| Max. | 120 | |
| Min. | 108 | |
| Robust SD | 5 | |
| Robust CV | 4.4% | |






Figure 22

Table 26

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Al |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 660 | 70 | -1.54 | -0.86 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 820 | 160 | 0.51 | 0.20 |
| 8 | 811 | 28.6 | 0.40 | 0.25 |
| 9 | 654 | 100 | -1.62 | -0.81 |
| 10 | 937 | 197 | 2.01 | 0.68 |
| 11 | 790 | 160 | 0.13 | 0.05 |
| 12 | NT | NT | | |
| 13 | NT | NT | | |

Statistics

| | | |
|--------------------------|-----|-----|
| Assigned Value | 780 | 120 |
| Spike | 853 | 24 |
| Homogeneity Value | 840 | 170 |
| Robust Average | 780 | 120 |
| Median | 800 | 120 |
| Mean | 779 | |
| N | 6 | |
| Max. | 937 | |
| Min. | 654 | |
| Robust SD | 122 | |
| Robust CV | 16% | |






Figure 23

Table 27

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | As |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 226 | NR | -0.85 | -1.50 |
| 2 | NT | NT | | |
| 3 | 260 | 44 | 0.53 | 0.28 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 250 | 50 | 0.12 | 0.06 |
| 8 | 239 | 10 | -0.32 | -0.46 |
| 9 | 247 | 30 | 0.00 | 0.00 |
| 10 | 239 | 53 | -0.32 | -0.15 |
| 11 | 312 | 69 | 2.63 | 0.92 |
| 12 | NT | NT | | |
| 13 | 247 | 37 | 0.00 | 0.00 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 247 | 14 |
| Spike | 258 | 7 |
| Homogeneity Value | 270 | 50 |
| Robust Average | 247 | 14 |
| Median | 247 | 10 |
| Mean | 253 | |
| N | 8 | |
| Max. | 312 | |
| Min. | 226 | |
| Robust SD | 15 | |
| Robust CV | 6.1% | |






Figure 24

Table 28

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Ba |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 360 | 42 | 0.47 | 0.36 |
| 4 | NT | NT | | |
| 6 | 340 | 50 | -0.12 | -0.08 |
| 7 | 340 | 68 | -0.12 | -0.06 |
| 8 | 312 | 11.3 | -0.93 | -1.70 |
| 9 | 351 | 63 | 0.20 | 0.11 |
| 10 | 342 | 94 | -0.06 | -0.02 |
| 11 | 336 | 48 | -0.23 | -0.16 |
| 12 | NT | NT | | |
| 13 | 368 | 55 | 0.70 | 0.42 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 344 | 15 |
| Spike | 367 | 10 |
| Homogeneity Value | 350 | 70 |
| Robust Average | 344 | 15 |
| Median | 341 | 9 |
| Mean | 344 | |
| N | 8 | |
| Max. | 368 | |
| Min. | 312 | |
| Robust SD | 17 | |
| Robust CV | 4.9% | |






Figure 25

Table 29

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Cd |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 136 | NR | 0.38 | 0.83 |
| 2 | NT | NT | | |
| 3 | 120 | 14 | -0.84 | -0.72 |
| 4 | NT | NT | | |
| 6 | 130 | 20 | -0.08 | -0.05 |
| 7 | 140 | 28 | 0.69 | 0.31 |
| 8 | 127 | 4.46 | -0.31 | -0.54 |
| 9 | 139 | 21 | 0.61 | 0.37 |
| 10 | 127 | 30 | -0.31 | -0.13 |
| 11 | 125 | 28 | -0.46 | -0.21 |
| 12 | NT | NT | | |
| 13 | 133 | 20 | 0.15 | 0.10 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 131 | 6 |
| Spike | 141 | 6 |
| Homogeneity Value | 130 | 30 |
| Robust Average | 131 | 6 |
| Median | 130 | 6 |
| Mean | 131 | |
| N | 9 | |
| Max. | 140 | |
| Min. | 120 | |
| Robust SD | 8 | |
| Robust CV | 6.1% | |






Figure 26

Table 30

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Cr |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 173 | NR | -0.39 | -1.00 |
| 2 | NT | NT | | |
| 3 | 200 | 20 | 1.11 | 0.94 |
| 4 | NT | NT | | |
| 6 | 180 | 30 | 0.00 | 0.00 |
| 7 | 190 | 38 | 0.56 | 0.26 |
| 8 | 180 | 6.34 | 0.00 | 0.00 |
| 9 | 176 | 37 | -0.22 | -0.11 |
| 10 | 173 | 47 | -0.39 | -0.15 |
| 11 | 184 | 34 | 0.22 | 0.12 |
| 12 | NT | NT | | |
| 13 | 174 | 26 | -0.33 | -0.22 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 180 | 7 |
| Spike | 186 | 5 |
| Homogeneity Value | 190 | 40 |
| Robust Average | 180 | 7 |
| Median | 180 | 7 |
| Mean | 181 | |
| N | 9 | |
| Max. | 200 | |
| Min. | 173 | |
| Robust SD | 8 | |
| Robust CV | 4.4% | |






Figure 27

Table 31

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Cu |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 286 | NR | -0.24 | -0.47 |
| 2 | NT | NT | | |
| 3 | 280 | 30 | -0.44 | -0.39 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 300 | 60 | 0.24 | 0.11 |
| 8 | 296 | 10.4 | 0.10 | 0.16 |
| 9 | 296 | 56 | 0.10 | 0.05 |
| 10 | 314 | 73 | 0.72 | 0.28 |
| 11 | 307 | 74 | 0.48 | 0.19 |
| 12 | NT | NT | | |
| 13 | 252 | 38 | -1.40 | -1.00 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 293 | 15 |
| Spike | 314 | 9 |
| Homogeneity Value | 310 | 60 |
| Robust Average | 293 | 15 |
| Median | 296 | 13 |
| Mean | 291 | |
| N | 8 | |
| Max. | 314 | |
| Min. | 252 | |
| Robust SD | 17 | |
| Robust CV | 5.8% | |






Figure 28

Table 32

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Fe |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 333 | NR | 0.12 | 0.31 |
| 2 | NT | NT | | |
| 3 | 310 | 34 | -0.58 | -0.52 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 330 | 66 | 0.03 | 0.01 |
| 8 | 331 | 11.7 | 0.06 | 0.11 |
| 9 | 339 | 51 | 0.30 | 0.19 |
| 10 | 314 | 111 | -0.46 | -0.13 |
| 11 | 360 | 44 | 0.94 | 0.68 |
| 12 | NT | NT | | |
| 13 | 324 | 49 | -0.15 | -0.10 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 329 | 13 |
| Spike | 363 | 10 |
| Homogeneity Value | 379 | 76 |
| Robust Average | 329 | 13 |
| Median | 331 | 9 |
| Mean | 330 | |
| N | 8 | |
| Max. | 360 | |
| Min. | 310 | |
| Robust SD | 15 | |
| Robust CV | 4.6% | |






Figure 29

Table 33

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Mn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 320 | 38 | 0.88 | 0.66 |
| 4 | NT | NT | | |
| 6 | 290 | 60 | -0.14 | -0.07 |
| 7 | 290 | 58 | -0.14 | -0.07 |
| 8 | 290 | 12.1 | -0.14 | -0.24 |
| 9 | 307 | 46 | 0.44 | 0.27 |
| 10 | 280 | 77 | -0.48 | -0.18 |
| 11 | 295 | 12 | 0.03 | 0.06 |
| 12 | NT | NT | | |
| 13 | 285 | 43 | -0.31 | -0.20 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 294 | 11 |
| Spike | 309 | 9 |
| Homogeneity Value | 300 | 60 |
| Robust Average | 294 | 11 |
| Median | 290 | 6 |
| Mean | 295 | |
| N | 8 | |
| Max. | 320 | |
| Min. | 280 | |
| Robust SD | 12 | |
| Robust CV | 4.1% | |






Figure 30

Table 34

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Mo |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 150 | 19 | 0.42 | 0.29 |
| 4 | NT | NT | | |
| 6 | 140 | 20 | -0.28 | -0.19 |
| 7 | 150 | 30 | 0.42 | 0.19 |
| 8 | 132 | 4.80 | -0.83 | -1.29 |
| 9 | 148 | 22 | 0.28 | 0.17 |
| 10 | 133 | 30 | -0.76 | -0.35 |
| 11 | 150 | 39 | 0.42 | 0.15 |
| 12 | NT | NT | | |
| 13 | 149 | 22 | 0.35 | 0.21 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 144 | 8 |
| Spike | 150 | 5 |
| Homogeneity Value | 150 | 30 |
| Robust Average | 144 | 8 |
| Median | 149 | 2 |
| Mean | 144 | |
| N | 8 | |
| Max. | 150 | |
| Min. | 132 | |
| Robust SD | 9 | |
| Robust CV | 6.3% | |






Figure 31

Table 35

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Ni |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 140 | 28 | -0.07 | -0.03 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 160 | 32 | 1.35 | 0.58 |
| 8 | 142 | 5.00 | 0.07 | 0.11 |
| 9 | 142 | 24 | 0.07 | 0.04 |
| 10 | 143 | 32 | 0.14 | 0.06 |
| 11 | 136 | 55 | -0.35 | -0.09 |
| 12 | NT | NT | | |
| 13 | 129 | 19 | -0.85 | -0.58 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 141 | 8 |
| Spike | 154 | 4 |
| Homogeneity Value | 150 | 30 |
| Robust Average | 141 | 8 |
| Median | 142 | 3 |
| Mean | 142 | |
| N | 7 | |
| Max. | 160 | |
| Min. | 129 | |
| Robust SD | 8 | |
| Robust CV | 5.7% | |






Figure 32

Table 36

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | P |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 604 | NR | 0.41 | 0.24 |
| 2 | NT | NT | | |
| 3 | 500 | 92 | -1.38 | -0.59 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 560 | 110 | -0.34 | -0.13 |
| 8 | 392 | 26.6 | -3.24 | -1.82 |
| 9 | 548 | 148 | -0.55 | -0.18 |
| 10 | 630 | 198 | 0.86 | 0.23 |
| 11 | 850 | NR | 4.66 | 2.70 |
| 12 | NT | NT | | |
| 13 | 625 | 94 | 0.78 | 0.33 |

Statistics

| | | |
|--------------------------|-----|-----|
| Assigned Value | 580 | 100 |
| Spike | 598 | 50 |
| Homogeneity Value | 500 | 100 |
| Robust Average | 580 | 100 |
| Median | 582 | 56 |
| Mean | 589 | |
| N | 8 | |
| Max. | 850 | |
| Min. | 392 | |
| Robust SD | 120 | |
| Robust CV | 21% | |






Figure 33

Table 37

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Pb |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 130 | NR | 0.92 | 1.57 |
| 2 | NT | NT | | |
| 3 | 97 | 13 | -1.85 | -1.49 |
| 4 | NT | NT | | |
| 6 | 120 | 20 | 0.08 | 0.05 |
| 7 | 130 | 26 | 0.92 | 0.41 |
| 8 | 119 | 4.31 | 0.00 | 0.00 |
| 9 | 120 | 23 | 0.08 | 0.04 |
| 10 | 118 | 26 | -0.08 | -0.04 |
| 11 | 114 | 28 | -0.42 | -0.17 |
| 12 | NT | NT | | |
| 13 | 117 | 17 | -0.17 | -0.11 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 119 | 7 |
| Spike | 134 | 4 |
| Homogeneity Value | 120 | 20 |
| Robust Average | 119 | 7 |
| Median | 119 | 2 |
| Mean | 118 | |
| N | 9 | |
| Max. | 130 | |
| Min. | 97 | |
| Robust SD | 8 | |
| Robust CV | 6.7% | |






Figure 34

Table 38

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Sb |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 310 | 40 | 0.03 | 0.02 |
| 4 | NT | NT | | |
| 6 | 310 | 50 | 0.03 | 0.02 |
| 7 | 320 | 64 | 0.36 | 0.17 |
| 8 | 289 | 12.4 | -0.65 | -0.99 |
| 9 | 308 | 52 | -0.03 | -0.02 |
| 10 | 288 | 64 | -0.68 | -0.32 |
| 11 | 340 | 170 | 1.00 | 0.18 |
| 12 | NT | NT | | |
| 13 | 313 | 47 | 0.13 | 0.08 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 309 | 16 |
| Spike | 314 | 9 |
| Homogeneity Value | 300 | 60 |
| Robust Average | 309 | 16 |
| Median | 310 | 8 |
| Mean | 310 | |
| N | 8 | |
| Max. | 340 | |
| Min. | 288 | |
| Robust SD | 18 | |
| Robust CV | 5.8% | |






Figure 35

Table 39

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Se |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 143 | NR | 0.00 | 0.00 |
| 2 | NT | NT | | |
| 3 | 140 | 15 | -0.21 | -0.19 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 150 | 30 | 0.49 | 0.23 |
| 8 | 145 | 6.58 | 0.14 | 0.26 |
| 9 | 140 | 31 | -0.21 | -0.10 |
| 10 | 137 | 34 | -0.42 | -0.18 |
| 11 | 146 | 68 | 0.21 | 0.04 |
| 12 | NT | NT | | |
| 13 | 146 | 22 | 0.21 | 0.13 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 143 | 4 |
| Spike | 153 | 4 |
| Homogeneity Value | 160 | 30 |
| Robust Average | 143 | 4 |
| Median | 144 | 4 |
| Mean | 143 | |
| N | 8 | |
| Max. | 150 | |
| Min. | 137 | |
| Robust SD | 5 | |
| Robust CV | 3.5% | |






Figure 36

Table 40

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Sn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 110 | 14 | -0.09 | -0.07 |
| 4 | NT | NT | | |
| 6 | 110 | 20 | -0.09 | -0.05 |
| 7 | 110 | 22 | -0.09 | -0.04 |
| 8 | 108 | 4.02 | -0.27 | -0.47 |
| 9 | 112 | 25 | 0.09 | 0.04 |
| 10 | 99.7 | 24.5 | -1.02 | -0.45 |
| 11 | 117 | 36 | 0.54 | 0.17 |
| 12 | NT | NT | | |
| 13 | 118 | 18 | 0.63 | 0.37 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 111 | 5 |
| Spike | 121 | 3 |
| Homogeneity Value | 110 | 20 |
| Robust Average | 111 | 5 |
| Median | 110 | 2 |
| Mean | 111 | |
| N | 8 | |
| Max. | 118 | |
| Min. | 99.7 | |
| Robust SD | 6 | |
| Robust CV | 5.4% | |






Figure 37

Table 41

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | TI |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 98 | 6 | -0.58 | -0.55 |
| 4 | NT | NT | | |
| 6 | 120 | 20 | 1.54 | 0.73 |
| 7 | 110 | 22 | 0.58 | 0.25 |
| 8 | 106 | 4.02 | 0.19 | 0.20 |
| 9 | 109 | 19 | 0.48 | 0.24 |
| 10 | 103 | 24 | -0.10 | -0.04 |
| 11 | 100 | 40 | -0.38 | -0.10 |
| 12 | NT | NT | | |
| 13 | 76.5 | 11 | -2.64 | -1.93 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 104 | 9 |
| Spike | 121 | 3 |
| Homogeneity Value | 110 | 20 |
| Robust Average | 104 | 9 |
| Median | 105 | 6 |
| Mean | 103 | |
| N | 8 | |
| Max. | 120 | |
| Min. | 76.5 | |
| Robust SD | 11 | |
| Robust CV | 11% | |






Figure 38

Table 42

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | U |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 120 | 6 | -0.40 | -0.46 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 140 | 28 | 1.20 | 0.51 |
| 8 | 126 | 4.45 | 0.08 | 0.10 |
| 9 | 121 | 19 | -0.32 | -0.19 |
| 10 | 111 | 25 | -1.12 | -0.53 |
| 11 | 126 | 13 | 0.08 | 0.06 |
| 12 | NT | NT | | |
| 13 | 129 | 19 | 0.32 | 0.19 |

Statistics

| | | |
|-----------------------|-----|---|
| Assigned Value | 125 | 9 |
| Spike | 138 | 4 |
| Robust Average | 125 | 9 |
| Median | 126 | 7 |
| Mean | 125 | |
| N | 7 | |
| Max. | 140 | |
| Min. | 111 | |
| Robust SD | 10 | |
| Robust CV | 8% | |






Figure 39

Table 43

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | V |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | NR | NR | | |
| 2 | NT | NT | | |
| 3 | 130 | 15 | -1.22 | -0.94 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 170 | 34 | 1.49 | 0.61 |
| 8 | 143 | 5.21 | -0.34 | -0.38 |
| 9 | 146 | 26 | -0.14 | -0.07 |
| 10 | 148 | 42 | 0.00 | 0.00 |
| 11 | 152 | 80 | 0.27 | 0.05 |
| 12 | NT | NT | | |
| 13 | 152 | 23 | 0.27 | 0.15 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 148 | 12 |
| Spike | 153 | 4 |
| Homogeneity Value | 156 | 31 |
| Robust Average | 148 | 12 |
| Median | 148 | 5 |
| Mean | 149 | |
| N | 7 | |
| Max. | 170 | |
| Min. | 130 | |
| Robust SD | 13 | |
| Robust CV | 8.8% | |






Figure 40

Table 44

Sample Details

| | |
|-------------------|-----------|
| Sample No. | S2 |
| Matrix. | Sea Water |
| Analyte. | Zn |
| Units | µg/L |

Participant Results

| Lab Code | Result | Uncertainty | z-Score | E_n-Score |
|-----------------|---------------|--------------------|----------------|----------------------------|
| 1 | 285 | NR | 0.18 | 0.50 |
| 2 | NT | NT | | |
| 3 | 290 | 36 | 0.36 | 0.27 |
| 4 | NT | NT | | |
| 6 | NT | NT | | |
| 7 | 270 | 54 | -0.36 | -0.18 |
| 8 | 286 | 10.3 | 0.21 | 0.42 |
| 9 | 286 | 80 | 0.21 | 0.07 |
| 10 | 271 | 66 | -0.32 | -0.13 |
| 11 | 286 | 46 | 0.21 | 0.13 |
| 12 | NT | NT | | |
| 13 | 258 | 39 | -0.79 | -0.55 |

Statistics

| | | |
|--------------------------|------|----|
| Assigned Value | 280 | 10 |
| Spike | 305 | 9 |
| Homogeneity Value | 290 | 60 |
| Robust Average | 280 | 10 |
| Median | 286 | 3 |
| Mean | 279 | |
| N | 8 | |
| Max. | 290 | |
| Min. | 258 | |
| Robust SD | 12 | |
| Robust CV | 4.3% | |






Figure 41

7 DISCUSSION OF RESULTS

7.1 Assigned Value

Sample S1 was filtered seawater while **Sample S2** was the same seawater unfiltered. A known amount of single element standard solutions was added to the two study samples.

Assigned Values were the robust average of participants' results. The robust averages and their associated expanded uncertainties were calculated using the procedure described in 'ISO13528:2015(E), Statistical methods for use in proficiency testing by interlaboratory comparisons'. Results less than 50% and more than 150% of the robust average were removed before calculation of each assigned value.⁶ Appendix 2 sets out the calculation for the robust average of V in Sample S1 and its associated uncertainty.

Spike Value includes both the incurred and the fortified value with the exception of U in S1.

Assigned values, spike values and homogeneity values were in agreement with each other within their estimates of uncertainty for all elements of interest with the exception of spike value for U in S1.

Traceability The consensus of participants' results (robust average) is not traceable to any external reference. So although expressed in SI units, the metrological traceability of the assigned value has not been established.

7.2 Measurement Uncertainty Reported by Participants

Participants were asked to report an estimate of the expanded measurement uncertainty associated with their results. Of 347 numerical results, 330 (95%) were reported with an expanded measurement uncertainty, indicating that the majority of laboratories have addressed this requirement of ISO 17025.⁸ The participants used a wide variety of procedures to estimate the expanded measurement uncertainty. These are presented in Table 3.

Approaches to estimating measurement uncertainty include: standard deviation of replicate analysis, Horwitz formula, professional judgement, bottom up approach, top down approach using precision and estimates of method and laboratory bias, and top down approach using only the reproducibility from inter-laboratory comparisons studies.^{9 – 15}

Proficiency tests allow a check of the reasonableness of uncertainty estimates. Results and the expanded MU are presented in the bar charts for each analyte (Figure 2 to 41). In this study, the reported expanded measurement uncertainty has been over-estimated in some cases (e.g. Labs 13, 6, 7 for Ag in S1, Lab 3 for Be in S1 or Lab 2 for Cd in S1) or under-estimated (e.g. Lab 8 for P in S1 and Al in S2 or Lab 9 for Al in S1). As a simple rule of thumb, when the uncertainty estimate is smaller than the assigned uncertainty value or larger than the uncertainty of the assigned value plus twice the target standard deviation then this should be viewed as suspect.

Stable control samples that cover the whole analytical process (including extraction) and **have a matrix similar** to the samples; **or**

- Stable control samples **and** duplicate analyses if control samples do not cover whole analytical process (e.g. the control sample is a synthetic sample- we have to take into consideration uncertainties arising from different matrices); **or**
- When control samples are not stable, from analysis of natural duplicates (gives within-day variation for sampling and measurement) and long-term uncertainty component from the variation in the instrument calibration; **or**

- Replicate analyses performed on the same sample at different times to obtain estimates of intermediate precision; within-batch replication provides estimates of repeatability only.

The most common sources for estimating the method bias component for the measurement uncertainty calculation are from:

- Certified reference material recoveries; **or**
- Participation in PT studies (laboratory bias from at least 6 successful PT studies); **or**
- From sample spike recoveries.

When a laboratory has successfully participated in at least 6 proficiency testing studies, the standard deviation from proficiency testing studies only, can also be used to estimate the uncertainty of their measurement results.^{10, 12} An example of estimating measurement uncertainty using proficiency testing data only is given in Appendix 3.

Some laboratories attached estimates of the expanded measurement uncertainty to results reported as less than their limit of detection. An estimate of uncertainty expressed as a value cannot be attached to a result expressed as a range.

In some cases the results were reported with an inappropriate number of significant figures. The recommended format is to write uncertainty to no more than two significant figures and then to write the result with the corresponding number of decimal places. For example, instead of $16.2 \pm 1.701 \mu\text{g/L}$, it is better to report $16.2 \pm 1.7 \mu\text{g/L}$ or instead of $10.75 \pm 1.20 \mu\text{g/L}$, it is better to report $10.8 \pm 1.2 \mu\text{g/L}$.⁹

7.3 E_n-score

E_n-score should be interpreted only in conjunction with z-scores. The E_n-score indicates how closely a result agrees with the assigned value taking into account the respective uncertainties. An unsatisfactory E_n score for an analyte can either be caused by an inappropriate measurement, an inappropriate estimation of measurement uncertainty, or both.

The dispersal of participants' E_n-scores is graphically presented in Figure 42. Where a laboratory did not report an expanded uncertainty with a result, an expanded uncertainty of zero (0) was used to calculate the E_n-score.

Of 347 results for which E_n-scores were calculated, 315 (91%) returned a satisfactory score of $|E_n| \leq 1.0$ indicating agreement of the participants' results with the assigned values within their respective expanded measurement uncertainties.

7.4 z-Score


The z-score compares the participant's deviation from the assigned value with the target standard deviation set for proficiency assessment.

The target standard deviation defines satisfactory performance in a proficiency test. Target standard deviations equivalent to 10% to 20% PCV were used to calculate z-scores. Unlike the standard deviation based on between laboratories CV, setting the target standard deviation as a realistic, set value enables z-scores to be used as fixed reference value points for assessment of laboratory performance, independent of group performance.

The between laboratory coefficient of variation predicted by the Thompson equation⁷ and the participants' coefficient of variation resulted in this study are presented for comparison in Table 45.


The dispersal of participants' z-scores is presented in Figure 43 (by laboratory code) and in Figure 44 (by test). Of 347 results for which z-scores were calculated, 328 (95%) returned a satisfactory score of $|z| \leq 2.0$ and 10 (3%) were questionable of $2.0 < |z| < 3.0$. Participants with multiple z-scores larger than 2 or smaller than -2 should check for laboratory bias.

Laboratories 1, 4, 6, 7, and 12 returned satisfactory z-scores for all analytes reported.



Scores of >10 or <-10 have been plotted as 10 or -10.

Figure 42 E_n -Score Dispersal by Laboratory




Scores of >10 or <-10 have been plotted as 10 or -10.

Figure 43 z-Score Dispersal by Laboratory

Table 45 Between Laboratory CV of this Study, Thompson CV and Set Target CV

| Sample | Analyte | Assigned value | Between Laboratory CV* | Thompson/Horwitz CV | Target SD (as PCV) |
|--------|---------|----------------|------------------------|---------------------|--------------------|
| S1 | Ag | 5.95 µg/L | 3% | 22% | 15% |
| S1 | Al | 15.7 µg/L | 22% | 22% | 15% |
| S1 | As | 1.94 µg/L | 10% | 22% | 15% |
| S1 | Be | 1.31 µg/L | 19% | 22% | 15% |
| S1 | Cd | 1.09 µg/L | 10% | 22% | 15% |
| S1 | Co | 1.21 µg/L | 13% | 22% | 15% |
| S1 | Cr | 2.36 µg/L | 21% | 22% | 15% |
| S1 | Cu | 12.7 µg/L | 7.9% | 22% | 15% |
| S1 | Fe | 15.9 µg/L | 17% | 22% | 15% |
| S1 | Hg | 0.175 µg/L | 21% | 22% | 15% |
| S1 | Mn | 3.69 µg/L | 9% | 22% | 15% |
| S1 | Ni | 1.68 µg/L | 25% | 22% | 15% |
| S1 | P | 140 µg/L | 36% | 22% | 20% |
| S1 | Pb | 0.995 µg/L | 6.1% | 22% | 15% |
| S1 | Se | 3.15 µg/L | 16% | 22% | 15% |
| S1 | Sn | 5.88 µg/L | 7.8% | 22% | 15% |
| S1 | Tl | 2.95 µg/L | 7.6% | 22% | 15% |
| S1 | U | 5.58 µg/L | 7% | 22% | 15% |
| S1 | V | 4.65 µg/L | 8.9% | 22% | 15% |
| S1 | Zn | 6.54 µg/L | 16% | 22% | 15% |
| S2 | Ag | 114 µg/L | 4.3% | 22% | 10% |
| S2 | Al | 780 µg/L | 16% | 17% | 10% |
| S2 | As | 247 µg/L | 6.2% | 20% | 10% |
| S2 | Ba | 344 µg/L | 5% | 19% | 10% |
| S2 | Cd | 131 µg/L | 5.9% | 22% | 10% |
| S2 | Cr | 180 µg/L | 4.6% | 21% | 10% |
| S2 | Cu | 293 µg/L | 5.7% | 19% | 10% |
| S2 | Fe | 329 µg/L | 4.6% | 19% | 10% |
| S2 | Mn | 294 µg/L | 4.2% | 19% | 10% |
| S2 | Mo | 144 µg/L | 6.2% | 21% | 10% |
| S2 | Ni | 141 µg/L | 6% | 21% | 10% |
| S2 | P | 580 µg/L | 20% | 17% | 10% |
| S2 | Pb | 119 µg/L | 6.8% | 22% | 10% |
| S2 | Sb | 309 µg/L | 5.7% | 19% | 10% |
| S2 | Se | 143 µg/L | 3.3% | 21% | 10% |
| S2 | Sn | 111 µg/L | 5% | 22% | 10% |
| S2 | Tl | 104 µg/L | 10% | 22% | 10% |
| S2 | U | 125 µg/L | 8% | 22% | 10% |
| S2 | V | 148 µg/L | 8.6% | 21% | 10% |
| S2 | Zn | 280 µg/L | 4.1% | 19% | 10% |

* Robust between Laboratories CV with outliers removed



Scores of >10 or <-10 have been plotted as 10 or -10.

Figure 44 z-Score Dispersal by Analyte

Table 46 Summary of Participants' Results and Performance for Total and Dissolved Elements

| Lab Code | S1 - Al µg/L | S2 - Al µg/L | S1 - Ag µg/L | S2 - Ag µg/L | S1 - As µg/L | S2 - As µg/L | S2 - Ba µg/L | S1 - Be µg/L | S1 - Cd µg/L | S2 - Cd µg/L | S1 - Co µg/L | S1 - Cr µg/L | S2 - Cr µg/L | S1 - Cu µg/L | S2 - Cu µg/L |
|----------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|
| AV | 15.7 | 780 | 5.95 | 114 | 1.94 | 247 | 344 | 1.31 | 1.09 | 131 | 1.21 | 2.36 | 180 | 12.7 | 293 |
| HV | 15.5 | 840 | 6.4 | 115 | 2.25 | 270 | 350 | 1.20 | 1.08 | 130 | 1.21 | 2.69 | 190 | 11.8 | 310 |
| SV | 12.2 | 853 | 6.08 | 122 | 2.34 | 258 | 367 | 1.30 | 1.11 | 141 | 1.19 | 2.21 | 186 | 10.9 | 314 |
| 1 | NT | NR | NT | NR | NT | 226 | NR | NT | NT | 136 | NT | NT | 173 | NT | 286 |
| 2 | 19 | NT | 1.2 | NT | 2 | NT | NT | 1 | 1 | NT | 1 | 3 | NT | 13 | NT |
| 3 | 7.4 | 660 | 6 | 110 | <10 | 260 | 360 | 1.1 | <1 | 120 | 1.4 | <5 | 200 | 12 | 280 |
| 4 | 14 | NT | 6.1 | NT | 1.8 | NT | NT | 1.6 | 1.1 | NT | 1.2 | 2 | NT | 13 | NT |
| 6 | NT | NT | 6.1 | 120 | NT | NT | 340 | 1.6 | 1.3 | 130 | NT | NT | 180 | NT | NT |
| 7 | 18 | 820 | 6.1 | 120 | 2.2 | 250 | 340 | 1.4 | 1 | 140 | 1.4 | 2.6 | 190 | 13 | 300 |
| 8 | 13.3 | 811 | 6.03 | 108 | 2.1 | 239 | 312 | 1.52 | 1.05 | 127 | 1.24 | 1.38 | 180 | 12.2 | 296 |
| 9 | 12.2 | 654 | 6 | 113.8 | 1.8 | 247 | 351 | 1.3 | 1.2 | 139 | 1.2 | 2.5 | 176 | 1.2 | 296 |
| 10 | 16.1 | 937 | 5.14 | 116 | 1.68 | 239 | 342 | 1.3 | 1.03 | 127 | 1.09 | 2.28 | 173 | 12.1 | 314 |
| 11 | 20 | 790 | 5.9 | 115 | <4 | 312 | 336 | NT | 0.97 | 125 | 1.2 | 2.88 | 184 | 13.6 | 307 |
| 12 | 13 | NT | 4.7 | NT | 2 | NT | NT | 1.3 | 1.2 | NT | 1.3 | 2.1 | NT | 14 | NT |
| 13 | NT | NT | 5.92 | 112 | 1.9 | 247 | 368 | 1.02 | 1.12 | 133 | 1.06 | 2.28 | 174 | 10.8 | 252 |

Shaded cells are results which returned a questionable or unsatisfactory z-score. A.V. = Assigned Value, H.V. = Homogeneity Value, S.V. = Spike Value

Table 46 Summary of Participants' Results and Performance for Total and Dissolved Elements (continued)

| Lab Code | S1 - Fe µg/L | S2 - Fe µg/L | S1 - Hg µg/L | S1 - Mn µg/L | S2 - Mn µg/L | S2 - Mo µg/L | S1 - Ni µg/L | S2 - Ni µg/L | S1 - P µg/L | S2 - P µg/L | S1 - Pb µg/L | S2 - Pb µg/L | S2 - Sb µg/L | S1 - Se µg/L | S2 - Se µg/L |
|----------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|----------------|----------------|-----------------|-----------------|-----------------|-----------------|-----------------|
| AV | 15.9 | 329 | 0.175 | 3.69 | 294 | 144 | 1.68 | 141 | 140 | 580 | 0.995 | 119 | 309 | 3.15 | 143 |
| HV | 14.7 | 379 | 0.192 | 3.85 | 300 | 150 | 1.69 | 150 | 140 | 500 | 1.04 | 120 | 300 | 4.42 | 160 |
| SV | 15.1 | 363 | 0.201 | 3.44 | 309 | 150 | 1.68 | 154 | 142 | 598 | 0.979 | 134 | 314 | 3.01 | 153 |
| 1 | NT | 333 | NT | NT | NR | NR | NT | NR | NT | 604 | NT | 130 | NR | NT | 143 |
| 2 | 16 | NT | 0.34 | 3 | NT | NT | 2 | NT | 195 | NT | <1 | NT | NT | 4 | NT |
| 3 | 14 | 310 | 0.14 | 4.9 | 320 | 150 | <10 | 140 | 105 | 500 | <1 | 97 | 310 | 3.2 | 140 |
| 4 | 14 | NT | 0.2 | 3.5 | NT | NT | 1.6 | NT | NR | NT | 1 | NT | NT | 3 | NT |
| 6 | NT | NT | 0.16 | NT | 290 | 140 | NT | NT | NT | NT | <1 | 120 | 310 | NT | NT |
| 7 | 18 | 330 | 0.22 | 3.9 | 290 | 150 | 1.8 | 160 | 140 | 560 | <1 | 130 | 320 | 3.7 | 150 |
| 8 | 11 | 331 | 0.13 | 3.6 | 290 | 132 | 1.05 | 142 | 76.8 | 392 | 1.02 | 119 | 289 | 3.17 | 145 |
| 9 | 15 | 339 | 0.17 | 4 | 307 | 148 | 1.5 | 142 | 119 | 548 | 0.1 | 120 | 308 | 3 | 140 |
| 10 | 22.6 | 314 | 0.2 | 3.78 | 280 | 133 | 1.35 | 143 | 194 | 630 | 0.99 | 118 | 288 | 1.85 | 137 |
| 11 | 15.4 | 360 | 0.209 | 3.72 | 295 | 150 | <7 | 136 | NT | 850 | 1.08 | 114 | 340 | <4 | 146 |
| 12 | 18 | NT | <0.5 | 3.4 | NT | NT | 2.1 | NT | NT | NT | 0.94 | NT | NT | 2.9 | NT |
| 13 | 17.1 | 324 | 0.148 | 3.65 | 285 | 149 | 2.02 | 129 | 150 | 625 | 0.937 | 117 | 313 | 3.11 | 146 |

Shaded cells are results which returned a questionable or unsatisfactory z-score. A.V. = Assigned Value, H.V.= Homogeneity Value, S.V. = Spike Value

Table 46 Summary of Participants' Results and Performance for Total and Dissolved Elements (continued)

| Lab Code | S1 - Sn µg/L | S2 - Sn µg/L | S1 - Tl µg/L | S2 - Tl µg/L | S1 - U µg/L | S2 - U µg/L | S1 - V µg/L | S2 - V µg/L | S1 - Zn µg/L | S2 - Zn µg/L |
|----------|-----------------|-----------------|-----------------|-----------------|----------------|----------------|----------------|----------------|-----------------|-----------------|
| AV | 5.88 | 111 | 2.95 | 104 | 5.58 | 125 | 4.65 | 148 | 6.54 | 280 |
| HV | 6.1 | 110 | 2.94 | 110 | NA | NA | 4.17 | 156 | 7.6 | 290 |
| SV | 5.98 | 121 | 3.03 | 121 | 2.50 | 138 | 5.21 | 153 | 7.59 | 305 |
| 1 | NT | NR | NT | NR | NT | NR | NT | NR | NT | 285 |
| 2 | 6 | NT | 3 | NT | 5 | NT | 5 | NT | 17 | NT |
| 3 | 6 | 110 | 2.5 | 98 | 5.3 | 120 | 2.5 | 130 | 6.4 | 290 |
| 4 | 6 | NT | 3.2 | NT | 6.1 | NT | 4.7 | NT | 7 | NT |
| 6 | 5.4 | 110 | 3.1 | 120 | NT | NT | NT | NT | NT | NT |
| 7 | 7.2 | 110 | 3 | 110 | 5.6 | 140 | 4.8 | 170 | 6.4 | 270 |
| 8 | 5.39 | 108 | 3.05 | 106 | 5.65 | 126 | 4.49 | 143 | 5.07 | 286 |
| 9 | 6 | 112 | 2.9 | 109 | 5.3 | 121 | 4.5 | 146 | 7.4 | 286 |
| 10 | 5.34 | 99.7 | 3.13 | 103 | 6.05 | 111 | 4.14 | 148 | 5.4 | 271 |
| 11 | 6.3 | 117 | 2.8 | 100 | 5.4 | 126 | 5 | 152 | 6.9 | 286 |
| 12 | NT | NT | 3 | NT | 5.8 | NT | 4.7 | NT | 8.1 | NT |
| 13 | 5.76 | 118 | 2.14 | 76.5 | 5.59 | 129 | 5.12 | 152 | 6.18 | 258 |

Shaded cells are results which returned a questionable or unsatisfactory z-score. A.V. = Assigned Value, H.V.= Homogeneity Value, NA = Not Available, S.V. = Spike Value

7.5 Participants' Results and Analytical Methods for Total and Dissolved Elements

Sample S1 was filtered seawater. The analytes' concentration in this sample ranged from 0.175 to 140 µg/L. **Sample S2** was unfiltered seawater and participants were asked to report results for total elements. Analytes' concentration in this sample was 2 to 7 times higher than in Sample S1. A summary of participants' results and performance in the two study samples is presented in Table 46 and in Figures 43 and 44.

Low level Al, Ni and P in S1 were the tests which had the highest coefficient of variation, ranging from 22% to 36%.

Individual Element Commentary

Participants were requested to analyse the samples using their normal test method and to report a single result as they would normally report to a client. With the exception of one participant, all who reported results for total elements in S2 performed digestion. Most used a digestion temperature of 85°C to 100°C and two digested their samples at 170°C. Six laboratories used nitric acid and hydrochloric acid for extraction and two used only nitric acid. Laboratory 11 reported using only hydrochloric acid for Ag extraction. No relationship was evident between the results reported for total elements in S2 and the digestion method employed.

Instrumental measurement was one of the main factors that influenced the results for total and dissolved elements in the seawater samples. However, participants' performance does not reflect only instrumental performance, but also the performance of the analyst and of the analytical method used by the testing laboratory. Thus, these results should not be construed as an evaluation of a particular instrument.

Participants used a wide variety of instrumental techniques, collision/reaction cells and cell gases. Most laboratories reported using ICP-MS with a collision/reaction cell; some used ICP-OES, and some only ICP-MS. One participant used ICP-MS measurements with a preconcentration and matrix separation step (seaFAST-ICP-MS) and one reported using ICP-MS/MS in collision mode or reaction mode with a H₂ and N₂O mix. Plots of participants' results and performance versus instrumental techniques used are presented in Figure 45.

Aluminium level in S1 was low at 15.7 µg/L and this posed significant problems for laboratories. The between laboratory CV for Al in S1 was 22%.

Laboratory 9 should check the method used for Al measurements as their reported results for S1 and S2 were both low, with z-scores close to -2, which is an indication of laboratory and/or method bias.

Arsenic measurements at low levels posed no significant problems for laboratories. The between-laboratory CV for As in S1 was 10%. The unsatisfactory z-score is high, likely indicating unsolved interference problems (Figure 45).

Ammonia as reaction gas is not effective in removing $^{40}\text{Ar}^{35}\text{Cl}^+$ interferences on As. ICP-MS with H₂ or O₂ as reaction gas may be a better option. When testing As in seawater, the largest and most common interference to overcome in analysis by ICP-MS is $^{40}\text{Ar}^{35}\text{Cl}^+$. Hydrogen as reaction gas has been proven to reduce Ar-based interferences while the mass shifting of $^{75}\text{As}^+$ to m/z 91 as $^{75}\text{AsO}^{16+}$, by O₂ is also considered an effective solution for overcoming As interferences in seawater.¹⁶





Figure 45 Participants' Results and Performance vs Instrumental Technique



*Scores of <-3 have been plotted as -3.

Figure 45 Participants' Results and Performance vs Instrumental Technique (continued)












Figure 45 Participants' Results and Performance vs Instrumental Technique (continued)



Scores of >3 or <-3 have been plotted as 3 or -3

Figure 45 Participants' Results and Performance vs Instrumental Technique (continued)



Score <-5 has been plotted as -3.

Figure 45 Participants' Results and Performance vs Instrumental Technique (continued)




Figure 45 Participants' Results and Performance vs Instrumental Technique (continued)

Chromium and Nickel measurements at the low level in S1 challenged participating laboratories: the between laboratories CV was high, at 21% and 25% respectively.

Participants used a wide variety of instrumental techniques to overcome the interference problems with Cr in the seawater sample while with the exception of one all laboratories used ICP-MS in collision mode for Ni measurements (Figure 45).

Iron and Zinc The coefficient of variation in S1 for Fe and Zn was 16% and 17% respectively. These elements are known to be ubiquitous in the environment and controlling Fe or Zn contamination is a challenge for laboratories. No relationship between the results reported for these elements and the instrumental technique used was evident.

Mercury Use of expired standards or those not prepared fresh from standard stock solutions before measurement could be a cause of high Hg results.

Phosphorus assigned value in S1 was 140 µg/L and the between laboratory coefficient of variation was high at 36%. Most laboratories used ICP-OES with a wavelength of 213.618 nm or 177.4 nm for P measurements in S1. One laboratory reported a satisfactory P result from ICP-MS/MS measurements in reaction mode using a mixture of H₂ and N₂O as reaction gases (Figure 45).

Selenium Participants reported using 6 different instrumental techniques: ICP-MS in collision, reaction or MS/MS mode and with various collision/reaction gases: He, HEHe, NH₃, H₂+N₂O and H₂. With the exception of one, all results reported for Se were in good agreement with each other and with the assigned value.

Vanadium level in S1 was low at 4.65 µg/L, and ICP-OES might not be the right technique for V measurements below 5 ppb level in the solution.

Comparisons of Participants' Performance in Samples S1 and S2

Plots of participants' performance in Samples S1 and S2 are presented in Figure 46. Plots of z-scores that are close to the zero axes are an indication of excellent accuracy.

Laboratories whose z-scores for an element in both samples S1 and S2 lie on the same side of the centre line may need to monitor their procedure as may be an indication of method bias.




Figure 46 Comparisons of Participants' Performance in S1 and S2





Figure 46 Comparisons of Participants' Performance in S1 and S2 (continued)



Scores of <-3 have been plotted as -3.

Figure 46 Comparisons of Participants' Performance in S1 and S2 (continued)




Figure 46 Comparisons of Participants' Performance in S1 and S2 (continued)

7.6 Comparison with Previous NMI Proficiency Tests of Metals in Water

AQA 19-16 is the 24th NMI proficiency test of metals in water and the 18th study in seawater.

Participants' performance in measurement of metals in seawater over last ten years is presented in Figure 47. Despite differences in the analytes' concentrations, on average participants' performance has improved over time.

Individual performance history reports are emailed to each participant at the end of the study; the consideration of z-scores for an analyte over time provides much more useful information than a single z-score.

Over time, laboratories should expect at least 95% of their scores to lie within the range $|z| \leq 2$. Scores in the range $2 < |z| < 3$ can occasionally occur, however these should be interpreted in conjunction with the other scores obtained by that laboratory. For example, a trend of z-scores on one side of the zero line is an indication of method or laboratory bias.

7.7 Reference Materials and Certified Reference Materials

Participants reported whether control samples (spiked samples, certified reference materials-CRMs or matrix specific reference materials-RMs) had been used (Table 47).

Table 47 Control Samples Used by Participants

| Lab. Code | RMs or CRMs identity |
|-----------|--|
| 3 | NMIA MX014 |
| 7 | MX014 |
| 9 | TM-25.5, TM-26.3, TMDA-52.4, TMDA-52.3, CASS-6 |
| 10 | HPS (CRMSW), saline water CRM |
| 11 | Nass-7 Seawater Certified Reference Material for Trace Metals and other constituents Cass-6 Nearshore Seawater certified Reference Material for Trace Metals and other constituents |
| 12 | NASS 7, CASS 6, NMI MX014 |
| 13 | SS |

A certified reference material for trace elements in seawater (MX014) is available from NMI.




Figure 47 Participants' Performance in Metals in Seawater PT Studies over Last Ten Years

8 REFERENCES

- [1] ISO17043:2010, Conformity assessment – *General requirements for proficiency testing*.
- [2] NMI 2019, *NMI Chemical Proficiency Testing Study Protocol*, viewed 7 January 2020, <<https://www.industry.gov.au/client-services/chemical-and-biological-measurement-services/proficiency-testing-services>>.
- [3] NMI 2019, *NMI Chemical Proficiency Testing Statistical Manual*, viewed 7 January 2020, <<https://www.industry.gov.au/client-services/chemical-and-biological-measurement-services/proficiency-testing-services>>.
- [4] Thompson, M, Ellison, S & Wood, R 2006, 'The international harmonized protocol for proficiency testing of (chemical) analytical laboratories', *Pure Appl. Chem*, vol 78, pp 145-196.
- [5] National Environmental Protection Council, Schedule B(1) – *Guidelines on the Investigation Levels for Soil and Groundwater*, viewed 7 January 2020, <<http://www.nepc.gov.au/nepms>>.
- [6] ISO13528:2015(E), *Statistical methods for use in proficiency testing by interlaboratory comparisons*.
- [7] Thompson, M 2000, Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing, *Analyst*, vol 125, pp 385-386.
- [8] ISO/IEC 17025:2018, *General requirements for the competence of testing and calibration laboratories*
- [9] Eurachem 2012, *Quantifying uncertainty in Analytical Measurement*, 3rd edition, viewed 10 May 2017, <http://www.eurachem.org/images/stories/Guides/pdf/QUAM2012_P1.pdf>.
- [10] Betil, M, Naykki, T, Hovind, H & Krysell, M 2004, *Nordtest Report Handbook for Calculation of Measurement Uncertainty in Environmental Laboratories*, Nordest Tekniikantie, Finland, Esopo.
- [11] Hibbert, B 2007, *Quality Assurance for the Analytical Chemistry Laboratory*, Oxford University Press.
- [12] NATA 2018, *General Accreditation Guidance Estimating and Reporting Measurement Uncertainty of Chemical Test Results*
- [13] ISO (2008), *Guide to the Expression of Uncertainty in Measurement (GUM)*, Geneva, Switzerland.
- [14] Eurolab 2002, Technical Report No 1/2002 - *Measurement Uncertainty in Testing*.
- [15] NMI, *Estimating Measurement Uncertainty for Chemists* – viewed 7 January 2020, <www.measurement.gov.au>.
- [16] Jackson, B & Liba, A 2015, 'Advantages of reaction cell ICP-MS on doubly charged interferences for arsenic and selenium analysis in foods' *J. Anal. At. Spectrometry*, vol. 30, pp1179-1183.

APPENDIX 1 – SAMPLE PREPARATION, ANALYSIS AND HOMOGENEITY TESTING

Sample Preparation

Sample S1 was prepared from seawater. Approximately 10 L of seawater from Sydney harbour was filtered through a 0.45 µm pore size filter, stabilised by adding 2% (v/w) nitric acid and further fortified with 20 elements.

Sample S2 was unfiltered seawater stabilised by adding 2% (v/w) nitric acid and further fortified for 20 elements.

Sample Analysis and Homogeneity Testing

With the exception of U in S1 and S2, a partial homogeneity test was conducted for all analytes of interest. Three bottles were analysed in duplicate and the average of the results was reported as the homogeneity value.

Sample Analysis for Dissolved and Total Elements

For analyses of total elements in Sample S2, a test portion of 30 mL was transferred to a 50 mL graduated polypropylene centrifuge tube. The samples were digested using 2 mL of nitric acid and 1 mL of hydrochloric acid on a hot block at 95±5°C for 120 min.

Testing involved measurements using ICP-MS or ICP-OES. The measurement instrument was calibrated using external standards for targeted analytes. A set of quality control samples consisting of blanks, a blank matrix spike, duplicates, sample matrix spikes and a certified reference material (MX014) was carried through the same set of procedures and analysed simultaneously with the samples. A summary of the ion/wavelength and instrument conditions used for each analyte is given in Table 48.

Table 48 Instrumental Technique used for Dissolved Elements

| Analyte | Instrument | Internal Standard | Reaction/ Collision Cell (if applicable) | Cell Mode/Gas (if applicable) | S1 Final Dilution Factor | S2 Final Dilution Factor | Ion/Wavelength |
|---------|------------|-------------------|--|-------------------------------------|--------------------------------|--------------------------------|------------------------------|
| Ag | ICP-MS | Rh | ORS | He | 10 | 10 | 107 m/z |
| Al | ICP-OES | Lu | NA | NA | 2 | 2 | 167.019 nm |
| As | ICP-MS | Rh | ORS | He | 10 | 10 | 75 m/z |
| Ba | ICP-MS | Rh | ORS | He | NA | 10 | 134 m/z |
| Be | ICP-MS | Rh | NA | NA | 10 | NA | 9 m/z |
| Cd | ICP-MS | Rh | ORS | He | 10 | 10 | 111 m/z |
| Co | ICP-MS | Rh | ORS | He | 10 | NA | 59 m/z |
| Cr | ICP-MS | Rh | ORS | He | 10 | 10 | 52 m/z |
| Cu | ICP-MS | Rh | ORS | He | 10 | 10 | 63 m/z |
| Fe | ICP-OES | Lu | NA | NA | 2 | 2 | 238.204 nm |
| Hg | ICP-MS | Rh | ORS | He | 10 | NA | 201 m/z |
| Mn | ICP-MS | Rh | ORS | He | 10 | 10 | 55 m/z |
| Mo | ICP-MS | Rh | ORS | He | NA | 10 | 60 m/z |
| Ni | ICP-MS | Rh | ORS | He | 10 | 10 | 60 m/z |
| P | ICP-MS | Rh | ORS | HEHe | 10 | 10 | 31 m/z |
| Pb | ICP-MS | Ir | NA | NA | 10 | 10 | Average of 206, 207, 208 m/z |
| Sb | ICP-MS | Rh | ORS | He | NA | 10 | 78 m/z |
| Se | ICP-MS | Rh | ORS | HEHe | 10 | 10 | 78 m/z |
| Sn | ICP-MS | Rh | NA | NA | 10 | 10 | 118 m/z |
| Tl | ICP-MS | Rh | ORS | He | 10 | 10 | 205m/z |
| V | ICP-MS | Rh | ORS | He | 10 | 10 | 51 m/z |
| Zn | ICP-MS | Rh | ORS | He | 10 | 10 | 66 m/z |

APPENDIX 2 – ASSIGNED VALUE, Z-SCORE AND E_N SCORE CALCULATION

The assigned value was calculated as the robust average using the procedure described in 'ISO13258:2015, Statistical methods for use in proficiency testing by interlaboratory comparisons – Annex C⁶ the uncertainty was estimated as:

$$u_{rob\ av} = 1.25 * S_{rob\ av} / \sqrt{p} \quad \text{Equation 4}$$

where:

$u_{rob\ av}$ robust average standard uncertainty
 $S_{rob\ av}$ robust average standard deviation
 p number of results

The expanded uncertainty ($U_{rob\ av}$) is the standard uncertainty multiplied by a coverage factor of 2 at approximately 95% confidence level.

A worked example is set out below in Table 49.

Table 49 Uncertainty of Assigned Value for V in Sample S1

| | |
|-----------------|-----------|
| No. results (p) | 10 |
| Robust Average | 4.65 µg/L |
| $S_{rob\ av}$ | 0.41 µg/L |
| $u_{rob\ av}$ | 0.17 µg/L |
| k | 2 |
| $U_{rob\ av}$ | 0.33 µg/L |

The assigned value for **V** in Sample S1 is **4.65 ± 0.33 µg/L**.

z-Score and E_N-score

For each participant's result a z-score and E_N-score are calculated according to Equation 2 and Equation 3 respectively (see page 11).

A worked example is set out below in Table 50.

Table 50 z-Score and E_N-score for V result reported by Laboratory 8 in S1

| V Result µg/L | Assigned Value µg/L | Set Target Standard Deviation | z-Score | E _N -Score |
|------------------|------------------------|---|---|---|
| 4.49±0.34 | 4.65±0.33 | 15% as CV or 0.15x4.65= =0.6975 µg/L | $z = \frac{(4.49 - 4.65)}{0.6975}$ $z = -0.23$ | $En = \frac{(4.49 - 4.65)}{\sqrt{0.34^2 + 0.33^2}}$ $En = -0.34$ |

APPENDIX 3 - USING PT DATA FOR UNCERTAINTY ESTIMATION

When a laboratory has successfully participated in at least 6 proficiency testing studies, the standard deviation from proficiency testing studies can also be used to estimate the uncertainty of their measurement results.^{10, 12} An example is given below.

Between 2007 and 2019, NMI carried out 24 proficiency tests for metals in water. These studies involved analyses of dissolved or total elements at low and high levels in potable, fresh (river), saline water, ground water and waste water. Laboratory X participated and submitted satisfactory Ni results in 12 of these PTs.

Table 51 Ni Results for Laboratory X From Proficiency Testing Studies of Metals in Water

| Study No. | Sample | Laboratory result* µg/L | Assigned value µg/L | Robust CV of all results (%) | Number of Results |
|-----------|--------------|----------------------------|------------------------|------------------------------|-------------------|
| AQA 08-02 | Fresh | 51 ± 7.2 | 52.0 ± 3.1 | 9.9 | 18 |
| AQA 08-10 | Fresh | 20 ± 3 | 18.9 ± 0.6 | 7.8 | 26 |
| | Fresh | 200 ± 20 | 191 ± 5 | 5.5 | 26 |
| AQA 09-05 | Saline | 5.0 ± 1.2 | 5.5 ± 0.6 | 13.3 | 14 |
| | Saline | 43 ± 5 | 44.7 ± 3.3 | 10.8 | 18 |
| AQA 09-18 | Fresh | 5.3 ± 0.5 | 5.04 ± 0.27 | 7.4 | 14 |
| | Fresh | 49 ± 4 | 48.9 ± 1.2 | 3.3 | 16 |
| AQA 10-06 | Potable | 49 ± 4 | 50 ± 1 | 5.9 | 20 |
| | Potable | 48 ± 4 | 50 ± 1 | 3 | 20 |
| AQA 11-17 | Waste water | 97 ± 9 | 99 ± 1 | 1.5 | 15 |
| | Waste water | 97 ± 9 | 98 ± 1 | 1.5 | 15 |
| AQA 12-09 | Fresh | 43 ± 6 | 45 ± 2 | 6.6 | 19 |
| | Fresh | 51 ± 7 | 53 ± 2 | 7.5 | 19 |
| AQA 12-20 | Sea water | 40 ± 4.4 | 38.4 ± 2.1 | 11 | 22 |
| AQA 13-09 | Fresh | 4.3 ± 0.5 | 4.09 ± 0.17 | 8.5 | 15 |
| | Fresh | 36 ± 4 | 36.1 ± 1.0 | 4.5 | 16 |
| AQA 14-08 | Ground water | 18.0 ± 2.0 | 19.1 ± 0.7 | 7.9 | 13 |
| AQA 15-18 | S2 | 32 | 33.6 | 4.9 | 6.2 |
| AQA 16-15 | S1 | 7.2 | 6.52 | 9.9 | 10 |
| | S2 | 46 | 40.9 | 11.7 | 13 |
| Average | | | | 7.3** | |

* Expanded uncertainty at approximately 95% confidence. ** The mean value of Robust CV was used.

Taking the average of the robust CV over these PT samples gives an estimate of the relative standard uncertainty of 7.3%. Using a coverage factor of 2 gives a relative expanded uncertainty of 15%, at a level of confidence of approximately 95%. Table 51 sets out the expanded uncertainty for results of the measurement of Ni in fresh, saline, waste or potable water over the range 5 – 200 µg/L.

Table 52 Uncertainty of Ni results estimated using PT data

| Results µg/L | Uncertainty µg/L |
|-----------------|---------------------|
| 5.0 | 0.75 |
| 20 | 3 |
| 50 | 7.5 |
| 200 | 30 |

The MU estimates made using PT data is close to Laboratory X's own uncertainty estimates reported with their PT results. The estimate of 15% passes the test of being reasonable, and the analysis of the 6 different matrices over 9 years can safely be assumed to include all the

relevant uncertainty components (different operators, reagents, calibrants etc), and so complies with ISO 17025.⁸

APPENDIX 4 - ACRONYMS AND ABBREVIATIONS

| | |
|--------------------|--|
| AAS | Atomic Absorption Spectrometry |
| CRI | Collision Reaction Interface |
| CRM | Certified Reference Material |
| CV | Coefficient of Variation |
| CVAAS | Cold Vapour Atomic Absorption Spectrometry |
| CVAFS | Cold Vapour Atomic Fluorescence Spectrometry |
| AFS | Atomic fluorescence spectroscopy |
| DRC | Dynamic Reaction Cell |
| FIA | Flow Injection Analyser |
| GFAAS | Graphite Furnace Atomic Absorption Spectrometry |
| HEHe | High energy He mode |
| ICP-OES-AV | Inductively Coupled Plasma – Optical Emission Spectrometry- axial view |
| ICP-OES-RV | Inductively Coupled Plasma – Optical Emission Spectrometry- radial view |
| Max | Maximum value in a set of results |
| Md | Median |
| Min | Minimum value in a set of results |
| NMI | National Measurement Institute (of Australia) |
| NR | Not Reported |
| NT | Not Tested |
| ORS | Octopole Reaction System |
| PT | Proficiency Test |
| ICP-MS | Inductively Coupled Plasma – Mass Spectrometry |
| RM | Reference Material |
| Robust CV | Robust Coefficient of Variation |
| Robust SD | Robust Standard Deviation |
| S | Spiked or formulated concentration of a PT sample |
| SS | Spiked sample |
| seaFAST-ICP-MS | Automated inductively coupled plasma spectrometry preconcentration system for undiluted seawater |
| SI | The International System of Units |
| s^2_{sam} | Sampling variance |
| s_a/σ | Analytical standard deviation divided by the target standard deviation |
| SRM | Standard Reference Material (Trademark of NIST) |
| Target SD | Target standard deviation |
| σ | Target standard deviation |
| UC | Universal Cell |
| VGA | Vapour Generator Accessory |

APPENDIX 5 - INSTRUMENT DETAILS FOR DISSOLVED ELEMENTS

Table 53 Instrument Conditions Al

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | | ORS | | 10 | NA | |
| 3 | ICP-OES-AV | | | | | | |
| 4 | ICP-MS | 45 | ORS | He | 1 | NA | 27 |
| 6 | | | | | | | |
| 7 | ICP-OES-AV | Y | NA | NA | 2 | 2 | 167.019 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 27 |
| 9 | ICP-MS/MS | Sc | ORS | H2-N2O | 1 | 1 | 27 |
| 10 | ICP-MS | Li6 | NA | NA | 1.05 | 21 | 27 |
| 11 | ICP-MS | Sc | UC | He | 20 | 20 | 27 |
| 12 | ICP-MS | Sc | NA | NA | 1 | NA | 27 |
| 13 | | | | | | | |

Table 54 Instrument Conditions Ag

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | GFAAS | | NA | | 10 | NA | 328.1 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NA | 107 |
| 6 | ICP-MS | In | | | 10 | 10 | 107 |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 107 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 107 |
| 9 | ICP-MS/MS | Te | ORS | H2-N2O | 1 | 1 | 107 |
| 10 | ICP-MS | Ir | ORS | He | 1.05 | 21 | 107 |
| 11 | ICP-MS | Rh | UC | He | 20 | 20 | 109 |
| 12 | ICP-MS | Rh | NA | NA | 1 | NA | 109 |
| 13 | ICP-MS | Rh | ORS | He | 10 | 10 | |

Table 55 Instrument Conditions As

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | DRC | O2 | NA | 25 | 91 |
| 2 | ICP-MS | Y | ORS | He | 10 | NA | 75 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 72 | ORS | HEHe | 1 | NA | 75 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 75 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 75 |
| 9 | ICP-MS/MS | Ge | ORS | H2-N2O | 1 | 1 | 75→91 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 75 |
| 11 | ICP-MS | Te | DRC | NH3 | 20 | 20 | 75 |
| 12 | ICP-MS | Rh | UC | He | 1 | NA | 75 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 56 Instrument Conditions Ba

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | NA | NA | NA | NA | NA | NA | NA |
| 3 | ICP-MS | | | | NA | | |
| 4 | NA | NA | NA | NA | NA | NA | NA |
| 6 | ICP-MS | In | | | NA | 10 | 137 |
| 7 | ICP-OES-AV | Y | NA | NA | NA | 2 | 455.403 |
| 8 | ICP-MS | Ge | ORS | He | NA | 10 | 137 |
| 9 | ICP-MS | Tb | ORS | He | NA | 1 | 137 |
| 10 | ICP-MS | Ir | ORS | He | NA | 21 | 137 |
| 11 | ICP-MS | Tb | UC | He | NA | 20 | 137 |
| 12 | NA | NA | NA | NA | NA | NA | NA |
| 13 | ICP-MS | Rh | ORS | He | NA | 10 | |

Table 57 Instrument Conditions Be

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Li | ORS | | 10 | NA | 9 |
| 3 | ICP-MS | | | | | NA | |
| 4 | ICP-OES-AV | Lu | | | 1 | NA | 313.107 |
| 6 | ICP-MS | Sc | | | 10 | NA | 9 |
| 7 | ICP-MS | Sc | NA | NA | 10 | NA | 9 |
| 8 | ICP-MS | Ge | ORS | He | 1 | NA | 9 |
| 9 | ICP-MS | Sc | ORS | NA | 1 | NA | 9 |
| 10 | ICP-MS | Li6 | NA | NA | 1.05 | NA | 9 |
| 11 | NA | NA | NA | NA | NA | NA | NA |
| 12 | ICP-MS | Sc | NA | NA | 1 | NA | 9 |
| 13 | ICP-MS | Ge | ORS | He | 10 | NA | |

Table 58 Instrument Conditions Cd

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Rh103 | NA | NA | NA | 25 | 111 |
| 2 | ICP-MS | In | ORS | | 10 | NA | 111 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NA | 114 |
| 6 | ICP-MS | In | | | 10 | 10 | 111 |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 111 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 111 |
| 9 | ICP-MS | Te | ORS | He | 1 | 1 | 111 |
| 10 | ICP-MS | Ir | ORS | He | 1.05 | 21 | 114 |
| 11 | ICP-MS | Rh | UC | He | 20 | 20 | 111 |
| 12 | ICP-MS | Rh | NA | NA | 1 | NA | 111 |
| 13 | ICP-MS | Rh | ORS | He | 10 | 10 | |

Table 59 Instrument Conditions Co

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Sc | ORS | | 10 | NA | 59 |
| 3 | ICP-MS | | | | | NA | |
| 4 | ICP-MS | 115 | ORS | He | 1 | NA | 59 |
| 6 | | | | | | NA | |
| 7 | ICP-MS | Rh | ORS | He | 10 | NA | 59 |
| 8 | ICP-MS | Ge | ORS | He | 1 | NA | 59 |
| 9 | ICP-MS | Ge | ORS | He | 1 | NA | 59 |
| 10 | ICP-MS | In | ORS | He | 1.05 | NA | 59 |
| 11 | ICP-MS | Ga | UC | He | 20 | NA | 59 |
| 12 | ICP-MS | Rh | UC | He | 1 | NA | 59 |
| 13 | ICP-MS | Ge | ORS | He | 10 | NA | |

Table 60 Instrument Conditions Cr

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | NA | He | NA | 25 | 52 |
| 2 | ICP-MS | Sc | ORS | | 10 | NA | 52 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 72 | ORS | He | 1 | NA | 52 |
| 6 | ICP-MS | Ga | DRC | NH3 | 10 | 10 | 52 |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 52 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 52 |
| 9 | ICP-MS/MS | Ge | ORS | H2-N2O | 1 | 1 | 52 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 52 |
| 11 | ICP-MS | Ga | DRC | NH3 | 20 | 20 | 52 |
| 12 | ICP-MS | Sc | UC | He | 1 | NA | 52 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 61 Instrument Conditions Cu

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | NA | He | NA | 25 | 63 |
| 2 | ICP-MS | Y | ORS | | 10 | NA | 65 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NA | 65 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 63 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 63 |
| 9 | ICP-MS | Ge | ORS | He | 1 | 1 | 63 |
| 10 | ICP-MS | Rh | ORS | He | 1.05 | 21 | 63 |
| 11 | ICP-MS | Ga | UC | He | 20 | 20 | 63 |
| 12 | seaFAST-ICP-MS | NA | UC | He | 1 | NA | 63 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 62 Instrument Conditions Fe

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | NA | He | NA | 25 | 56 |
| 2 | ICP-MS | Sc | ORS | | 10 | NA | 56 |
| 3 | ICP-OES-AV | | | | | | |
| 4 | ICP-MS | 72 | ORS | He | 1 | NA | 56 |
| 6 | | | | | | | |
| 7 | ICP-OES-AV | Y | NA | NA | 2 | 2 | 238.204 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 56 |
| 9 | ICP-MS/MS | Ge | ORS | H2-N2O | 1 | 1 | 56 |
| 10 | ICP-MS | Ge | ORS | H2 | 1.05 | 21 | 56 |
| 11 | ICP-MS | Ga | DRC | NH3 | 20 | 20 | 54 |
| 12 | seaFAST-ICP-MS | NA | UC | He | 1 | NA | 56 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 63 Instrument Conditions Hg

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|--------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | CVAAS | SnCl ₂ | NA | | 10 | NA | 3470 |
| 3 | CVAAS | | | | | NA | |
| 4 | VGA-ICP-OES | | | | 1 | NA | 194.164 |
| 6 | CVAAS | | | | 2.5 | NA | |
| 7 | CVAFS | NA | NA | NA | 2 | NA | NA |
| 8 | CVAFS | | | | 1 | NA | 253.7 |
| 9 | ICP-MS | Ir | ORS | He | 1 | NA | 202 |
| 10 | CVAFS | | NA | NA | 3 | NA | 253.7 |
| 11 | Fluorescence | NA | NA | NA | 5 | NA | 254 |
| 12 | ICP-MS | Ir | NA | NA | 1 | NA | 201 |
| 13 | ICP-MS | Ir | ORS | He | 10 | NA | |

Table 64 Instrument Conditions Mn

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------------------------------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Sc | ORS | | 10 | NA | 55 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 72 | ORS | He | 1 | NT | 55 |
| 6 | ICP-MS | Ga | | | 10 | 10 | 55 |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 55 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 55 |
| 9 | ICP-MS/MS | Sc | ORS | H ₂ -N ₂ O | 1 | 1 | 55 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 55 |
| 11 | ICP-MS | Ga | DRC | NH ₃ | 20 | 20 | 55 |
| 12 | seaFAST-ICP-MS | NA | UC | He | 1 | NA | 55 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 65 Instrument Conditions Mo

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | NA | NA | NA | NA | NA | NA | NA |
| 3 | ICP-MS | | | | NA | | |
| 4 | NA | NA | NA | NA | NA | NA | NA |
| 6 | ICP-MS | In | | | NA | 10 | 95 |
| 7 | ICP-MS | Rh | ORS | He | NA | 50 | 95 |
| 8 | ICP-MS | Ge | ORS | He | NA | 10 | 95 |
| 9 | ICP-MS/MS | In | ORS | H2-N2O | NA | 1 | 98→130 |
| 10 | ICP-MS | In | ORS | He | NA | 21 | 95 |
| 11 | ICP-MS | Rh | UC | He | NA | 20 | 98 |
| 12 | NA | NA | NA | NA | NA | NA | NA |
| 13 | ICP-MS | Rh | ORS | He | NA | 10 | |

Table 66 Instrument Conditions Ni

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Sc | ORS | He | 10 | NA | 60 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NT | 60 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 60 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 60 |
| 9 | ICP-MS | Ge | ORS | He | 1 | 1 | 60 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 60 |
| 11 | ICP-MS | Ga | UC | He | 20 | 20 | 60 |
| 12 | ICP-MS | Rh | UC | He | 1 | NA | 60 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 67 Instrument Conditions P

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Be9 | NA | NA | NA | 25 | 31 |
| 2 | ICP-OES-AV | Yb | NA | | 10 | NA | 177.4343 |
| 3 | ICP-OES-AV | | | | | | |
| 4 | | | | | | NA | |
| 6 | | | | | | | |
| 7 | ICP-OES-AV | Y | NA | NA | 2 | 2 | 213.618 |
| 8 | ICP-OES-AV | | | | 1 | 10 | 213.618 |
| 9 | ICP-MS/MS | Ge | ORS | H2-N2O | 1 | 1 | 31→47 |
| 10 | ICP-OES-AV | Lu | NA | NA | 1.05 | 10.5 | 177.434 |
| 11 | ICP-MS | Sc | UC | He | NA | 20 | 31 |
| 12 | | | | | | NA | |
| 13 | ICP-OES-AV | Te | ORS | NA | 1 | 1 | 213.618 |

Table 68 Instrument Conditions Pb

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Lu175 | NA | NA | NA | 25 | 208 |
| 2 | ICP-MS | Ir | ORS | | 10 | NA | 206 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NT | 208 |
| 6 | ICP-MS | Ir | | | 10 | 10 | 207 |
| 7 | ICP-MS | Ir | ORS | He | 10 | 50 | 207 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 208 |
| 9 | ICP-MS | Ir | ORS | He | 1 | 1 | 208 |
| 10 | ICP-MS | Ir | ORS | He | 1.05 | 21 | 208 |
| 11 | ICP-MS | Tb | UC | He | 20 | 20 | 206+207+208 |
| 12 | ICP-MS | Ir | NA | NA | 1 | NA | 206+207+208 |
| 13 | ICP-MS | Ir | ORS | He | 10 | 10 | |

Table 69 Instrument Conditions Sb

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | NA | NA | NA | NA | NA | NA | NA |
| 3 | ICP-MS | | | | NA | | |
| 4 | NA | NA | NA | NA | NA | NA | NA |
| 6 | ICP-MS | In | | | NA | 10 | 121 |
| 7 | ICP-MS | Rh | ORS | He | NA | 50 | 121 |
| 8 | ICP-MS | Ge | ORS | He | NA | 10 | 121 |
| 9 | ICP-MS/MS | Te | ORS | H2-N2O | NA | 1 | 121 |
| 10 | ICP-MS | In | ORS | He | NA | 21 | 121 |
| 11 | ICP-MS | Rh | UC | He | NA | 20 | 121 |
| 12 | NA | NA | NA | NA | NA | NA | NA |
| 13 | ICP-MS | Rh | ORS | He | NA | 10 | |

Table 70 Instrument Conditions Se

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | DRC | NH3 | NA | 25 | 82 |
| 2 | ICP-MS | Y | ORS | He | 10 | NA | 78 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | HEHe | 1 | NT | 78 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | HEHe | 10 | 50 | 78 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 78 |
| 9 | ICP-MS/MS | Ge | ORS | H2-N2O | 1 | 1 | 80→96 |
| 10 | ICP-MS | In | ORS | H2 | 1.05 | 21 | 78 |
| 11 | ICP-MS | Te | DRC | NH3 | 20 | 20 | 82 |
| 12 | ICP-MS | Rh | DRC | NH3 | 1 | NA | 82 |
| 13 | ICP-MS | Rh | ORS | HEHe | 10 | 10 | |

Table 71 Instrument Conditions Sn

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | In | ORS | He | 10 | NA | 118 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 115 | ORS | He | 1 | NT | 118 |
| 6 | ICP-MS | In | | | 10 | 10 | 118 |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 118 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 118 |
| 9 | ICP-MS/MS | In | ORS | H2-N2O | 1 | 1 | 118 |
| 10 | ICP-MS | In | ORS | He | 1.05 | 21 | 118 |
| 11 | ICP-MS | Rh | UC | He | 20 | 20 | 120 |
| 12 | | | | | | NA | |
| 13 | ICP-MS | Rh | ORS | He | 10 | 10 | |

Table 72 Instrument Conditions Tl

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Ir | ORS | | 10 | NA | 205 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 193 | ORS | He | 1 | NA | 205 |
| 6 | ICP-MS | Ir | | | 10 | 10 | 203 |
| 7 | ICP-MS | Ir | ORS | He | 10 | 50 | 205 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 205 |
| 9 | ICP-MS | Ir | ORS | He | 1 | 1 | 205 |
| 10 | ICP-MS | Ir | ORS | He | 1.05 | 21 | 205 |
| 11 | ICP-MS | Tb | UC | He | 20 | 20 | 205 |
| 12 | seaFAST-ICP-MS | NA | NA | NA | 1 | NA | 205 |
| 13 | ICP-MS | Ir | ORS | He | 10 | 10 | |

Table 73 Instrument Conditions U

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Ir | ORS | | 10 | NA | 238 |
| 3 | ICP-MS | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NA | 238 |
| 6 | | | | | | | |
| 7 | ICP-MS | Ir | ORS | He | 10 | 50 | 238 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 238 |
| 9 | ICP-MS | Tb | ORS | He | 1 | 1 | 238 |
| 10 | ICP-MS | Ir | ORS | He | 1.05 | 21 | 238 |
| 11 | ICP-MS | Tb | UC | He | 20 | 20 | 238 |
| 12 | ICP-MS | Ir | NA | NA | 1 | NA | 238 |
| 13 | ICP-MS | Ir | ORS | He | 10 | 10 | |

Table 74 Instrument Conditions V

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | NA | NA | NA | NA | NA | NA | NA |
| 2 | ICP-MS | Sc | ORS | | 10 | NA | 51 |
| 3 | ICP-OES-AV | | | | | | |
| 4 | ICP-MS | 45 | ORS | He | 1 | NT | 51 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 51 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 51 |
| 9 | ICP-MS/MS | Sc | ORS | H2-N2O | 1 | 1 | 80→96 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 51 |
| 11 | ICP-MS | Ga | DRC | NH3 | 20 | 20 | 51 |
| 12 | seaFAST-ICP-MS | NA | UC | He | 1 | NA | 51 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

Table 75 Instrument Conditions Zn

| Lab. Code | Instrument | Internal standard | Reaction/ Collision Cell | Cell Gas | S1 Final Dilution factor | S2 Final Dilution factor | Wavelength (nm)/ Ion(m/z) |
|-----------|----------------|-------------------|--------------------------|----------|--------------------------|--------------------------|---------------------------|
| 1 | ICP-MS | Ga71 | NA | He | NA | 25 | 66 |
| 2 | ICP-MS | Y | ORS | | 10 | NA | 66 |
| 3 | ICP-OES-AV | | | | | | |
| 4 | ICP-MS | 103 | ORS | He | 1 | NT | 66 |
| 6 | | | | | | | |
| 7 | ICP-MS | Rh | ORS | He | 10 | 50 | 64 |
| 8 | ICP-MS | Ge | ORS | He | 1 | 10 | 66 |
| 9 | ICP-MS/MS | Te | ORS | O2 | 1 | 1 | 66 |
| 10 | ICP-MS | Ge | ORS | He | 1.05 | 21 | 68 |
| 11 | ICP-MS | Te | DRC | NH3 | 20 | 20 | 66 |
| 12 | seaFAST-ICP-MS | NA | UC | He | 1 | NA | 66 |
| 13 | ICP-MS | Ge | ORS | He | 10 | 10 | |

END OF REPORT