

APPENDIX 1

NMI 14/2/170

Rev 0

14/03/13



Australian Government
National Measurement Institute

Pattern Approval Number: NMI 14/2/170

Manufacturer/Make: Electrex

Model	Type of meter	Connection / operation	Accuracy class	Phases	Wires	Reference frequency	Reference voltage	Rated or Basic current	Maximum current	Meter constant
Atto D4	Electronic	Transformer-operated	1	3	3,4	50 Hz	230-240 V AC	_[1]	1, 5 A	_[1]
Femto D4, Femto 96	Electronic	Transformer-operated	1	3	3,4	50 Hz	230-240 V AC	_[1]	1, 5 A	_[1]
Flash 96, Flash 96H	Electronic	Transformer-operated	1	3	3,4	50 Hz	20-500 V AC	_[1]	1, 5 A	_[1]
Zepto D6, Zepto 96	Electronic	Transformer-operated	1	3	3,4	50 Hz	230-240 V AC	_[1]	1, 5 A	_[1]

Note [1]: Values are not specified. Various values may be applicable.