



Australian Government

National Measurement  
Institute

Bradfield Road, West Lindfield NSW 2070

## Certificate of Approval

### NMI 14/2/19

Issued by the Chief Metrologist under Regulation 60  
of the  
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

EDMI Model Mk10 Atlas 2000-1010-1C Electricity Meter

submitted by EDMI Pty Ltd  
162 South Pine Road  
Brendale QLD 4500

**NOTE:** This Certificate relates to the suitability of the pattern of the instrument for use as a legal measuring instrument only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

This approval has been granted with reference to document NMI M 6, *Pattern Approval and Initial Verification of Electricity Meter and Associated Transformers. Definitions, Metrological and Technical Requirements*, July 2004.

This approval becomes subject to review on **1/12/15**, and then every 5 years thereafter.

#### DOCUMENT HISTORY

Rev	Reason/Details	Date
0	Pattern & variant 1 approved – interim certificate issued	26/09/05
1	Pattern & variant 1 approved – certificate issued	8/11/05
2	Pattern & variant 1 updated & reviewed – certificate issued	9/11/12

#### CONDITIONS OF APPROVAL

##### General

Instruments purporting to comply with this approval shall be marked with pattern approval number 'NMI 14/2/19' and only by persons authorised by the submittor.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

Auxiliary devices used with this instrument shall comply with the requirements of General Supplementary Certificates Nos S1/0/A or S1/0B.

Signed by a person authorised by the Chief Metrologist to exercise his powers under Regulation 60 of the *National Measurement Regulations 1999*.

A handwritten signature in black ink, consisting of a series of loops and a long horizontal stroke at the bottom.

## TECHNICAL SCHEDULE No 14/2/19

### 1. Description of Pattern

approved on 26/09/05

An EDM1 model Mk10 Atlas 2000-1010-1C Class 1 electronic polyphase direct connect static watt hour meter (Figure 1) used to measure electrical energy.

#### 1.1 Field of Operation

The field of operation of the measuring system is determined by the following characteristics:

- Number of phases 3
- Number of wires 4
- Reference frequency 50 Hz
- Reference ambient temperature ranges:
  - specified range of operation -25 to 80°C
  - limit range of operation -20 to 70°C
- Rated voltage 240/415 V AC
- Rated currents: Rated current,  $I_b$  10 A  
Maximum current,  $I_{max}$  100 A
- Meter constant 1000 imp/kWh
- Accuracy class 1

#### 1.2 Features/Functions

- Three (3) elements
- Electronic (LCD) digital indicator
- IEC standard optical interface (either FLAG or ANSI configuration)
- 8 terminal bottom connect rectangular base
- 4 LED pulse indicators
- a communications port, either RS242 or RS485
- load profile memory

#### 1.3 Verification Provision

Provision is made for the application of a verification mark.

#### 1.4 Sealing Provision

Provision is made for the instrument to be sealed by the application of one or more mechanical seals (Figures 1 & 2).

## 1.5 Descriptive Markings and Notices

Instruments are marked with the following data, together in one location, in the form shown at right:

Manufacturer's name or mark	...
Model designation	...
Serial number	...
Pattern approval mark	NMI 14/2/19
Number of phases	...
Number of wires	...
Reference frequency	... Hz
Meter constant	...
Rated voltage	... AC
Rated currents:	$I_b$ ... A
	$I_{max}$ ... A
Accuracy index	Class 1

## 2. Description of Variant 1

approved on 26/09/05

The model Mk10 Atlas 2000-1010-1A (Figure 3) which is configured for current transformer connection.

The field of operation is the same as for the pattern except for the following:

- Rated currents: Rated current,  $I_b$  5 A  
Maximum current,  $I_{max}$  20 A
- Meter constant 1000 imp/kWh

### TEST PROCEDURE No 14/2/19

Instruments tested for initial verification shall comply with the certificate of approval and technical schedule, and the maximum permissible errors for verifications at the operating conditions in effect at the time of verification.

Meters shall be verified in accordance with NITP 14 *National Instrument Test Procedures for Utility Meters*.

NOTE: NMI reserves the right to vary this procedure. Any such variation shall be notified in writing by NMI.

FIGURE 14/2/19 – 1



EDMI Model Mk10 Atlas 2000-1010-1C Electricity Meter

FIGURE 14/2/19 – 2



Typical Mechanical Sealing

FIGURE 14/2/19 – 3



Model Mk10 Atlas 2000-1010-1A – Variant 1

~ End of Document ~