



Bradfield Road, West Lindfield NSW 2070

## Certificate of Approval

### No 14/2/62

Issued by the Chief Metrologist under Regulation 60  
of the  
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

Elster Model A1100-(LB3AABBS-A) Electricity Meter

submitted by Elster Metering Systems  
Paton Drive Tolgate Business Park  
Beaconside, Stafford, Staffordshire ST16 3EF  
UNITED KINGDOM

**NOTE:** This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

This approval has been granted with reference to document NMI M 6-1, *Electricity Meters, Part 1: Metrological and Technical Requirements*, July 2012.

This approval becomes subject to review on 1/04/18, and then every 5 years thereafter.

#### DOCUMENT HISTORY

| Rev | Reason/Details  | Date     |
|-----|---|----------|
| 0   | Pattern approved – certificate issued   | 1/03/13  |
| 1   | Pattern amended (Figure 1 replaced, without 'For indoor use only' label) – certificate issued | 10/04/14 |
|     |   |          |

## CONDITIONS OF APPROVAL

### General

Instruments purporting to comply with this approval shall be marked with pattern approval number 'NMI 14/2/62' and only by persons authorised by the submitter.

It is the submitter's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

Auxiliary devices used with this instrument shall comply with the requirements of General Supplementary Certificates No S1/0/A or No S1/0B.

Signed by a person authorised by the Chief Metrologist to exercise their powers under Regulation 60 of the *National Measurement Regulations 1999*.

A handwritten signature in black ink, appearing to read 'Dr A Rawlinson', with a horizontal line underneath.

**Dr A Rawlinson**

## TECHNICAL SCHEDULE No 14/2/62

### 1. Description of Pattern

approved on 1/03/13

An Elster model A1100-(LB3AABBS-A) Class 1 electronic polyphase direct connect static watt hour meter (Figure 1) used to measure electrical energy.

Note: A1100 is the basic series identifier. The full model number may have a variety of additional alpha numeric characters representing non metrological functions e.g. **LB3AABDSSNNS-AN**).

#### 1.1 Field of Operation

The field of operation of the measuring system is determined by the following characteristics:

- |   |                                       |                    |
|---|---------------------------------------|--------------------|
| • | Number of phases                      | 3                  |
| • | Number of wires                       | 4                  |
| • | Reference frequency                   | 50 Hz              |
| • | Reference ambient temperature ranges: |                    |
|   | specified range of operation          | -40 to 60°C        |
|   | limit range of operation              | -40 to 70°C        |
| • | Rated voltage                         | 3 × 240 (415) V AC |
| • | Rated currents:                       |                    |
|   | Basic current, $I_b$                  | 10 or 20 A         |
|   | Maximum current, $I_{max}$            | 100 A              |
| • | Meter constant                        | 1000 imp/kWh       |
| • | Accuracy class                        | 1                  |

#### 1.2 Features/Functions

- Three (3) elements
- Electronic (LCD) digital indicator
- Meters may have an import register only or both import and export registers
- Auxiliary output
- IrDA data output
- Optical serial port
- Bottom connect rectangular base.

Note: Meters are NOT fitted with an internal clock.

#### 1.3 Verification Provision

Provision is made for the application of a verification mark.

#### 1.4 Sealing Provision

Provision is made for the instrument to be sealed by the application of one or more mechanical seals (Figure 1).

## 1.5 Descriptive Markings and Notices

Instruments are marked with the following data, together in one location, in the form shown at right:

|                             |                 |
|-----------------------------|-----------------|
| Manufacturer's name or mark | ...             |
| Model designation           | ...             |
| Serial number               | ...             |
| Pattern approval mark       | NMI 14/2/62     |
| Number of phases            | ...             |
| Number of wires             | ...             |
| Reference frequency         | ... Hz          |
| Meter constant              | ...             |
| Rated voltage               | ... AC          |
| Rated currents:             | $I_b$ ... A     |
|                             | $I_{max}$ ... A |
| Accuracy index              | Class 1         |

### TEST PROCEDURE No 14/2/62

Instruments tested for initial verification shall comply with the certificate of approval and technical schedule, and the maximum permissible errors for verifications at the operating conditions in effect at the time of verification.

Meters shall be verified in accordance with NITP 14 *National Instrument Test Procedures for Utility Meters*.

NOTE: NMI reserves the right to vary this procedure. Any such variation shall be notified in writing by NMI.

FIGURE 14/2/62 – 1



Elster A1100 Series Electricity Meter

~ End of Document ~