



Australian Government
Department of Industry,
Science and Resources

**National
Measurement
Institute**

36 Bradfield Road, West Lindfield NSW 2070

Certificate of Approval
NMI 14/2/124

Issued by the Chief Metrologist under Regulation 60
of the
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

DZG Metering GmbH model MDVH3016 Class 0.2 S Electricity Meter

submitted by Heidelberg Str. 32
16515 Oranienburg
Germany

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

This approval has been granted with reference to document NMI M 13-1 *Active-energy electricity meters (a.c.) Part 1: Metrological and Technical Requirements*, June 2022.

This approval is subject to review at the decision of the Chief Metrologist in accordance with the conditions specified in the document NMI P 106.

DOCUMENT HISTORY

Rev	Reason/Details	Date
0	Pattern & Variants 1 to 4 approved – certificate issued	19/06/25

CONDITIONS OF APPROVAL

General

Instruments purporting to comply with this approval shall be marked with approval number 'NMI 14/2/124' and only by persons authorised by the submitter.

It is the submitter's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

Signed by a person authorised by the Chief Metrologist
to exercise their powers under Regulation 60 of the
National Measurement Regulations 1999.

**Phillip Mitchell**

Acting Manager
Policy and Regulatory Services

TECHNICAL SCHEDULE No 14/2/124

1. Description of Pattern

approved on 19/06/25

A DZG Metering GmbH model MDVH3016 polyphase class 0.2 S transformer-operated static watt hour meter (Figure 1) used to measure electrical energy.

1.1 Field of Operation

The field of operation of the measuring system is determined by the following characteristics:

- | | |
|------------------------------|--|
| • Number of phases | 3 |
| • Number of wires | 3 or 4 |
| • Reference frequency | 50 Hz |
| • Rated voltage | 3 × 110 V AC (3P3W) or
3 × 57.7/100 V to
240/415 V AC (3P4W) |
| • Rated current, I_n | 1 A |
| • Maximum current, I_{max} | 20 A |
| • Meter constant | 50000 imp/kWh |
| • Accuracy class | 0.2 S |
| • Reference temperature | 23 °C |

1.2 Features/Functions

- Liquid crystal digital indicators having a maximum display of 9999 9999 9999 9 Wh, kWh and MWh
- Internal crystal-controlled clock
- Measurement in positive and negative directions (import and export)
- Communication interfaces: Optical, RS485, RS232 and Ethernet ports

1.3 Verification Provision

Provision is made for the application of a verification mark.

1.4 Sealing Provision

Provision is made for the sealing devices and parameters that have a metrologically significant effect and that determine the measurement result by the application of mechanical seals (Figure 1).

1.5 Descriptive Markings

Instruments are clearly and permanently marked with the following data, in the vicinity of the indicating device, in the form shown right:

Manufacturer's mark, or name written in full
Model designation
Serial number
Pattern approval mark	NMI 14/2/124
Number of phases
Number of wires
Reference frequency Hz
Meter constant
Rated voltage AC
Rated currents:	I_n A
	I_{max} A
Accuracy class	...
Environment	Indoor meter or IM

1.6 Harmonics

Instruments purporting to comply with this approval are suitable for use where the harmonics do not exceed those specified in AS 62053.22:2018.

2. Description of Variant 1

approved on 19/06/25

A DZG Metering GmbH model MDVH3016 polyphase class 0.2 S transformer-operated static watt hour meter used to measure electrical energy.

This variant has the same Field of Operations and Features/Functions as the pattern except for the following:

- Maximum current, I_{max} 6 A

3. Description of Variant 2

approved on 19/06/25

A DZG Metering GmbH model MDVH3016 polyphase class 0.2 S transformer-operated static watt hour meter used to measure electrical energy.

This variant has the same Field of Operations and Features/Functions as the pattern except for the following:

- Maximum current, I_{max} 10 A

4. Description of Variant 3**approved on 19/06/25**

A DZG Metering GmbH model MDVH3016 polyphase class 0.2 S transformer-operated static watt hour meter used to measure electrical energy.

This variant has the same Field of Operations and Features/Functions as the pattern except for the following:

- Rated current, I_n 5 A
- Maximum current, I_{max} 10 A

5. Description of Variant 4**approved on 19/06/25**

A DZG Metering GmbH model MDVH3016 polyphase class 0.2 S transformer-operated static watt hour meter used to measure electrical energy.

This variant has the same Field of Operations and Features/Functions as the pattern except for the following:

- Rated current, I_n 5 A
- Maximum current, I_{max} 20 A

TEST PROCEDURE No 14/2/124

Instruments tested for verification shall comply with the certificate of approval and technical schedule, and the maximum permissible errors for verifications at the operating conditions in effect at the time of verification.

The maximum permissible errors are specified in the *National Trade Measurement Regulations 2009* (Cth).

Electricity meters shall be verified in accordance with the following National Instrument Test Procedures:

- NITP 14.0 – Utility Meters – general requirements
- NITP 14.2 – Utility Meters – electricity meters

NOTE: NMI reserves the right to vary this procedure. Any such variation shall be notified in writing by NMI.

FIGURE 14/2/124 – 1



DZG Metering GmbH model MDVH3016 Class 0.2 S Electricity Meter (the Pattern, including example of required markings and typical mechanical sealing)

~ End of Document ~