



Australian Government
**Department of Industry,
Science and Resources**

**National
Measurement
Institute**

36 Bradfield Road, West Lindfield NSW 2070

Conversion Certificate of Approval
NMI 6B/257

Issued by the Chief Metrologist under Regulation 60
of the
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

Nuweigh Model IH-60-20 (Converted) Weighing Instrument

submitted by Industrial Weighing Solutions Pty Ltd
14 Bradman Avenue
Bowral NSW 2576

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

This approval is subject to review at the decision of the Chief Metrologist in accordance with the conditions specified in the document NMI P 106.

DOCUMENT HISTORY

Rev	Reason/Details	Date
0	Pattern & variants 1 and 2 approved – certificate issued	18/03/26

CONDITIONS OF APPROVAL

General

Instruments purporting to comply with this approval shall be marked with pattern approval number 'NMI 6B/257' and only by persons authorised by the submittor.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

The values of the performance criteria (maximum number of scale intervals etc.) applicable to the instrument shall be within the limits specified herein and in any approval documentation for the components where they are approved separately.

Special:

Instruments shall not be manufactured using this approval. Instruments shall only be converted or modified as described herein or with substitute load cells and/or indicator, and in other capacities, in accordance with General Certificate of Approval No 6B/0.

Signed by a person authorised by the Chief Metrologist to exercise their powers under Regulation 60 of the *National Measurement Regulations 1999*.



Phillip Mitchell
A/g Manager
Policy and Regulatory Services

TECHNICAL SCHEDULE No 6B/257

1. Description of Pattern **approved on 18/03/26**

A converted Nuweigh model IH-60-20 non-automatic self-indicating weighing instrument of 20 000 kg maximum capacity with a verification scale interval of 20 kg and with a minimum capacity of 400 kg. The instrument (as described in approval NMI 6/10B/80) uses four IWS model P510A C3 load cells of 30 000 kg capacity (as described in the documentation of approval NMI S416), and has been converted by replacement of the indicator and now uses a Rinstrum model C320 digital indicator (as described in the documentation of approval NMI S869).

2. Description of Variant 1 **approved on 18/03/26**

With the basework of any approved mechanical or lever/load cell weighing instrument converted in accordance with General Certificate of Approval No 6B/0 (refer Special Condition of Approval).

3. Description of Variant 2 **approved on 18/03/26**

With the indicator of any approved full load cell weighing instrument replaced by another approved digital indicator in accordance with General Certificate of Approval No 6B/0 (refer to the Special Condition of Approval) and Certificate of Approval of the instrument converted or modified.

TEST PROCEDURE No 6B/257

Instruments shall be tested in accordance with any relevant tests specified in the National Instrument Test Procedures.

The instrument shall not be adjusted to anything other than as close as practical to zero error, even when these values are within the maximum permissible errors.

Maximum Permissible Errors

The maximum permissible errors are specified in Schedule 1 of the *National Trade Measurement Regulations 2009*.