

VARIATION No 1

This is to certify that the following modifications of the pattern and variants of the
Epex Bulk Dispenser

approved in Certificate No 5/6D/14 dated 8 April 1970

submitted by Engineering Products Pty Ltd, 418-428 Burnley Street, Burnley,
Victoria, 3121,

have been approved under the Weights and Measures (Patterns of Instruments)
Regulations as being suitable for use for trade.

The approved modifications are as follows:

- (a) fitting an OPW 1 $\frac{1}{4}$ -inch IA automatic nozzle; and
- (b) converting all flowmeter models to indicate in metric units in accordance with
Appendix 13 of the General Specifications for Measuring Instruments to be
Used for Trade.


Approval was granted on 15 May 1974.

This variation is described in Technical Schedule No 5/6D/14, Variation No 1, and
in drawings and specifications lodged with the Commission.

The approval is subject to review on or after 31 December 1975.

All instruments conforming to this approval shall be marked with the approval
number "NSC No 5/6D/14".

Signed



Executive Officer

Indra



Weights and Measures
(National Standards)
Act 1960-1964
Weights and Measures
(Patterns of Instruments)
Regulations

COMMONWEALTH OF AUSTRALIA

NATIONAL STANDARDS COMMISSION

Certificate of Approval

CERTIFICATE NUMBER 5/6D/14

In respect of the pattern of

Epex Bulk Dispenser and Variants 1 and 2.

Submitted and
manufactured by: Engineering Products Pty. Ltd.,
418-428 Burnley Street,
Burnley,
Victoria. 3121.

This is to certify that the pattern and variants of the instrument described in this Certificate have been examined by the National Standards Commission under the provisions of the abovementioned Regulations and have been approved as being suitable for use for trade.

Approval was granted for the pattern and variants 1 and 2 on 3rd April, 1970.

Approval was granted on condition that all instruments made in conformity with the pattern or variants:

1. are appropriately marked NSC No 5/6D/14 and, where required by State legislation, with the State approval number also; and
2. comply with the General Specifications for Weighing and Measuring Instruments to be Used for Trade in respect of that part of the instrument which was not previously approved by a State.

8/4/70

Cont'd over

This Certificate comprises pages 1 and 2 dated 8th April, 1970.

Pursuant to regulation 12 of the abovementioned Regulations, this Certificate is applicable only in the State of Victoria.

Date of issue 8th April, 1970.

Signed



A person authorised by the Commission
to sign Certificates under the
abovementioned Regulations.

DESCRIPTION OF PATTERN

The pattern is of a wholesale flowmeter as described in Certificate No 5/6D/6 dated 16th February, 1970.

DESCRIPTION OF VARIANTS

1. The variants 1 to 7 as described in Certificate No 5/6D/6 dated 16th February, 1970.
2. The pattern or variant 1 fitted with a Gilbarco T279-5 nozzle as approved in Victorian Notice of Approval No 527.

8/4/70



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No 5/6D/14

VARIATION No 1

Pattern: Epex Bulk Dispenser

Submittor: Engineering Products Pty Ltd,
418-428 Burnley Street,
Burnley, Victoria, 3121.

Date of Approval of Variants: 15 May 1974

The modifications described in this schedule apply to the pattern and variants described in the following pages and figures of Certificate No 5/6D/14 dated 8 April 1970:

Page 2 dated 8 April 1970


All instruments conforming to this approval shall be marked "NSC No 5/6D/14".

Description:

This variation approves:


1. fitting an OPW $1\frac{1}{4}$ -inch 1A automatic hose nozzle (see Figures 1 and 2);
2. converting all flowmeter models to indicate in metric units in accordance with Appendix 13 of the General Specifications for Measuring Instruments to be Used for Trade.

FIGURE 5/6D/14 - 1



OPW 1 1/4-inch 1A Nozzle

FIGURE 5/6D/14 - 2



OPW 1 1/4-inch 1A Nozzle — Schematic Diagram