



**Australian Government**  
**National Measurement  
Institute**

Bradfield Road, West Lindfield NSW 2070

**Cancellation**  
**Certificate of Approval**  
**No 6/4D/307A**

Issued by the Chief Metrologist under Regulation 60  
of the  
*National Measurement Regulations 1999*

This is to certify that the approval for use for trade granted in respect of the

Bizerba Model SC-H 200 Weighing Instrument

submitted by        Toshiba TEC Australia Pty Ltd  
                          Unit 1, 9-11 South Street  
                          RYDALMERE    NSW    2116

has been cancelled in respect of new instruments as from 1 June 2012.

Signed by a person authorised by the Chief Metrologist  
to exercise his powers under Regulation 60 of the  
*National Measurement Regulations 1999*.

A handwritten signature in black ink, consisting of a series of loops and flourishes, positioned to the right of the signature text.



**Australian Government**  

---

**National Measurement  
Institute**

Bradfield Road, West Lindfield NSW 2070

## **Certificate of Approval**

### **No 6/4D/307A**

Issued by the Chief Metrologist under Regulation 60  
of the  
*National Measurement Regulations 1999*

This is to certify that an approval for use for trade has been granted in respect of the

Bizerba Model SC-H 200 Weighing Instrument

submitted by Toshiba TEC Australia Pty Ltd  
Unit 1, 9-11 South Street  
RYDALMERE NSW 2116.

**NOTE:** This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

This Certificate is issued upon completion of a review of approval NSC 6/4D/307.  
This approval has been granted with reference to document NMI R 76, *Non-automatic weighing instruments, Parts 1 and 2*, dated July 2004.

#### **CONDITIONS OF APPROVAL**

This approval becomes subject to review on 1 February 2012, and then every 5 years thereafter.

Instruments purporting to comply with this approval shall be marked NMI 6/4D/307A and only by persons authorised by the submittor.

It is the submitter's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

The National Measurement Institute reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

Auxiliary devices used with this instrument shall comply with the requirements of General Supplementary Certificate No S1/0/A.

#### DESCRIPTIVE ADVICE

**Pattern:** approved 19 January 2007

- A Bizerba model SC-H 200 multi-interval self-indicating price-computing weighing instrument with a maximum capacity of 15 kg.

**Variants:** approved 19 January 2007

1. Certain models of the SC-H series as single interval instruments.
2. Certain models of the SC-H 100 series.
3. Certain models of the SC-H 500 series.
4. The model SC-H 800.
5. Models of the SC-H series connected in a network.
6. Models of the SC-H series connected in a network using wireless LAN (local area network) technology.
7. With an alternative display, in which case the model number has an 'OD' suffix.
8. The model SC-H 400 OD.

Technical Schedule No 6/4D/307A describes the pattern and variants 1 to 8.

#### FILING ADVICE

The documentation for this approval comprises:

Certificate of Approval No 6/4D/307A dated 22 January 2007  
Technical Schedule No 6/4D/307A dated 22 January 2007 (incl. Test Procedure)  
Figures 1 to 7 dated 22 January 2007

Signed by a person authorised by the Chief Metrologist to exercise his powers under Regulation 60 of the *National Measurement Regulations 1999*.



TECHNICAL SCHEDULE No 6/4D/307A

**Pattern:** Bizerba Model SC-H 200 Weighing Instrument

**Submittor:** Toshiba TEC Australia Pty Ltd  
Unit 1, 9-11 South Street  
RYDALMERE NSW 2116

## 1. Description of Pattern

A Bizerba model SC-H 200 multi-interval self-indicating price-computing weighing instrument (Figures 1 and 2) with a verification scale interval ( $e_1$ ) of 0.002 kg up to 6 kg and with a verification scale interval ( $e_2$ ) of 0.005 kg from 6 kg up to the maximum capacity of 15 kg.

The instrument has an operator keyboard and display integrated into the instrument body, and a column-mounted customer display. Each display is a liquid crystal dot matrix type, on which weight, unit price and price information, together with alphanumeric information relating to product look up (PLU) items, is displayed. In addition when a tare or pre-set tare is operational a display of the tare value is provided.

Instruments may be fitted with an integral label or ticket printer. A suffix to the model number indicates the printer type, e.g. the model SC-H 200 T is provided with a thermal receipt printer while the model SC-H 200 E has a thermal label printer.

Instruments have unit price to \$9999.99/kg, price to \$9999.99, a product look up (PLU) facility, and may be fitted with output sockets (output interfacing capability) for the connection of peripheral and/or auxiliary devices.

Instruments may be provided with a 53 key PLU keyboard.

### 1.1 Zero

Zero is automatically corrected to within  $\pm 0.25e_1$  whenever power is applied and whenever the instrument comes to rest within  $0.5e_1$  of zero.

The initial zero-setting device has a nominal range of not more than 20% of the maximum capacity of the instrument.

One of the instrument keys may be configured to operate a semi-automatic zero-setting device with a nominal range of not more than 4% of the maximum capacity of the instrument.

### 1.2 Tare

A semi-automatic subtractive tare device of up to maximum capacity may be fitted. In addition, a keyboard-entered pre-set subtractive tare device of up to 6 kg maximum capacity may be fitted.

A separate display of the tare value is provided whenever a tare is active.

Pre-set tare values may also be associated with PLU keys.

### 1.3 Display Check

A display check is initiated whenever power is applied.

### 1.4 Levelling

The instrument is provided with adjustable feet and adjacent to the level indicator is a notice advising that the instrument must be level when in use.

### 1.5 Verification/Certification Provision

Provision is made for the application of a verification/certification mark.

### 1.6 Sealing Provision

Provision is made for the calibration adjustments to be sealed by means of a destructible adhesive label, or lead and wire seal, used to prevent removal of the plate that provides access to the pins that allow calibration (Figure 3).

### 1.7 Descriptive Markings

Instruments carry the following markings:

Manufacturer's mark, or name written in full	Bizerba, Germany
Name or mark of manufacturer's agent	.....
Indication of accuracy class	Ⓜ
Pattern approval mark for the instrument	NMI 6/4D/307A
Maximum capacity	<i>Max</i> ...../..... kg #1
Minimum capacity	<i>Min</i> ..... kg #1
Verification scale interval	<i>e</i> = ...../..... kg #1
Tare capacity	<i>T</i> = - ..... kg #2
Serial number of the instrument	.....

#1 These markings are also shown near the display of the result if they are not already located there.

#2 This marking is required if *T* is not equal to *Max*.

## 2. Description of Variants

### 2.1 Variant 1

The pattern and certain other models of the SC-H 100, SC-H 200, SC-H 500, and SC-H 800 series as single interval instruments. Instruments may be fitted with a semi-automatic subtractive tare device and/or a keyboard-entered pre-set subtractive tare device, each of up to maximum capacity. Instruments may be in capacities as listed below:

- Of 6 kg maximum capacity with a verification scale interval of 0.002 kg.
- Of 15 kg maximum capacity with a verification scale interval of 0.005 kg.
- Of 30 kg maximum capacity with a verification scale interval of 0.01 kg.

## 2.2 Variant 2

Certain models of the SC-H 100 series as listed below.

- Model SC-H 100 (Figures 2 and 4) which is similar to the pattern but has the customer display integrated into the instrument body.
- Model SC-H 100 G (Figure 2) which is similar to the model SC-H 100 but has the basework in a separate unit to the display/printer.

## 2.3 Variant 3

(i) Certain models of the SC-H 500 series as listed below.

- Model SC-H 500 SB1 (Figures 2 and 5) which is similar to the pattern but with the display, printer and keyboard mounted on a column. It has an additional PLU keyboard with up to 96 keys.
- Model SC-H 500 SB2 (Figure 2) which is similar to the model SC-H 500 SB1 but has two additional PLU keyboards each with up to 96 keys.

(ii) Certain other models of the SC-H 500 series as listed below. These instruments are intended for self-service operation only, and all the operator interface keys of the pattern (other than the PLU keyboards) are covered.

The use of the totalisation across instruments arrangement described in variant 5 is not approved in this self-service arrangement.

- Model SC-H 500 QS1 which is similar to the model SC-H 500 SB1.
- Model SC-H 500 QS2 which is similar to the model SC-H 500 SB2.

## 2.4 Variant 4

The model SC-H 800 (Figure 6) which is similar to the pattern but with the display, printer and keyboard mounted on a column.

## 2.5 Variant 5

The models of the SC-H series may be connected in a network with compatible approved Bizerba instruments, to share common PLU data, for totalisation across instruments (that is, transactions entered by an operator at different instruments can be totalised at any instrument in the network), and to accumulate and retrieve management information.

In addition, the network may be interfaced with a computer for the collection of management data, or the downloading of PLU data.

Note: The weighing and price-computing functions of each weighing instrument in the network are independent, and the removal, repair or replacement of a particular weighing instrument does not necessitate reverification of any other weighing instrument in the network.

## 2.6 Variant 6

The models of the SC-H series may be connected in a network (as described in variant 5) but using wireless LAN (local area network) technology rather than a physical connection.

## 2.7 Variant 7

Models of the pattern and variants with an orange liquid crystal display (LCD), and identified by use of 'OD' as a suffix to the model name, e.g. the pattern, model SC-H 200 becomes model SC-H 200 OD.

Note: In addition to the T and E suffixes mentioned in clause 1. **Description of Pattern** an 'F' suffix may also be used to designate an instrument with a column mounted display facing the customer.

## 2.8 Variant 8

The model SC-H 400 OD which is similar to the pattern, but is in the form of a suspended (hanging) scale (Figure 7).

As the instrument is freely suspended, the levelling arrangements described for the pattern are not required.

## TEST PROCEDURE

Instruments should be tested in accordance with any relevant tests specified in the Uniform Test Procedures.


### Maximum Permissible Errors at Verification/Certification

For single range instruments, the maximum permissible errors for increasing and decreasing loads on initial verification/certification for loads,  $m$ , expressed in verification scale intervals,  $e$ , are:

- $\pm 0.5e$  for loads  $0 \leq m \leq 500$ ;
- $\pm 1.0e$  for loads  $500 < m \leq 2\,000$ ; and
- $\pm 1.5e$  for loads  $2\,000 < m \leq 10\,000$ .

For multi-interval instruments with verification scale intervals of  $e_1, e_2 \dots$ , apply  $e_1$  for zero adjustment, and maximum permissible errors apply  $e_1, e_2 \dots$ , as applicable for the load.

FIGURE 6/4D/307A – 1



Bizerba Model SC-H 200 Weighing Instrument

FIGURE 6/4D/307A – 2





FIGURE 6/4D/307A – 3




Sealing of Calibration Access

FIGURE 6/4D/307A – 4



Bizerba Model SC-H 100 Weighing Instrument

FIGURE 6/4D/307A – 5




Bizerba Model SC-H 500 SB1 Weighing Instrument

FIGURE 6/4D/307A – 6



Bizerba Model SC-H 800 Weighing Instrument

FIGURE 6/4D/307A – 7



Bizerba Model SC-H 400 OD