



Australian Government
Department of Industry,
Innovation and Science

National Measurement Institute

Interim Certificate of Approval NMI 6/4D/373

VALID FOR VERIFICATION PURPOSES UNTIL 26 JUNE 2016

Issued by the Chief Metrologist under Regulation 60
of the
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the instruments herein described.

Bizerba Model KHII 800 Weighing Instrument

submitted by Bizerba Australia Pty Ltd
2 McGregors Drive
Keilor Park Victoria 3042

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

DOCUMENT HISTORY

Rev	Reason/Details	Date
0	Pattern & variants 1 to 4 approved – interim certificate issued	29/05/14
1	Pattern & variants 1 to 4 approved – certificate issued	19/01/15
2	Variant 5 provisionally approved – interim certificate issued	26/06/15
3	Variant 5 amended (extra site) – interim certificate issued	13/10/15

CONDITIONS OF APPROVAL

General

Instruments purporting to comply with this approval shall be marked with pattern approval number 'NMI 6/4D/373' and only by persons authorised by the submittor.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

Auxiliary devices used with this instrument shall comply with the requirements of General Supplementary Certificate No S1/0B.

Special Conditions of Approval: (Provisional Approval)

This approval is limited to two (2) instruments only, located at the Coles store, 180 Gaffney Street, Coburg North VIC, and one (1) instrument only, located at the Coles store, Cockburn Gateway, Shop M2, 182 Beeliar Drive Success WA.

Instruments purporting to comply with this approval shall be marked with approval number 'NMI P6/4D/373' and only by persons authorised by the submittor. (Note: The 'P' in the approval number may be a temporary marking.)

The approval will remain provisional pending completion of satisfactory testing and evaluation.

The submittor shall provide NMI with copies of test results from the initial verification and all subsequent tests.

In the event of unsatisfactory performance the approval may be cancelled (or altered).

The submittor shall implement such modifications as required by NMI. In the event that such modifications (if any are required by NMI) are not made to the satisfaction of NMI, this approval may be withdrawn.

1. Description of Pattern approved on 29/05/14

A Bizerba model KHII 800 class III multi-interval self-indicating price-computing non-automatic weighing instrument with a verification scale interval e_1 of 0.002 kg up to 6 kg and with a verification scale interval e_2 of 0.005 kg from 6 kg to the maximum capacity of 15 kg.

Technical Schedule No 373 dated 19/01/15 describes the pattern & variants 1 to 4.

2. Description of Variant 5 provisionally approved on 26/06/15

A Bizerba model KF II GSP H KF class III multi-interval self-indicating price-computing non-automatic weighing instrument with a verification scale interval e_1 of 0.001 kg up to 3 kg and with a verification scale interval e_2 of 0.002 kg from 3 kg to the maximum capacity of 6 kg. The basework of the pattern is fitted with an integral model GSP H slicer.

Instruments are fitted with a column-mounted touchscreen operator display/ keyboard and a column-mounted single-sided colour customer display. The operator touchscreen consists of displays for presentation of tare, weight, unit price and price information, zero, 'net' indicators and functions relating to product look up (PLU) items.

Instruments are fitted with a printer, for printing of labels or tickets.

Instruments display unit price to \$9999.99/kg, total price to \$9999.99, and have a product look up (PLU) facility.

Signed by a person authorised by the Chief Metrologist to exercise their powers under Regulation 60 of the *National Measurement Regulations 1999*.

A handwritten signature in black ink, appearing to read 'A Rawlinson', with a horizontal line underneath.

Dr A Rawlinson

~ End of Document ~