



NATIONAL STANDARDS COMMISSION

6/10A/151
21/12/88

NATIONAL MEASUREMENT (PATTERNS OF INSTRUMENTS) REGULATIONS

REGULATION 9

CERTIFICATE OF APPROVAL No 6/10A/151

This is to certify that an approval for use for trade has been granted in respect of the pattern of the

Toledo Model 840 Weighing Instrument of Capacities between 15 000 kg and 150 000 kg.

CONDITIONS OF APPROVAL

Special:

This approval permits existing instruments in the field to be verified. It does NOT permit the manufacture of new instruments to the pattern, and as such it does not automatically become subject to review.

Instruments purporting to comply with this approval shall be marked NSC No 6/10A/151 prior to verification.

Instruments purporting to comply with this approval may be verified by a Weights and Measures authority for as long as they remain within the appropriate maximum permissible errors.

Instruments may be relocated and reinstalled but shall not be altered in any way other than to replace damaged components with equivalent components or as provided for in General Certificate No 6B/0.

Signed

Executive Director

Descriptive Advice

Pattern: approved 8/11/88

- Toledo model 840 non-self-indicating weighing instruments of capacities between 15 000 kg and 150 000 kg.

Test Procedure

Instruments should be tested in accordance with the relevant sections of the Inspector's Handbook.

The results shall not exceed the maximum permissible errors given in Document 118, 2nd Edition, October 1986.

Filing Advice

The documentation for this approval comprises only this Certificate.