



# NATIONAL STANDARDS COMMISSION

G.H.  
6/10A/152  
21/8/89

## NATIONAL MEASUREMENT (PATTERNS OF INSTRUMENTS) REGULATIONS

### REGULATION 9

#### CERTIFICATE OF APPROVAL No 6/10A/152

This is to certify that an approval for use for trade has been granted in respect of the pattern of the

Asco Weighing Instrument of Capacities between 15 000 kg and 150 000 kg.

#### CONDITIONS OF APPROVAL

##### Special:

This approval does NOT permit the verification of new instruments of the pattern. It does not become due for review.

This approval applies only to instruments in the field which have been previously verified; they may be re-verified, or verified after repairs, at the discretion of the relevant Weights and Measures authority.

Instruments purporting to comply with this approval shall be marked NSC No 6/10A/152 prior to re-verification.

Instruments may be relocated and reinstalled but shall not be altered in any way other than to replace damaged components with equivalent components or as provided for in General Certificate No 6B/0.

Signed

Executive Director

#### Descriptive Advice

Pattern: approved 15/2/89

- Asco non-self-indicating weighing instruments of capacities between 15 000 kg and 150 000 kg.

#### Test Procedure

Instruments should be tested in accordance with the relevant sections of the Inspector's Handbook.

The results shall not exceed the maximum permissible errors given in Document 118, 2nd Edition, October 1986.

#### Filing Advice

The documentation for this approval comprises only this Certificate.