



Australian Government

National Standards Commission

12 Lyonpark Road, North Ryde NSW 2113 Australia

Conversion Certificate of Approval

No 6B/237

Issued under Regulation 60
of the
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the

Brimin Model BWB-80 (Converted) Weighing Instrument

submitted by Brimin Engineering Pty Ltd
 RSD 2563 Dugay's Bridge Road
 Rutherglen VIC 3685.

NOTE: This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

CONDITIONS OF APPROVAL

This approval becomes subject to review on 1 February 2009, and then every 5 years thereafter.

Instruments purporting to comply with this approval shall be marked NSC No 6B/237 and only by persons authorised by the submitter.

It is the submitter's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the Commission and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with the Commission's Document NSC P 106.

The values of the performance criteria (maximum number of scale intervals etc.) applicable to the instrument shall be within the limits specified herein and in any approval documentation for the components where they are approved separately.

The Commission reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

Special:

Instruments shall not be manufactured using this approval. Instruments shall only be converted or modified as described herein or with substitute load cells and/or indicator, and in other capacities, in accordance with General Certificate No 6B/0.

DESCRIPTIVE ADVICE

Pattern: approved 28 January 2004

- A converted Brimin model BWB-80 self-indicating weighing instrument. The pattern has been converted by replacement of the indicator and forms an instrument with a verification scale interval (e) of 50 kg and a maximum capacity of 80 000 kg. The instrument is fitted with Eight Precision Transducers model HPC-30 load cells of 30 000 kg (as described in the documentation of NSC approval No S412) and now uses a Ranger model 5000 digital indicator (as described in the documentation of NSC approval No S363).

Variants: approved 28 January 2004

1. With the basework of any Commission-approved mechanical or lever/load cell weighing instrument converted in accordance with General Certificate No 6B/0. (refer Special Condition of Approval)
2. With the indicator of any Commission-approved full load cell weighing instrument replaced by another Commission-approved digital indicator in accordance with General Certificate No 6B/0. (refer Special Condition of Approval)

TEST PROCEDURE

Instruments should be tested in conjunction with any tests specified in the approval documentation for the indicator used, and in accordance with any relevant tests specified in the Uniform Test Procedures.

Maximum Permissible Errors at Verification/Certification

The maximum permissible errors for increasing and decreasing loads on initial verification/certification for loads, m , expressed in verification scale intervals, e , are:

- $\pm 0.5e$ for loads $0 \leq m \leq 500$;
- $\pm 1.0e$ for loads $500 < m \leq 2\,000$; and
- $\pm 1.5e$ for loads $2\,000 < m \leq 10\,000$.

FILING ADVICE

The documentation for this approval comprises only this Certificate.

Signed by a person authorised under Regulation 60 of the National Measurement Regulations 1999 to exercise the powers and functions of the Commission under this Regulation.

A handwritten signature in black ink, appearing to be 'J. H. T.', located to the right of the signature text.