



COMMONWEALTH OF AUSTRALIA

NATIONAL STANDARDS COMMISSION

Weights and Measures  
(National Standards)  
Act 1960-1964

Weights and Measures  
(Patterns of Instruments)  
Regulations

## *Certificate of Approval*

---

CERTIFICATE NUMBER 6/10B/3

*In respect of the pattern of*

Avery Type 4390 BHA Pit-bank Weighbridge and Variant 1.

Submitted and  
manufactured by: W. & T. Avery (Australia) Pty. Ltd.,  
Birmingham Avenue,  
Villawood,  
New South Wales. 2163.

This is to certify that the pattern and variant of the instrument described in this Certificate have been examined by the National Standards Commission under the provisions of the abovementioned Regulations and have been approved as being suitable for use for trade.

Approval was granted for the pattern and variant 1 on 17th November, 1966.

Approval was granted on condition that all instruments made in conformity with the pattern and variant:

1. are appropriately marked NSC No 6/10B/3 and, where required by State legislation, with the State approval number also; and
2. comply with the General Specifications for Weighing and Measuring Instruments to be Used for Trade, in respect of that part of the instrument which was not previously approved by a State.

16/3/70

Cont'd over

This Certificate comprises:

Pages 1 to 3 dated 16th March, 1970.

Pursuant to regulation 12 of the abovementioned Regulations, this Certificate is applicable in the States of Victoria and New South Wales in respect of the pattern and in all States in respect of the variant.

Date of issue 16th March, 1970.

Signed

A handwritten signature in cursive script, appearing to read "Philip A. Mansfield". The signature is written in dark ink and is positioned to the right of the word "Signed".

A person authorised by the Commission  
to sign Certificates under the  
abovementioned Regulations.

16/3/70

---

DESCRIPTION OF PATTERN

The pattern is of a pit-bank weighbridge previously approved in the States of Victoria and New South Wales, modified by fitting a chart graduated in decimal submultiples of a ton and similarly modifying the graduations on the tare bars. The modified instrument has a capacity of 2.8 tons, 2 tons being on the chart, which has 0.005 ton graduations, 0.7 ton on the major tare bar which has 0.1 ton graduations and 0.1 ton on the minor tare bar with 0.005 ton graduations.

DESCRIPTION OF VARIANT

1. Having decimal graduated charts fitted to other instruments approved by a State.

GENERAL NOTES

Notice of approval of the pattern and variant described in this Certificate was given in Memorandum of Approval No 34 dated 17th November, 1966.

\*\*\*  
\*\*\*  
\*\*\*




## NATIONAL STANDARDS COMMISSION

### CERTIFICATE OF APPROVAL No 6/10B/3

This is to certify that Certificate of Approval No 6/10B/3\* dated 16 March 1970 was cancelled on 16 September 1974.

Signed



Executive Officer

---

\* Instruments made in accordance with Certificate No 6/10B/3 should be marked with the approval number "NSC No 6/9C/105" when converted to read in metric units.

24/9/74