

CERTIFICATE OF APPROVAL No 6/4A/6

This is to certify that the patterns of the

Bizerba R3000N Weighing Instrument

CANCELLED

submitted by Globus-Bizerba Pty Ltd,
152 Edinburgh Road,
Marrickville, New South Wales, 2204,

0/1

have been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade.

Date of Approval: 14 November 1974.

The patterns are described in Technical Schedule No 6/4A/6, and in drawings and specifications lodged with the Commission.

The approval is subject to review on or after 1 November 1979.

All instruments conforming to this approval shall be marked with the approval number "NSC No 6/4A/6".

Signed



Executive Officer

43 5

Indexed

14/11/74

CANCELLED

Cert 100

9/1



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No 6/4A/6

Pattern: Bizerba R3000N Weighing Instrument

Submitter: Globus-Bizerba Pty Ltd,
152 Edinburgh Road,
Marrickville, New South Wales, 2204.

Date of Approval: 14 November 1974

All instruments conforming to this approval shall be marked "NSC No 6/4A/6".

Description:

The pattern (see Figure 1) is a partly self-indicating weighing instrument of maximum capacity 3 kg with a centre-zero reading face which indicates a weight difference of up to ± 30 g with a maximum of 3 graduations per degree over not more than $\pm 11^\circ$ deflection.

The lever system is illustrated in Figures 2 and 3.

The instrument is marked "not for retail counter use".


FIGURE 6/4A/6 - 1



Bizerba R3000N

22/5/75


FIGURE 6/4A/6 - 2



Bizerba R3000N

22/5/75

FIGURE 6/4A/6 - 3



Bizerba R3000N — Schematic Diagram

22/5/75