

CERTIFICATE OF APPROVAL No 6/4D/101

VARIATION No 1

This is to certify that the following modification of the pattern and variants of the

Ajax Fan Scale Model AK 102

approved in Certificate No 6/4D/101 dated 14 March 1974

submitted by Broughton Scales Pty Ltd, 72-86 Brenan Street, Leichhardt, N.S.W., 2040,

has been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade.

The approved modification is:

adding an 1800-g capacity by 10-g graduations weight and price chart with unit-price range 2 c/kg to 250 c/kg.


Approval was granted on 13 May 1974.

This variation is described in Technical Schedule No 6/4D/101, Variation No 1, and in drawings and specifications lodged with the Commission.

The approval is subject to review on or after 31 December 1975.

All instruments conforming to this approval shall be marked with the approval number "NSC No 6/4D/101".

Signed



Executive Officer

revised

13/5/74



C/100
23/1/80
NATIONAL STANDARDS COMMISSION

CERTIFICATE OF APPROVAL No 6/4D/101

Approved: Ajax Fan Scale Model AK 102

This is to certify that the pattern and variants of the abovementioned instrument (also known as Austral Model AK 102) have been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade when converted to read in metric units in accordance with Appendix 8 of the General Specifications for Measuring Instruments to be Used for Trade.

Date of Approval: 19 December 1973

Description of Pattern:

The pattern and variants may be identified by reference to the following approval numbers:

Victoria	Approval No 102
Western Australia	Approval No 155

A typical instrument is shown in Figure 1.

Conditions of Approval:

1. This approval is subject to review on or after 31 December 1975 to determine the extent to which the pattern may need to be modified to comply with the Commission's pattern approval requirements at the date of review.
2. All instruments conforming to variants in accordance with this approval shall be marked "NSC No 6/4D/101".

Signed

A person authorised by the Commission to sign Certificates under the abovementioned Regulations.



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No 6/4D/101

VARIATION No 1

Pattern: Ajax Fan Scale Model AK 102

Submittor: Broughton Scales Pty Ltd,
72-86 Brenan Street,
Leichhardt, New South Wales, 2040.

Date of Approval of Variant: 13 May 1974

The modification described in this schedule applies to the pattern and variants described in the following pages and figures of Certificate No 6/4D/101 dated 14 March 1974:

Page 1 dated 14 March 1974


Figure 6/4D/101 - 1 dated 14 March 1974

All instruments conforming to this approval shall be marked "NSC No 6/4D/101".

Description:

This variation approves the fitting of a chart which is graduated to 1800 g by 10-g graduations and a series of price scales for unit prices from 2 c/kg to 250 c/kg (see Figure 2).


FIGURE 6/4D/101 - 1



Ajax Fan Scale Model AK 102

14/3/74

FIGURE 6/4D/101 - 2



22/5/74

2-250-c/kg Chart and Indicator