P.O. BOX 282 NOH 1 H RYDE N.S.W. 2113

Telegrams: Telephone: NATSTANCOM SYDNEY

888 3922

#### CERTIFICATE OF APPROVAL No 6/4D/103

#### VARIATION No 1

This is to certify that the following modification of the pattern of the

Pioneer Self-indicating Counter Machine Model B24

approved in Certificate No 6/4D/103 dated 6 February 1974

submitted by Pioneer Scale Co. Pty Ltd,
Bray Avenue,
Torrensville, South Australia, 5031,

has been approved under the Weights and Measures (Patterns of Instruments) egulations as being suitable for use for trade.

Date of Approval: 11 May 1976

The approved modification, described in Technical Schedule No 6/4D/103 - Variation No 1 and in drawings and specifications lodged with the Commission, provides for an 8-kg capacity by 0,02-kg graduation weight and price chart with unit-price range 2 c/kg to 440 c/kg.

The approval expires on 31 December 1977.

All instruments conforming to this approval shall be marked with the approval number "NSC No 6/4D/103".

Signed

Acting Executive Officer







### NATIONAL STANDARDS COMMISSION

### CERTIFICATE OF APPROVAL No 6/4D/103

Approved: Pioneer Self-indicating Counter Machine Model B24

This is to certify that the pattern and variants of the Pioneer Self-indicating Counter Machine Model B24 have been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade when converted to read in metric units in accordance with Appendix 8 of the General Specifications for Measuring Instruments to be Used for Trade.

Date of Approval: 27 November 1973

#### Description of Pattern:

The pattern and variants may be identified by reference to the following approval numbers:

New South Wales ... Approval No 191

Supplementary Approval Nos 265,

266, 271

South Australia ... Approval Nos 49, 176

Tasmania ... Approval No T/V 92

Victoria ... Approval Nos 92, 178, 220, 590,

602

Western Australia ... Approval No 101

A typical instrument is shown in Figure 1.

## Conditions of Approval:

- 1. This approval is subject to review on or after 31 December 1975 to determine the extent to which the pattern may need to be modified to comply with the Commission's pattern approval requirements at the date of review.
- 2. All instruments conforming to the pattern or variants shall be marked "NSC No 6/4D/103".

Signed
Silver follhornefir

A person authorised by the Commission to sign Certificates under the abovementioned Regulations.



### NATIONAL STANDARDS COMMISSION

# TECHNICAL SCHEDULE No 6/4D/103

#### VARIATION No 1

Pattern: Pioneer Self-indicating Counter Machine Model B24

Submittor:

Pioneer Scale Co. Pty Ltd,

Bray Avenue,

Torrensville, South Australia, 5031.

Date of Approval of Variation: 11 May 1976

The modification described in this Schedule applies to the pattern described in Certificate No 6/4D/103 dated 8 February 1974.

#### Conditions of Approval:

- 1. The approval expires on 31 December 1977.
- 2. All instruments conforming to this approval shall be marked "NSC No 6/4D/103".

### Description:

The approved modification provides for an 8-kg capacity by 0,02-kg graduation weight and price chart with unit-price range 2 c/kg to 440 c/kg.





### NATIONAL STANDARDS COMMISSION

NOTIFICATION OF CHANGE

CERTIFICATE OF APPROVAL No 6/4D/103

CHANGE No 1

The approval of the

Pioneer Self-indicating Counter Machine Model B24

given in Certificate No 6/4D/103 - Variation No 1 dated 21 June 1976

is varied by extending the expiry date to coincide with the latest date at which a non-metric instrument may be used under State regulations.

