



NATIONAL STANDARDS COMMISSION

CERTIFICATE OF APPROVAL No 6/4D/143

Approved: Hobart-Dayton Model 970 Counter Scale

This is to certify that the pattern and variants of the abovementioned instrument have been approved under the Weights and Measures (Patterns of Instruments) Regulations as being suitable for use for trade when:

- (a) fitted with an 8000-g capacity by 20-g graduation weight-and-price chart with unit-price range 1 c/kg to 300 c/kg; and
- (b) converted to read in metric units in accordance with Appendix 8 of the General Specifications for Measuring Instruments to be Used for Trade.

Date of Approval: 14 June 1974

The pattern and variants may be identified by reference to the following approval numbers:

Queensland	Approval Nos 253, 279
New South Wales	Approval No 161
Victoria	Approval Nos 129, 637
South Australia	Approval No 74
Western Australia	Approval No 50

A typical instrument is shown in Figure 1.

Conditions of Approval:


1. This approval is subject to review on or after 31 December 1975 to determine the extent to which the pattern and variants may need to be modified to comply with the Commission's pattern approval requirements at the date of review.
2. All instruments conforming to this approval shall be marked "NSC No 6/4D/143".

Signed

Executive Officer

20/8/75

FIGURE 6/4D/143 - 1



Hobart-Dayton Model 970

20/8/75