National Standards Commission



Certificate of Approval

No 6/4D/264

Issued under Regulation 9
of the
National Measurement (Patterns of Instruments) Regulations

This is to certify that an approval for use for trade has been granted in respect of the

Ishida Model AC-2000 Weighing Instrument

submitted by Ishida Scales Mfg Co. Ltd

44 Sanno-cho, Shogoin Sakyo-ku Kyoto JAPAN.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

J. Birch

CONDITIONS OF APPROVAL

This approval is subject to review on or after 1/6/96. This approval expires in respect of new instruments on 1/6/97.

Instruments purporting to comply with this approval shall be marked NSC No 6/4D/264 and only by persons authorised by the submittor.

It is the submittor's responsibility to ensure that all instruments marked with this approval number are constructed as described in the drawings and specifications lodged with the Commission and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with the Commission's Document 106.

The Commission reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

DESCRIPTIVE ADVICE

Pattern:

approved 21/5/91

 An Ishida model AC-2000 dual-interval self-indicating price-computing weighing instrument of 15 kg maximum capacity.

Technical Schedule No 6/4D/264 describes the pattern.

FILING ADVICE

The documentation for this approval comprises.

Certificate of Approval No 6/4D/264 dated 27/8/91 Technical Schedule No 6/4D/264 dated 27/8/91 (incl. Test Procedure) Figure 1 dated 27/8/91



National Standards Commission

TECHNICAL SCHEDULE No 6/4D/264

Pattern: Ishida Model AC-2000 Weighing Instrument.

Submittor: Ishida Scales Mfg Co. Ltd

44 Sanno-cho, Shogoin Sakyo-ku Kyoto JAPAN.

1. Description of Pattern

An Ishida model AC-2000 dual-interval self-indicating price-computing weighing and label printing instrument (Figure 1) with a verification scale interval of 0.002 kg up to 6 kg and with a verification scale interval of 0.005 kg from 6 kg up to the maximum capacity of 15 kg. Instruments have unit price to \$999.99/kg, price to \$999.99, a price-look-up (PLU) facility, and may be fitted with output sockets for the connection of peripheral and/or auxiliary devices.

1.1 Zero

Zero is automatically corrected to within ±0.25e whenever power is applied and whenever the instrument comes to rest within 0.5e of zero.

The instrument has an initial zero-setting device with a nominal range of 3.2% of the maximum capacity of the instrument.

1.2 Tare

A semi-automatic subtractive taring device of up to 5.998 kg capacity may be fitted.

1.3 Display Check

A display check is initiated when the ZERO button is pressed.

1.4 Levelling

The instrument is provided with adjustable feet and adjacent to the level indicator is a notice advising that the instrument must be level when in use.

1.5 Verification/Certification Provision

Provision is made for a verification/certification mark to be applied.

1.6 Markings

Instruments are marked with the following data, together in one location:

Manufacturer's name or mark Serial number NSC approval number NSC No 6/4D/264 Accuracy class Low range Maximum capacity Max kg * Verification scale interval e = d = kg *High range Maximum capacity kg * Max Verification scale interval e = d = kg'...... kg * Minimum capacity Maximum subtractive tare T = -..... ka -5°C to +40°C Special temperature limits

* Repeated adjacent to each reading face.

TEST PROCEDURE

Instruments should be tested in conjunction with any relevant tests specified in the Inspector's Handbook.

Maximum Permissible Errors at Verification/Certification

The maximum permissible errors for increasing and decreasing loads, expressed in terms of verification scale interval (e), with the instrument adjusted to zero within $\pm 0.25e$ at no load, are:

```
\pm 0.5e for loads from 0 to 500e;
\pm 1.0e for loads over 500e up to 2 000e; and
\pm 1.5e for loads over 2 000e.
```

National Standards Commission



NOTIFICATION OF CHANGE VARIOUS CERTIFICATES OF APPROVAL

The following change is made to the approval documentation for various approvals

submitted by

Ishida Scales Mfg Co. Ltd 44 Sanno-cho, Shogoin

Sakyo-ku Ky

Kyoto JAPAN.

In the Certificates and Technical Schedules for the approvals listed below, the name of the submittor should be changed to read;

"Ishida Co. Ltd" (the address remains unchanged)

NOTE: Only 'current' approvals are listed; not those that have expired or have been cancelled.

APPROVAL NUMBER	PATTERN
6/4C/62 6/4C/73 6/4C/78 6/4C/79	MT-30W Weighing Instrument WL-600SS Weighing Instrument VEGA Weighing Instrument MG-3000 Weighing Instrument
6/4D/228 6/4D/250 6/4D/262 6/4D/263 6/4D/264 6/4D/265	MS-5060SP Weighing Instrument Nova 150 Weighing Instrument Libra-150 Weighing Instrument Libra II Weighing Instrument AC-2000 Weighing Instrument IP-21EX Weighing Instrument
6/9C/227	MT-150W Weighing Instrument
S281	MTC-20 Digital Indicator

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Measuring Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

J. Birk

FIGURE 6/4D/264 - 1

