



NATIONAL STANDARDS COMMISSION

6/4/150
21/8/89

NATIONAL MEASUREMENT (PATTERNS OF INSTRUMENTS) REGULATIONS

REGULATION 9

INSTRUMENT CERTIFICATE OF APPROVAL No 6/4/150

This is to certify that an approval for use for trade has been granted in respect of the pattern of the

Lorch Weighing Instrument.

CONDITIONS OF APPROVAL

Special:

This approval does NOT permit the verification of new instruments of the pattern. It does not become due for review.

This approval applies only to the instrument located at Thermal Quilts, Belconnen ACT, which has previously been verified; it may be re-verified, or verified after repairs, at the discretion of the relevant Weights and Measures authority.

The instrument purporting to comply with this approval shall be marked NSC No 6/4/150 prior to re-verification.

The instrument may be relocated and reinstalled but shall not be altered in any way other than to replace damaged components with equivalent components.

Signed

Executive Director

Descriptive Advice

Pattern: approved 15/2/89

- A Lorch non-self-indicating weighing instrument of 10 kg capacity with steelyards of 7.5 kg and 0.5 kg capacity. The steelyards have verification scale intervals of 0.5 kg and 0.005 kg, respectively.

Test Procedure

The instrument should be tested in accordance with the relevant sections of the Inspector's Handbook. The results shall not exceed the maximum permissible errors given in Document 118, 2nd Edition, October 1986.

Filing Advice

The documentation for this approval comprises only this Certificate.