



NATIONAL STANDARDS COMMISSION
WEIGHTS AND MEASURES (PATTERNS OF INSTRUMENTS) REGULATIONS

REGULATION 9

CERTIFICATE OF APPROVAL No 6/9C/81

This is to certify that an approval has been granted by the Commission that the pattern of the

Mercury Model SB-LP 1200 Platform Weighing Instrument

submitted by Mercury Weighing And Control Systems Pty Ltd
32 Dew Street
Thebarton, South Australia, 5031

is suitable for use for trade.

The approval is subject to review on or after 1/11/88.

Instruments purporting to comply with this approval shall be marked NSC No 6/9C/81.

The approval may be withdrawn if instruments are used other than in accordance with the drawings and specifications lodged with the Commission.

Condition of Approval

The load cells to be used in these instruments shall be subject to regular certification by the Commission.

Signed

Executive Director

Descriptive Advice

Pattern: approved 7/10/83

- Mercury model SB-LP 1200 platform weighing instrument of 1000 kg maximum capacity and with up to 1000 scale intervals.

Filing Advice

The documentation for this approval comprises:

Certificate of Approval No 6/9C/81 dated 28/10/83
Technical Schedule No 6/9C/81 dated 28/10/83
Test Procedure No 6/9C/81 dated 28/10/83
Figure 1 dated 28/10/83.

28/10/83



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No 6/9C/81

Pattern: Mercury Model SB-LP 1200 Platform Weighing Instrument

Submitter: Mercury Weighing And Control Systems Pty Ltd
32 Dew Street
Thebarton, South Australia, 5031.

1. Description of Pattern

A platform weighing instrument (Figure 1) of 1000 kg maximum capacity and with up to 1000 scale intervals.

1.1 Basework

The model SB-LP 1200 basework uses 4 Transducers Inc. model B5112- 1K-0052 450 kg load cells. The basework is provided with 4 adjustable feet attached directly to the load cells and adjacent to the level indicator is a notice advising that the instrument must be level when in use.

1.2 Indicator

The Mercury model 579 digital indicator is described in Technical Schedule No S114 and is approved for use with up to 3072 verification scale intervals.

1.3 Verification

Provision is made for a verification mark to be applied.

1.4 Markings

Instruments are to be marked with the following data, together in a clearly visible location:

Manufacturer's name or mark	
NSC approval number	NSC No 6/9C/81
Serial number	
Accuracy class	(III)
Maximum capacity	Max 1000 kg
Minimum capacity	Min 50 kg
Verification scale interval	$e = d = 1 \text{ kg}$
Load cell serial number(s)	

The following is the minimum data required to be marked on the load cells:

Manufacturer's name or mark
Model number
Serial number
Maximum capacity

TEST PROCEDURE No 6/9C/81

All load applications to the instrument should be in accordance with the Commission's recommended testing procedure for the elimination of rounding error as set out in Document 104.

The maximum permissible errors are:

- $\pm 0.5e$ for loads between 0 and 500e;
- $\pm 1.0e$ for loads between 501e and 1000e.

1. Zero Range

Check that the range of the zero adjustment is not more than 4% of the maximum capacity ($\pm 2\%$ approximately).

2. Zero Test

Check, by means of Document 104, that when the zero light illuminates, zero is set within 0.25e.

3. Range of Indication

- (a) The maximum mass indicated should not exceed the maximum capacity (Max) by more than 10 scale intervals; above this indicated mass the indicator should be blank.
- (b) Below zero the indicator should display the mass prefixed by a minus sign or be blank.

4. Test Loads

Test loads are to be applied to the complete weighing instrument increasing in not less than 5 approximately equal steps to maximum capacity, followed by decreasing loads in not less than 5 approximately equal steps to zero load.

5. Multiple Indicators

Where more than one indicating system is used, the variation between indications or printings for the same load shall not be greater than the absolute value of the maximum permissible error for that load registered on the device with the largest verification scale interval.



25/3/88

NATIONAL STANDARDS COMMISSION

NOTIFICATION OF CHANGE

VARIOUS CERTIFICATES OF APPROVAL

The following changes are made to the approval documentation for the approvals listed overleaf

submitted by Mercury Weighing and Control Systems Pty Ltd
 32 Dew Street
 Thebarton SA 5031.

In the Certificates and Technical Schedules listed, the following changes should be made:

- 1) The submitter should be changed to read;

A & D Mercury Pty Ltd

(the address remains unchanged)

- 2) Any Mercury instrument or component of an instrument approved in the documentation, may now also be known as "AND Mercury" or similar.

Signed

A handwritten signature in cursive script, appearing to read 'J. Birch'.

Executive Director

APPROVAL PATTERN

TYPE: weighing instruments counter scales

6/3/007 Model 92

6/3/008 Model 131

TYPE: counter machines semi-self-indicating

6/4A/012 Model 304A

TYPE: counter machines freely-suspended < 30 kg (spring scales)

6/5/011 Model 211 DA

TYPE: weighing instruments non-self-indicating

6/9A/001 Models 692 and 682

6/9A/004 Model 522D

6/9A/007 Model 211

6/9A/008 Model 600

TYPE: weighing instruments self-indicating

6/9C/005 Model 211D

6/9C/013 Up to 2500 lb or 1200 kg

6/9C/066 Model 522 AL

6/9C/067 Model SM100/479/522D

6/9C/081 Model SB-LP 1200

6/9C/088 Model 522D LT-10K

TYPE: weighbridges self-indicating

6/10B/040 Model WB-LT

6/10B/045A Model RVB-H20

TYPE: automatic weighing instruments (except belt conveyors)

6/14B/012 Model HSD automatic hopper

TYPE: overhead weighing instrument (suspended load or receptor)

6/18/005 With 211DA headwork

6/18/017 Model OHT 500

TYPE: digital indicators

S114 Model 579

S128 Model 1300

S132 Model 900

S161 Model AD4316

S199 Model AD-4321

TYPE: load cells

S117 Interface model SM25-12 kg

S163 Transducers model B5112.1K

S221 HBM model TRT-50 (Mercury model TRT3K-50)



NATIONAL STANDARDS COMMISSION

NOTIFICATION OF CHANGE

CERTIFICATE OF APPROVAL No 6/9C/81

CHANGE No 1

The following change is made to the approval documentation of the

Mercury model SB-LP 1200 Platform Weighing Instrument

submitted by Mercury Weighing And Control Systems Pty Ltd

32 Dew Street

THEBARTON SA 5031.

In Technical Schedule No 6/9C/81 dated 28/10/83, the following should be added to paragraph 1.4 Markings which gives the minimum data to be marked on the load cells;

NSC approval number


NSC No ...

Signed

Executive Director

23/12/83

FIGURE 6/9C/81 - 1



Mercury Model SB-LP 1200