

NATIONAL STANDARDS COMMISSION

NATIONAL MEASUREMENT (PATTERNS OF INSTRUMENTS) REGULATIONS

REGULATION 9

SUPPLEMENTARY CERTIFICATE OF APPROVAL No S195

This is to certify that an approval for use for trade has been granted in respect of the pattern and variants of the

Transducers Model B5352-200-10M1 Load Cell

submitted by Transducers Inc. 14030 Bolsa Lane Cerritos California 90701 USA.

Conditions of Approval

This approval is subject to review on or after 1/7/90.

Load cells purporting to comply with this approval shall be marked NSC No S195. Instruments incorporating a load cell purporting to comply with this approval shall be marked NSC No S195 in addition to the approval number of the instrument.

This approval may be withdrawn if load cells are constructed and used other than in accordance with the drawings and specifications lodged with the Commission.

The number of scale intervals applicable to the instrument shall be no greater than the number of verification scale intervals approved for the basework or the load cell(s) or the indicator, whichever is the smallest.

The load cells used shall be subject to regular certification by the Commission.

Signed

Executive Director

Descriptive Advice

Pattern: approved 21/6/85

. Transducer model B5352-200-10M1 load cell of 200 kg capacity.

Variant: approved 21/6/85

1. With various length cable.

Technical Schedule No S195 describes the pattern and variant 1.

..../2

S195 29/8/86

S195 29/8/86

Page 2

Variant: approved 27/5/86

2. Of 1000 kg capacity.

Variant: approved 2/7/86

3. Of 500 kg capacity.

Technical Schedule No S195 Variation No 1 describes variants 2 and 3.

Filing Advice

Supplementary Certificate of Approval No S195 dated 1/11/85 is superseded by this Certificate and may be destroyed. Table 1 dated 1/11/85 is replaced by the Table included herein.

Note: Paragraph <u>1.1 Method of Mounting</u> in Technical Schedule No S195 dated 1/11/85, should be amended to read, in part: ".... the methods shown in Figures 2 to 5".

The documentation for this approval now comprises:

Supplementary Certificate of Approval No S195 dated 29/8/86 Technical Schedule No S195 dated 1/11/85 Technical Schedule No S195 Variation No 1 dated 29/8/86 (including Table 1) Figures 1 to 3 dated 1/11/85 Figures 4 and 5 dated 29/8/86



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No S195

Pattern: Transducers Model B5352-200-10M1 Load Cell

Submittor: Transducers Inc. 14030 Bolsa Lane Cerritos California 90701 USA

1. Description of Pattern

The pattern is a Transducers model B5352-200-10M1 load cell of 200 kg capacity (refer Figure 1 and Table 1).

1.1 Method of Mounting

Mounting is to be in accordance with one of the methods shown in Figures 2 and 3. The method shown in Figure 3 is for use in a Commission-approved overhead-track weighing instrument.

1.2 Marking

The following is the minimum data required to be marked on the load cell:

Manufacturer's name or mark Model number Serial number NSC approval number Maximum rated capacity

NSC No S195

.

2. Description of Variant 1

With various length cable as designated by the model number viz. B5352-200-**M1 where ** multiplied by 0.3 is the cable length in metres.

TABLE 1

Type: Transducers	B5352-200-10M1
Maximum capacity	200 kg
Maximum number of	(a) 6000
verification	(b) 6000
scale intervals	(c) 3000
	(d) 3000
Minimum value of	(a) 0 . 01 kg
verification	(b) 0.01 kg
scale interval	(c) 0.05 kg
	(d) 0.05 kg
Output rating (nominal)	2.0 mV/V
Input impedance (nominal)	350 ohms
Supply voltage (AC or DC)	10-15 V
Cable length (± 0.1 m)	(refer Variant 1)
Number of leads	5 (including shield)

(a) Instruments with automatic zero track - multi cell applications
(b) Instruments with automatic zero track - single cell applications
(c) Instruments without automatic zero track - multi cell applications

(d) Instruments without automatic zero track - single cell applications



NATIONAL STANDARDS COMMISSION

TECHNICAL SCHEDULE No S195

VARIATION No 1

Pattern: Transducers Model B5352-200-10M1 Load Cell

Submittor: 14030 Bolsa Lane Cerritos California 90701 USA

1. Description of Variants

- 1.1 Variant 2

A Transducers model B5352-1000-10M1 load cell of 1000 kg capacity (refer Table 1).

1.2 Variant 3

A Transducers model B5352-500-10M1 load cell of 500 kg capacity (refer Table 1).

S195 29/8/86

TABLE 1	

Note: This Table replaces Table 1 included in Technical Schedule No S195 dated 1/11/85.

Type: Transducers Moximum cabacity		B5352-200-10M1 200 ka		85352-1000- 1000 kg	1 OM 1	ងខ្លួ	352-500-1 0M1 30 kg	
Maximum number of (c	(D	0009	(o)	5500	0	200	8	
verification (t	ĥ	0009	<u>م</u>	5500	<u>е</u>	20	8	
scale intervals (4	()	3000	(°)	5000	0	200	8	
	Ð	3000	(P)	4000	р)	, 40	8	
Minimum value of (c	() ()	0.01 kg	(o	0.06 kg	<u> </u>	0.0	03 kg	
verification (i	<u>م</u>	0.01 kg	(q)	0.10 kg	9 9	。 。	05 kg	
scale interval (c	.	0.05 kg	()	0.15 kg	<u></u>		D6 kg	
	Ð	0.05 kg	(p)	0.20 kg	P	。 。	lo kg	
Output rating (nominal)		2.0 mV/V		2.0 mV/V		N	.0 mV/V	
Input impedance (nominal		350 ohms		350 ohms		ы С	o chimis	
Supply voltage (AC or DC	_	10-15 V		10-15 V		<u>-</u>	-15 V	
Cable length (+ 0.1 m) Number of leads	Ë	efer Variant 1)	<u>ц</u>	efer Variant	-	(refe	r Variant 1)	
(including shield)		Ŋ		ß		ß		

Instruments with automatic zero track - multi cell applications Instruments with automatic zero track - single cell applications Instruments without automatic zero track - multi cell applications Instruments without automatic zero track - single cell applications

0000

S195 29/8/86

National Standards Commission



NOTIFICATION OF CHANGE

SUPPLEMENTARY CERTIFICATE OF APPROVAL No S195

CHANGE No 1

The following change is made to the approval documentation for the

Transducers Model B5352-200-10M1 Load Cell

submitted by Transducers Inc. 14030 Bolsa Lane Cerritos California 90701 USA.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

4

Lain,

In Supplementary Certificate No S195 dated 29/8/86, the following Condition of Approval should be added:

"This approval expires in respect of new instruments on 1/7/91."

NOTE: This approval was incorrectly cancelled as part of Cancellation Certificate No 0/3 dated 31/12/90.



FIGURE S195 - 1



FIGURE S195 - 2



FIGURE S195 - 3







S195 29/8/86

Alternative Mounting Method



FIGURE S195-4







S195 29/8/86