

Supplementary Certificate of Approval No S263

Issued under Regulation 9 of the National Measurement (Patterns of Instruments) Regulations

This is to certify that an approval for use for trade has been granted in respect of the

Philips Model PR6222/34H Load Cell

submitted by Philips Scientific & Industrial Pty Ltd 25 Paul Street North North Ryde NSW 2113.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

J. Binh

S263 27/4/90

CONDITIONS OF APPROVAL

This approval is subject to review on or after 1/3/95. This approval expires in respect of new instruments on 1/3/96.

Instruments purporting to comply with this approval shall be marked NSC No S263 and only by persons authorised by the submittor. Instruments incorporating a load cell purporting to comply with this approval shall be marked NSC No S263 in addition to the approval number of the instrument.

This approval may be withdrawn if load cells are constructed other than in accordance with the drawings and specifications lodged with the Commission and/or perform other than in accordance with these approval documents.

The values of the performance criteria (maximum number of scale intervals etc.) applicable to an instrument incorporating these load cells shall be within the limits specified in this approval and in any approval documentation for the other components, excepting any limitations imposed by mechanical indicators on mechanical baseworks in such documentation.

The Commission reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

DESCRIPTIVE ADVICE

Pattern: approved 19/2/90

- A Philips model PR6222/34H load cell of 30 t maximum capacity.
- Technical Schedule No S263 describes the pattern.

FILING ADVICE

The documentation for this approval comprises:

Supplementary Certificate of Approval No S263 dated 27/4/90 Technical Schedule No S263 dated 27/4/90 (incl. Table 1) Figures 1 to 4 dated 27/4/90



TECHNICAL SCHEDULE No S263

- Pattern: Philips Model PR6222/34H Load Cell.
- Submittor: Philips Scientific & Industrial Pty Ltd 25 Paul Street North North Ryde NSW 2113.

1. Description of Pattern

The pattern is a Philips model PR6222/34H load cell of 30 t capacity (refer Figure 1 and Table 1) approved for use with a maximum of 5500 verification scale intervals.

1.1 Method of Mounting

Mounting is to be in accordance with the manufacturer's instructions and in accordance with one of the methods shown in Figures 2 to 4.

1.3 Markings

The following is the minimum data required to be marked on the load cells:

Manufacturer's name or mark	
Model number	
Serlai number	
NSC approval number	NSC No S263
Maximum rated capacity	

TABLE 1

Type: Philips model PR6222/34H Maximum capacity Maximum number of verification scale intervals	(a) (b)	30 5500 5500	t
Minimum value of verification scale interval	(a) (b)	0.95 2.36	kg kg
Output rating (nominal) Input impedance (nominal) Supply voltage (AC or DC) Number of leads (plus shield) Cable length (<u>+</u> 0.1m)	4	1.0 650 1-24 4 12	mV/V ohms V (★) m

- (a) Instruments with automatic zero track.
- (b) Instruments without automatic zero track.
- (*) Recommended voltage range; maximum supply voltage 32 V.

S263 11/10/91

National Standards Commission



NOTIFICATION OF CHANGE

SUPPLEMENTARY CERTIFICATE OF APPROVAL No S263

CHANGE No 1

The following changes are made to the approval documentation for the

Philips Model PR6222/34H Load Cell

submitted by Philips Scientific & Industrial Pty Ltd 25 Paul Street North NORTH RYDE NSW 2113.

In Technical Schedule No S263 dated 27/4/90, Table 1 is replaced by the attached Table 1, in which the data relating to the verification scale interval has been amended.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.

lacen



TECHNICAL SCHEDULE No S263

TABLE 1

. The sum he was a second seco		
Maximum capacity Maximum number of verification scale intervals	30 (a) 3000 (b) 3000	t
Minimum value of verification scale interval	(a) 0.622 (b) 1.55	kg kg
Output rating (nominal) Input impedance (nominal) Supply voltage (AC or DC) Number of leads (plus shield) Cable length (±0.1m)	1.0 650 4-24 4 12	mV/V ohms V (*) m

(a) Instruments with automatic zero track.

Type: Philips model PR6222/34H

(b) Instruments without automatic zero track.

(*) Recommended voltage range; maximum supply voltage 32 V.

NOTE: This Table replaces Table 1 issued as part of Technical Schedule No S263 dated 27/4/90.



Notification of Change

Supplementary Certificate of Approval No S263

Change No 2

The following changes are made to the approval documentation for the

Philips Model PR6222/34H Load Cell

- submitted by Philips Scientific & Industrial Pty Ltd 25 Paul Street North Ryde NSW 2113.
- 1. In Supplementary Certificate of Approval No S263 and its Technical Schedule both dated 27 April 1990;
- (i) All references to the submittor should be amended to read:

"GWT GLOBAL Weighing Technologies GmbH Meiendorfer Strasse 205 22145 Hamburg GERMANY."

- (ii) All references to "Philips" load cells should be amended to read "Philips or GLOBAL Weighing" load cells.
- NOTE: This approval expired in respect of new instruments on 1 March 1996.

Signed and sealed by a person authorised under Regulation 9 of the National Measurement (Patterns of Measuring Instruments) Regulations to exercise the powers and functions of the Commission under this Regulation.





Philips PR6222 Lood Cell



Figure S263 - 3



