



Bradfield Road, West Lindfield NSW 2070

# **Cancellation**

## **Supplementary Certificate of Approval**

### **NMI S457**

Issued by the Chief Metrologist under Regulation 60  
of the  
*National Measurement Regulations 1999*

This is to certify that the approval for use for trade granted in respect of the

Avery Weigh-Tronix Model E1010 Digital Indicator

submitted by      Avery Weigh-Tronix Ltd  
                         Foundry Lane  
                         Smethwick  
                         West Midlands      B66 2LP      UK

has been cancelled in respect of new instruments as from 1 August 2015.

#### **DOCUMENT HISTORY**

<b>Rev</b>	<b>Reason/Details</b>	<b>Date</b>
0	Pattern & variant 1 approved – interim certificate issued	17/06/05
1	Pattern & variant 1 approved – certificate issued	23/06/05
2	Pattern & variant 1 reviewed– notification of change issued	14/01/11
3	Pattern & variant 1 cancelled – cancellation certificate issued	6/07/15

Signed by a person authorised by the Chief Metrologist  
to exercise their powers under Regulation 60 of the  
*National Measurement Regulations 1999*.

A handwritten signature in black ink, appearing to be "Dr A Rawlinson".

**Dr A Rawlinson**



**Australian Government**  
**National Measurement  
Institute**

12 Lyonpark Road, North Ryde NSW 2113

**Supplementary Certificate of Approval**  
**No S457**

Issued by the Chief Metrologist under Regulation 60  
of the  
National Measurement Regulations 1999

This is to certify that an approval for use for trade has been granted in respect of the  
Avery Weigh-Tronix Model E1010 Digital Indicator

submitted by Avery Weigh-Tronix Ltd  
Foundry Lane  
Smethwick  
West Midlands B66 2LP UK.

**NOTE:** This Certificate relates to the suitability of the pattern of the instrument for use for trade only in respect of its metrological characteristics. This Certificate does not constitute or imply any guarantee of compliance by the manufacturer or any other person with any requirements regarding safety.

**CONDITIONS OF APPROVAL**

This approval becomes subject to review on 1 July 2010 and then every 5 years thereafter.

Instruments purporting to comply with this approval shall be marked with approval number 'NMI S457' and only by persons authorised by the submitter.

Instruments incorporating a component purporting to comply with this approval shall be marked 'NMI S457' in addition to the approval number of the instrument.

It is the submitter's responsibility to ensure that all instruments marked with this approval number are constructed as described in the documentation lodged with the National Measurement Institute (NMI) and with the relevant Certificate of Approval and Technical Schedule. Failure to comply with this Condition may attract penalties under Section 19B of the National Measurement Act and may result in cancellation or withdrawal of the approval, in accordance with document NMI P 106.

The National Measurement Institute reserves the right to examine any instrument or component of an instrument purporting to comply with this approval.

Auxiliary devices used with this instrument shall comply with the requirements of General Supplementary Certificate No S1/0/A.

The values of the performance criteria (maximum number of scale intervals etc.) applicable to an instrument incorporating the pattern approved herein shall be within the limits specified herein and in any approval documentation for the other components.

#### DESCRIPTIVE ADVICE

**Pattern:** approved 17 June 2005

- An Avery Weigh-Tronix model E1010 digital mass indicator.

**Variant:** approved 17 June 2005

1. An Avery Weigh-Tronix model E1005 digital mass indicator.

Technical Schedule No S457 describes the pattern and variant 1.

#### FILING ADVICE

The documentation for this approval comprises:

Supplementary Certificate of Approval No S457 dated 23 June 2005  
Technical Schedule No S457 dated 23 June 2005 (incl. Table 1 and Test Procedure)  
Figures 1 and 2 dated 23 June 2005

Signed by a person authorised by the Chief Metrologist to exercise his powers under Regulation 60 of the National Measurement Regulations 1999.



TECHNICAL SCHEDULE No S457

**Pattern:** Avery Weigh-Tronix Model E1010 Digital Indicator

**Submittor:** Avery Weigh-Tronix Ltd  
Foundry Lane  
Smethwick  
West Midlands B66 2LP UK

## 1. Description of Pattern

An Avery Weigh-Tronix model E1010 digital mass indicator (Table 1 and Figure 1) which is of stainless steel housing construction and which may be configured to form part of:

- A weighing instrument with a single weighing range of up to 10 000 verification scale intervals; or
- A multi-interval weighing instrument with up to two partial weighing ranges (each with its own verification scale interval) in which case it is approved for use with up to 5000 verification scale intervals per partial weighing range.

Instruments may be fitted with output sockets (output interfacing capability) for the connection of auxiliary and/or peripheral devices.

Instruments are powered directly by mains AC power adaptor, or by an internal rechargeable battery.

This approval does not include the use of the indicator as an automatic weighing instrument, unless specifically mentioned in a certificate of approval for such an instrument.

TABLE 1 – Specifications

Maximum number of verification scale intervals	10 000 or 5000 per range
Minimum sensitivity	0.5 $\mu$ V / scale interval
Excitation voltage	5 V DC
Maximum excitation current	57.5 mA

### 1.1 Zero

Zero is automatically corrected to within  $\pm 0.25e$  whenever the instrument comes to rest within  $0.5e$  of zero

The instrument has a semi-automatic zero setting device (to set the instrument to within  $\pm 0.25e$  of zero) with a nominal range of not more than 4% of the maximum capacity of the instrument.

### 1.2 Tare

A semi-automatic subtractive taring device of up to the maximum capacity of the instrument may be fitted. A pre-set taring device (keyboard-entered and/or stored) of up to the maximum capacity (or of up to the  $Max_1$  (the maximum capacity of the lower interval range) for multi-interval instruments) may also be fitted.

### 1.3 Display Check

A display check is initiated whenever power is applied.

### 1.4 Linearisation Facility

Instruments are fitted with a programmable multi-point linearisation correction facility.


### 1.5 Additional Features

The indicator may also have certain additional functions including setpoints ('cut-offs') and 'under/accept/over' functions. Some functions can be assigned to a function key of the indicator. However, this approval does not include the use of the indicator as an automatic weighing instrument, unless specifically mentioned in a certificate of approval for such an instrument.

The additional functions (other than the indications of measured mass, i.e. gross, tare, net, totals, displayed either on the indicator or on an auxiliary or peripheral device) are not approved for trade use.

### 1.6 Markings and Notices

Instruments carry the following markings:

Manufacturer's mark, or name written in full	Avery Weigh-Tronix Ltd	
Name or mark of manufacturer's agent	.....	
Indication of accuracy class		
Maximum capacity	<i>Max</i> ..... kg	#1
Minimum capacity	<i>Min</i> ..... kg	#1
Verification scale interval	<i>e</i> = ..... kg	#1
Maximum subtractive tare	<i>T</i> = - ..... kg	#2
Serial number of the instrument	.....	
Pattern approval mark for the indicator	NMI S457	
Pattern approval mark for other components	.....	#3
#1	These markings are also shown near the display of the result if they are not already located there.	
#2	This marking is required if <i>T</i> is not equal to <i>Max</i> .	
#3	May be located separately from the other markings.	

In addition, instruments not greater than 100 kg capacity shall carry a notice stating NOT TO BE USED FOR TRADING DIRECT WITH THE PUBLIC, or similar wording.

Note:

For multi-interval instruments the markings shall be as above, with the exception that the 'Maximum capacity' and 'Verification scale interval' shall be marked for both interval ranges, e.g. as follows:

Maximum capacity	<i>Max</i> ...../..... kg	#1
Verification scale interval	<i>e</i> = ...../..... kg	#1

## **1.7 Verification/Certification Provision**

Provision is made for the application of a verification/certification mark.

## **1.8 Sealing Provision**

Access to the configuration and calibration facility is password protected (note that the passwords required to configure or calibrate the instrument are not the same as the user password shown in the instructions below). The indicator increments a configuration and/or calibration value (audit trail number) each time the indicator is re-configured and/or calibrated.

The value(s) of these counters may be recorded on a destructible adhesive label attached to the instrument (e.g. as CFG x , CAL y).

Any subsequent alteration to the calibration or configuration will be evident as the recorded values and the current counter values will differ.

The current value can be displayed by using the following sequence (starting from the normal weighing mode):

- enter the user mode by holding the 'Esc' key for approximately 5 seconds.
- Enter the user password '111' and press the 'Enter' key.
- Press the 'Units' key to enter the Audit mode.
- Press the 'Print' key to enter the Configuration Counter display mode.
- Press the 'Print' key. The value displayed for a short period is the configuration counter value (CFG).
- Press the 'Units' key to enter the Calibration Counter display mode.
- Press the 'Print' key. The value displayed for a short period is the calibration counter value (CAL).
- Press 'ESC' twice to return to the normal weighing mode.

## **2. Description of Variant 1**

The Avery Weigh-Tronix model E1005 (Figure 2), which is similar to the pattern (model E1010), and has the same specifications (Table 1), but which has a plastic housing and a reduced set of functions and operator keys. In particular the model E1005 does not have the capability for pre-set tare (keyboard-entered or stored).

## 2.1 Sealing

The instructions for accessing the configuration and calibration audit trail differ from those of the model E1010, and are as follows (starting from the normal weighing mode):

- enter the user mode by holding the 'Zero' key for approximately 5 seconds.
- Enter the user password '111' by the following set of keystrokes – press 'Select' twice, 'Units', 'Select' twice, 'Units', 'Select' twice and then press the 'F1' key.
- Press the 'Units' key to enter the Audit mode.
- Press the 'Print' key to enter the Configuration Counter display mode.
- Press the 'Print' key. The value displayed for a short period is the configuration counter value (CFG).
- Press the 'Units' key to enter the Calibration Counter display mode.
- Press the 'Print' key. The value displayed for a short period is the calibration counter value (CAL).
- Press the 'Zero' key to return to the normal weighing mode.

## TEST PROCEDURE

Instruments should be tested in accordance with any relevant tests specified in the Uniform Test Procedures.

### Maximum Permissible Errors at Verification/Certification

For single range instruments, the maximum permissible errors for increasing and decreasing loads on initial verification/certification for loads,  $m$ , expressed in verification scale intervals,  $e$ , are:

- $\pm 0.5e$  for loads  $0 \leq m \leq 500$ ;
- $\pm 1.0e$  for loads  $500 < m \leq 2\,000$ ; and
- $\pm 1.5e$  for loads  $2\,000 < m \leq 10\,000$ .

For multi-interval instruments with verification scale intervals of  $e_1, e_2 \dots$ , apply  $e_1$  for zero adjustment, and maximum permissible errors apply  $e_1, e_2 \dots$ , as applicable for the load.



Australian Government

National Measurement  
Institute

Bradfield Road, West Lindfield NSW 2070

## **Notification of Change**

### **Supplementary Certificate of Approval No S457**

### **Change No 1**

Issued by the Chief Metrologist under Regulation 60  
of the  
*National Measurement Regulations 1999*

The following changes are made to the approval documentation for the  
Avery Weigh-Tronix Model E1010 Digital Indicator

submitted by      Avery Weigh-Tronix Ltd  
                         Foundry Lane  
                         Smethwick  
                         West Midlands    B66 2LP    UK.

- A.    In Supplementary Certificate of Approval No S457 dated 23 June 2005;
1.    The Condition of Approval referring to the review of the approval should be amended to read:
- “This approval becomes subject to review on 1 July **2015**, and then every 5 years thereafter.”
2.    The FILING ADVICE should be amended by adding the following:
- “Notification of Change No 1 dated 14 January 2011”

Signed by a person authorised by the Chief Metrologist  
to exercise his powers under Regulation 60 of the  
*National Measurement Regulations 1999*.

A handwritten signature in black ink, consisting of a series of loops and a long horizontal stroke at the bottom.



FIGURE S457 – 1



Avery Weigh-Tronix Model E1010 Digital Indicator

S457  
23 June 2005

FIGURE S457 – 2



Avery Weigh-Tronix Model E1005 Digital Indicator