




REFERENCE MATERIAL PRODUCT INFORMATION SHEET

NMIA D916b: Methyldienolone

Report ID: D916b.2020.02 (Ampouled 170727)

Chemical Formula: C₁₉H₂₆O₂

Molecular Weight: 286.4 g/mol



Property value

Batch No.	CAS No.	Purity estimate by HPLC-UV
16-S-11	14531-89-6	82.9 ± 2.4%

The uncertainty is based on the standard deviation of five analyses in duplicate and has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit ($k = 2$).

IUPAC name: 17 β -Hydroxy-17 α -methyl-estra-4,9-dien-3-one

Expiration of certification: The property values are valid till 18 March 2024, i.e. three years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to ampoules that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials. The material will be re-tested on an annual basis to ensure that the property values are still valid. In the event a product fails the stability trial, notification will be sent to all impacted customers.

Description: Material was prepared by synthesis, and certified for identity and purity by NMIA. The compound is supplied as a dried aliquot in a sealed ampoule and is intended for a single use to prepare a standard solution. The ampoules were prepared by dispensing 1006 ± 2 μ g of D916b. Based on the current purity estimate this equates to 824 ± 54 μ g of methyldienolone.

Intended use: This reference material should be used for qualitative analysis only.

Instructions for use: Open the ampoule and carefully rinse the interior at least three times with a suitable organic solvent (e.g. toluene and one drop of methanol).

Recommended storage: When not in use, this material should be stored at or below 4 °C in a closed container in a dry, dark area.

Metrological traceability: Metrological traceability of the stated purity value to the SI unit for mass (kg) has not been established. This is not a certified reference material, and should not be considered for calibration.

Stability: The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by GC-FID on seven randomly selected ampoules of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Safety: Treat as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,
Team Leader,
Chemical Reference Materials, NMI.
15 November 2022

This report supersedes any issued prior to 15 November 2022.

NATA Accreditation No. 198 / Corporate Site No. 14214.

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

Characterisation Report:

GC-FID:	Instrument:	Agilent 6890
	Column:	HP-1, 30 m × 0.32 mm I.D. × 0.25 µm
	Program:	180 °C (1 min), 30 °C/min to 250 °C (10 min), 30 °C/min to 300 °C (3 min)
	Injector:	200 °C
	Carrier:	Helium
		Detector Temp: 320 °C
		Split ratio: 20/
	Relative peak area of the main component:	
	Initial analysis:	Mean = 98.4%, s = 0.3% (7 ampoules in duplicate, August 2017)
	Re-analysis:	Mean = 92.2%, s = 0.3% (5 ampoules in duplicate, August 2018)
	Re-analysis:	Mean = 92.5%, s = 0.1% (5 ampoules in duplicate, January 2019)
	Re-analysis:	Mean = 83.0%, s = 1.6% (5 ampoules in duplicate, March 2020)
	Re-analysis:	Mean = 82.8%, s = 1.7% (5 ampoules in duplicate, February 2021)
HPLC:	Instrument:	Thermo Scientific Ultimate 3000 RS pump, RS autosampler
	Column:	Alltima C-18, 5 µm (4.6 mm x 150 mm)
	Column oven:	40 °C
	Mobile Phase:	Acetonitrile/MilliQ water (45:55 v/v)
	Flow rate:	1.0 mL/min
	Detector:	RS Diode Array Detector operating at 310 nm
	Relative peak area of the main component:	
	Initial analysis:	Mean = 99.5%, s = 0.01% (7 sub samples in duplicate, December 2016)
	Re-analysis:	Mean = 82.9%, s = 1.2% (9 ampoules in duplicate, March 2021)

The following analytical data was obtained on the bulk material subsequently used in the preparation of the ampoules.

Characterisation Report:

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and ¹H NMR spectroscopy. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

I_{ORG} = Organic impurities of related structure, I_{VOL} = volatile impurities, I_{NVR} = non-volatile residue.

Supporting evidence is provided by HPLC with UV detection, qualitative headspace GC-MS analysis of occluded solvents and elemental microanalysis.

GC-FID:	Instrument:	Agilent 6890
	Column:	HP-1, 30 m × 0.32 mm I.D. × 0.25 µm
	Program:	180 °C (1 min), 30 °C/min to 250 °C (10 min), 30 °C/min to 300 °C (1 min)
	Injector:	250 °C
	Carrier:	Helium
		Detector Temp: 320 °C
		Split ratio: 20/1
	Relative peak area of the main component:	
	Initial analysis:	Mean = 99.0%, s = 0.03% (7 sub samples in duplicate, November 2016)
HPLC:	Instrument:	Thermo Scientific Ultimate 3000 RS pump, RS autosampler
	Column:	Alltima C-18, 5 µm (4.6 mm x 150 mm)
	Column oven:	40 °C
	Mobile Phase:	Acetonitrile/MilliQ water (45:55 v/v)
	Flow rate:	1.0 mL/min
	Detector:	RS Diode Array Detector operating at 310 nm
	Relative peak area of the main component:	
	Initial analysis:	Mean = 99.5%, s = 0.01% (7 sub samples in duplicate, December 2016)
Thermogravimetric analysis:	Volatile content 0.2% and non volatile residue < 0.2% mass fraction	
Karl Fischer analysis:	Moisture content 0.2% mass fraction (November 2016)	

Spectroscopic and other characterisation data

ESI-MS:	Instrument:	Micromass Quatro Micro
	Operation:	Positive ion mode, direct infusion at 2 $\mu\text{L}/\text{min}$
	Ionisation:	ESI spray voltage at 3.2 kV positive ion
	EM voltage:	500 V
	Cone voltage:	30 V
	Peak:	287 (M+H ⁺) <i>m/z</i>
GC-MS:	Parent compound:	
	Instrument:	Agilent 6890/5973
	Column:	HP Ultra 2, 17 m \times 0.20 mm I.D. \times 0.25 μm
	Program:	80 $^{\circ}\text{C}$ (0.5 min), 50 $^{\circ}\text{C}$ to 200 $^{\circ}\text{C}$ (0.5 min), 15 $^{\circ}\text{C}$ to 310 $^{\circ}\text{C}$ (5 min)
	Injector:	290 $^{\circ}\text{C}$
	Transfer line temp:	300 $^{\circ}\text{C}$
	Carrier:	Helium, 1.0 mL/min
	Split ratio:	30/1
	<i>Bis</i> -TMS derivative:	
	Instrument:	Agilent 6890/5973
	Column:	HP Ultra 1, 17 m \times 0.22 mm I.D. \times 0.11 μm
	Program:	180 $^{\circ}\text{C}$ (0.5 min), 12 $^{\circ}\text{C}/\text{min}$ to 310 $^{\circ}\text{C}$ (3 min)
	Injector:	260 $^{\circ}\text{C}$
	Transfer line temp:	300 $^{\circ}\text{C}$
	Carrier:	Helium, 1.0 mL/min
	Split ratio:	30/1
	The retention times of the parent compound and <i>bis</i> -TMS derivative are reported with the major peaks in the mass spectra. The latter are reported as mass to charge ratios and (in brackets) as a percentage relative to the base peak.	
	Parent (7.7 min):	286 (M ⁺ , 100), 268 (21), 253 (35), 228 (91), 215 (88), 199 (22), 186 (14), 174 (24), 160 (65), 129 (29), 117 (38), 91 (45), 77 (20), 55 (14), 43 (48) <i>m/z</i>
	<i>Bis</i> -TMS (13.5 min):	430 (M ⁺ , 100), 415 (7), 325 (13), 285 (24), 246 (6), 143 (3), 73 (23) <i>m/z</i>
HS-GC-MS:	Instrument:	Agilent 6890/5973/G1888
	Column:	DB-624, 30 m \times 0.25 mm I.D. \times 1.4 μm
	Program:	50 $^{\circ}\text{C}$ (5 min), 7 $^{\circ}\text{C}/\text{min}$ to 120 $^{\circ}\text{C}$, 15 $^{\circ}\text{C}/\text{min}$ to 220 $^{\circ}\text{C}$ (8.3 min)
	Injector:	150 $^{\circ}\text{C}$
	Transfer line temp:	280 $^{\circ}\text{C}$
	Carrier:	Helium, 1.2 mL/min
	Split ratio:	50/1
	Solvents detected:	Benzene
IR:	Instrument:	Biorad FTS300MX FT-IR
	Range:	4000-400 cm^{-1} , KBr pellet
	Peaks:	3498, 2941, 2862, 1644, 1608, 1450, 1388, 1278, 1215, 1166, 1069, 955, 856, 728 cm^{-1}
¹ H NMR:	Instrument:	Bruker DMX600
	Field strength:	600 MHz
	Solvent:	CDCl_3 (7.26 ppm)
	Spectral data:	δ 1.01 (3H, s), 1.21 (3H, s), 1.23-1.28 (1H, m), 1.32-1.44 (3H, m), 1.58-1.66 (2H, m), 1.74-1.81 (1H, m), 1.85-1.93 (2H, m), 2.12 (1H, bm), 2.25 (1H, bm), 2.32-2.47 (4H, m), 2.53 (1H, m), 2.82 (1H, dd, $J = 16.1, 2.5$ Hz), 2.89 (1H, ddd, $J = 14.6, 4.9, 4.9$ Hz), 5.66 (1H, s) ppm
	Benzene estimated at 1% mass fraction was observed in the ¹ H NMR	
¹³ C NMR:	Instrument:	Bruker DMX600
	Field strength:	150 MHz
	Solvent:	CDCl_3 (77 ppm)
	Spectral data:	δ 13.3, 23.3, 25.7, 25.8, 25.8, 27.3, 30.9, 31.5, 37.0, 38.9, 40.2, 45.3, 50.9, 81.3, 122.1, 125.5, 146.3, 157.3, 199.8 ppm
Melting point:	159-161 $^{\circ}\text{C}$	
Microanalysis:	Found:	C = 79.6%; H = 9.2% (November 2016)
	Calculated:	C = 79.7%; H = 9.2% (Calculated for $\text{C}_{19}\text{H}_{26}\text{O}_2$ containing 1% benzene and 0.2% water)