




CERTIFIED REFERENCE MATERIAL CERTIFICATE OF ANALYSIS

NMIA D975: Naphyrone hydrochloride

Report ID: D975.2014.03

Chemical Formula: C₁₉H₂₃NO.HCl

Molecular Weight: 317.9 g/mol (HCl), 281.4 g/mol (base)



Certified value

Batch No.	CAS No.	Purity (mass fraction)
11-D-26	850352-11-3 (HCl) 850352-53-3 (base)	98.1 ± 1.7%

The uncertainty has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit ($k = 2$).

IUPAC name: 1-(2-Naphthalenyl)-2-(1-pyrrolidinyl)-1-pentanone hydrochloride

Expiration of certification: The property values are valid till 21st October 2017, i.e. three years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

Description: Off-white powder sourced from an external supplier, certified for identity and purity by NMIA. Packaged in amber glass bottles with a septum and crimped aluminium cap or screw top cap.

Intended use: This certified reference material is suitable for use as a primary calibrator.

Instructions for use: Equilibrate the bottled material to room temperature before opening.

Recommended storage: When not in use this material should be stored at or below 25 °C in a closed container in a dry, dark area.

Metrological traceability: The certified purity value is traceable to the SI unit for mass (kg) through Australian national standards via balance calibration. In the mass balance approach all impurities are quantified as a mass fraction and subtracted from 100%.

Stability: This material has demonstrated stability over a minimum period of three years. The measurement uncertainty at the 95% confidence interval includes a stability component which has been estimated from annual stability trials. The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by HPLC with UV detection on ten randomly selected 1-2 mg sub samples of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Caution: Treat as a hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,
Team Leader,
Chemical Reference Materials, NMI.
3 August 2021

This report supersedes any issued prior to 3 August 2021.

NATA logo notice: Accredited for compliance with ISO 17034. Accreditation No. 198 / Corporate Site No. 14214. The results of the tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

Characterisation Report:

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The certified purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and ¹H NMR spectroscopy. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

I_{ORG} = Organic impurities of related structure, I_{VOL} = volatile impurities, I_{NVR} = non-volatile residue.

Supporting evidence is provided by qualitative headspace GC-MS analysis of occluded solvents and elemental microanalysis.

GC-FID:

Instrument:	Agilent 6890N
Column:	HP-1, 30 m × 0.32 mm I.D. × 0.25 μm
Program:	200 °C (1 min), 5 °C/min to 240 °C, 30 °C/min to 300 °C (3 min)
Injector:	250 °C
Detector Temp:	320 °C
Carrier:	Helium
Split ratio:	20/1

Relative mass fraction of the main component:

Initial analysis:	Mean = 98.9%, s = 0.04% (5 sub samples in duplicate, November 2013)
Re-analysis:	Mean = 98.7%, s = 0.03% (5 sub samples in duplicate, October 2014)

HPLC:

Instrument:	Shimadzu Binary pump LC-20AB, SIL-20 A HT autosampler
Column:	Alltech Alltima C-18, 5 μm (4.6 mm x 150 mm)
Column oven:	40 °C
Mobile Phase:	Methanol/MilliQ water (50:50) The aqueous phase contained 0.05% v/v formic acid
Flow rate:	0.6 mL/min
Detector:	Shimadzu SPD-M20A Photodiode Array Detector operating at 254 nm

Relative mass fraction of the main component:

Initial analysis:	Mean = 99.4%, s = 0.04% (10 sub samples in duplicate, December 2011)
Re-analysis:	Mean = 99.3%, s = 0.06% (5 sub samples in duplicate, December 2012)

Karl Fischer analysis:

- Moisture content 0.3% mass fraction (November 2011)
- Moisture content 0.2% mass fraction (November 2012)
- Moisture content < 0.2% mass fraction (November 2013)
- Moisture content 0.4% mass fraction (October 2014)

Thermogravimetric analysis:

- Non volatile residue < 0.2% mass fraction (November 2011)
- The volatile content (e.g. organic solvents and/or water) could not be determined because of the inherent volatility of the material

Spectroscopic and other characterisation data

GC-MS:	Instrument:	Agilent 6890/5973
	Column:	TG-1MS, 30 m x 0.25 mm I.D. x 0.25 µm
	Program:	90 °C (1 min), 10 °C/min to 180 °C (7 min), 30 °C/min to 300 °C (3 min)
	Injector:	250 °C
	Transfer line temp:	280 °C
	Carrier:	Helium, 1.0 mL/min
	Split ratio:	20/1
	The retention time of the free base is reported along with the major peaks in the mass spectrum. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.	
	Free base (20.8 min):	155 (5), 127 (20), 126 (100), 124 (3), 55 (3) <i>m/z</i>
LC/ESI -MS:	Instrument:	Micromass Quatro LC Micro
	Operation:	Positive ion mode, direct infusion at 10 µL/min
	Ionisation:	ESI spray voltage at 3.5 kV positive ion
	EM voltage:	650 V
	Cone voltage:	40 V
	Peak:	282.4 (M+H ⁺) <i>m/z</i>
HS-GC-MS:	Instrument:	Agilent 6890/5973/G1888
	Column:	DB-624, 30 m x 0.25 mm I.D. x 1.4 µm
	Program:	50 °C (5 min), 7 °C/min to 120 °C, 15 °C/min to 220 °C (8.3 min)
	Injector:	150 °C
	Transfer line temp:	280 °C
	Carrier:	Helium, 1.2 mL/min
	Split ratio:	50/1
	Solvents detected:	Diethyl ether, acetone, ethyl acetate, n-butanol
TLC:	Conditions:	Kieselgel 60F ₂₅₄ . TBME/diethyl ether/diethyl amine (18:18:1) Single spot observed, R _f = 0.7. Visualisation with UV at 254 nm
IR:	Instrument:	Biorad FTS3000MX FT-IR
	Range:	4000-400 cm ⁻¹ , KBr powder
	Key peaks:	3042, 2959, 2928, 2871, 2562, 2467, 2362, 1668, 1626, 1449, 1355, 1285, 1120, 818, 590 cm ⁻¹
¹ H NMR:	Instrument:	Bruker Avance-III-400
	Field strength:	400 MHz
	Solvent:	DMSO- <i>d</i> ₆ (2.50 ppm)
	Spectral data:	δ 0.77 (3H, t, <i>J</i> = 7.2 Hz), 1.10 (1H, m), 1.27 (1H, m), 1.93-2.08 (6H, m), 3.14 (1H, m), 3.28 (1H, m), 3.52 (1H, m), 3.67 (1H, m), 5.74 (1H, m), 7.69 (1H, m), 7.75 (1H, m), 8.04 (1H, dd, <i>J</i> = 1.8, 8.6 Hz), 8.06 (1H, d, <i>J</i> = 8.0 Hz), 8.12 (1H, d, <i>J</i> = 8.8 Hz), 8.16 (1H, d, <i>J</i> = 8.0 Hz), 8.92 (1H, s), 10.75 (1H, s) ppm n-Butanol estimated at 0.2% mass fraction was observed in the ¹ H NMR
¹³ C NMR:	Instrument:	Bruker Avance-III-400
	Field strength:	100 MHz
	Solvent:	DMSO- <i>d</i> ₆ (39.5 ppm)
	Spectral data:	δ 13.7, 17.5, 22.9, 31.9, 52.0, 53.6, 67.2, 123.4, 127.4, 127.8, 129.0, 129.7, 129.8, 131.7, 131.9, 132.0, 135.7, 196.6 ppm
Melting point:	227-232 °C (dec)	
Microanalysis:	Found:	C = 71.3%; H = 7.6%; N = 4.3% (November, 2011)
	Calculated:	C = 71.2%; H = 7.3%; N = 4.6% (Calculated for C ₁₉ H ₂₃ NO.HCl)