

Australian Government

Department of Industry, Science and Resources National Measurement Institute




CERTIFIED REFERENCE MATERIAL CERTIFICATE OF ANALYSIS

NMIA S044: 17α-Methyl-4-chloroandrost-4-ene-3α,17β-diol

Report ID: S044.2023.02 (Ampouled 171005)

Chemical Formula: C₂₀H₃₁ClO₂

Molecular Weight: 338.9 g/mol



Certified value

Batch No.	CAS No.	Mass per ampoule
17-S-04	2389041-69-2	972 ± 20 μg

The uncertainty has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit (k = 2).

IUPAC name: (3α,17β)-4-Chloro-17-methylandrost-4-ene-3,17-diol.

Expiration of certification: The property values are valid till 24 January 2028, i.e. five years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

Description: The compound is supplied as a dried aliquot in a sealed ampoule and is intended for a single use to prepare a standard solution containing S044. Material was sourced from an external supplier, and certified for identity and purity by NMIA.

Intended use: This certified reference material is suitable for use as a primary calibrator.

Instructions for use: Open the ampoule and carefully rinse the interior at least three times with a suitable organic solvent (e.g. chloroforml). This will transfer 972 \pm 20 μ g of anhydrous 17 α -methyl-4-chloroandrost-4-ene-3 α ,17 β -diol.


Recommended storage: When not in use, this material should be stored at or below 4 °C in a closed container in a dry, dark area.

Metrological traceability: The certified purity value is traceable to the SI unit for mass (kg) through Australian national standards via balance calibration. In the mass balance approach all impurities are quantified as a mass fraction and subtracted from 100%.

Stability: This material has demonstrated stability over a minimum period of three years. The measurement uncertainty at the 95% confidence interval includes a stability component which has been estimated from annual stability trials. The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by GC-FID on seven randomly selected ampoules of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Safety: Treat as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.



Dr Stephen R. Davies, Team Leader, Chemical Reference Materials, NMI. 19 May 2025

This report supersedes any issued prior to 19 May 2025.

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

Characterisation Report:

GC-FID:	Instrument:	Varian CP-3800 or Agilent 7890
	Column:	HP-1 or HP-5, 30 m \times 0.32 mm l.D. \times 0.25 μm
	Program:	180 °C (1 min), 30 °C/min to 240 °C (10 min), 30 °C/min to 280 °C (8 min)
	Injector:	250 °C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative mass fraction of the main component as the <i>Bis</i> -TMS derivative:	
	Initial analysis:	Mean = 99.4%, s = 0.01% (7 ampoules in duplicate, November 2017)
	Re-analysis:	Mean = 99.3%, s = 0.05% (5 ampoules in duplicate, September 2018)
	Re-analysis:	Mean = 99.2%, s = 0.01% (5 ampoules in duplicate, August 2019)
	Re-analysis:	Mean = 99.3%, s = 0.01% (5 ampoules in duplicate, August 2020)
	Re-analysis:	Mean = 99.4%, s = 0.01% (5 ampoules in duplicate, January 2023)

The following analytical data was obtained on the bulk material subsequently used in the preparation of the ampoules.

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The certified purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and ¹H NMR spectroscopy. The purity value is calculated as per Equation 1

Purity = (100 % - I_{ORG}) x (100 % - I_{VOL} - I_{NVR})

Equation 1

IORG = Organic impurities of related structure, IVOL = volatile impurities, INVR = non-volatile residue.

Supporting evidence is provided by qualitative headspace GC-MS analysis of occluded solvents and elemental microanalysis.

GC-FID:	Instrument:	Varian CP-3800
	Column:	HP-1MS, 30 m × 0.32 mm l.D. × 0.25 μm
	Program:	180 °C (1 min), 30 °C/min to 250 °C (10 min), 30 °C/min to 280 °C (8 min)
	Injector:	250 °C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative mass fraction of the main component as the <i>bis</i> -TMS derivative	
	Initial analysis:	Mean = 99.3%, s = 0.01% (7 sub samples in duplicate, August 2017)
Karl Fischer analysis:		Moisture content 0.1% mass fraction (August 2017)
Thermogravimetric analysis:		Volatiles content 2.6% and non-volatile residue < 0.2% mass fraction (August 2017)

Spectroscopic and other characterisation data

opectioscopi			
GC-MS:	Parent compound: Instrument: Column:	Agilent 6890/5973 HP-1MS, 30 m x 0.25 mm I.D. x 0.25 μm	
	Program:	180 °C (1 min), 15 °C/min to 240 °C (12 min), 30 °C/min 300 °C (3 min)	
	Injector:	250 °C Split ratio: 20/1	
	Transfer line temp:	280 °C Carrier: Helium, 1.0 mL/min	
	Scan range:	50-550 <i>m/z</i>	
	<i>Bis</i> -TMS derivative: Instrument:	Shimadzu GCMS-QP2010 Ultra	
	Column:	HP Ultra 1, 25 m x 0.22 mm I.D. x 0.11 μm	
	Program:	115 °C (0.8 min), 90 °C/min to 180 °C, 5 °C/min to 190 °C, 3 °C/min to 230 °C, 10 °C/min to 265 °C, 30 °C/min to 320 °C (4 min)	
	Injector:	250 °C Split ratio: 20/1	
	Transfer line temp: Scan range:	250 °C Carrier: Helium 50-550 <i>m/z</i>	
	The retention times of t	es of the parent compound and <i>bis</i> -TMS derivative are reported with the major peaks in the e latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the	
	Parent (14.5 min):	304 (29), 302 (83), 289 (28), 287 (76), 267 (38), 251 (21), 227 (10), 191 (24), 179 (24), 177 (25), 161 (37), 143 (25), 141 (31), 133 (36), 131 (34), 128 (32), 121 (30), 119 (36), 115 (38), 107 (52), 105 (100), 91 (88), 79 (45), 77 (44) <i>m/z</i>	
	<i>Bis</i> -TMS (18.8 min):	482 (M⁺, < 1), 467 (2), 447 (12), 357 (47), 341 (3), 267 (5),143 (100), 130 (12), 75 (20), 73 (46) <i>m/z</i>	
ESI-MS:	Instrument:	Micromass Quatro LC Micro	
	Operation:	Negative ion mode, direct infusion at 20 μ L/min	
	lonisation:	ESI spray voltage at 3.5 kV positive ion	
	EM voltage:	500 V	
	Cone voltage: Peak:	40 ∨ 337.2 (M-Cl ³⁵ -H⁺), 339.1(M-Cl ³⁷ -H⁺) <i>m/z</i>	
HS-GC-MS:	Instrument:	Agilent 6890/5973/G1888	
	Column:	DB-624, 30 m x 0.25 mm l.D. x 1.4 μm	
	Program:	50 °C (5 min), 7 °C/min to 120 °C, 15 °C/min to 220 °C (8.3 min) 150 °C	
	Injector: Transfer line temp:	280 °C Split ratio: 50/1	
	Carrier:	Helium, 1.2 mL/min	
	Solvents detected:	No solvents detected	
TLC:	Conditions:	Kieselgel 60F ₂₅₄ . Hexane/ethyl acetate (1:1) Single spot observed, R_f = 0.7. Visualisation with vanillin.	
IR:	Instrument:	Bruker Alpha Platinum ATR	
	Range:	4000-400 cm ⁻¹ , neat	
	Peaks:	3254, 2971, 2932, 2874, 2847, 1450, 1373, 1312, 1290, 1150, 1084, 943, 819, 753, 705, 649, 551 cm ⁻¹	
¹ H NMR:	Instrument:	Bruker Avance III-500	
	Field strength:	500 MHz	
	Solvent:	CDCl₃ (7.26ppm)	
	Spectral data:	δ 0.81-0.94 (2H, m), 0.88 (3H, s), 1.05 (3H, s), 1.15 (1H, m), 1.19 (3H, s), 1.24-1.33 (2H,	
		m), 1.38 (1H, ddd, J = 3.8, 12.9, 26.9 Hz), 1.49-1.65 (6H, m), 1.71-1.88 (5H, m), 1.93 (1H, dddd, $J = 2.1, 4.7, 14.2, $	
		(1H, dddd, <i>J</i> = 2.1, 4.7, 14.3, 14.3 Hz), 2.93 (1H, dddd, <i>J</i> = 2.8, 4.2, 14.3 Hz), 4.14 (1H, m) ppm	
		Methanol estimated at 2.4% mass fraction was observed in the ¹ H NMR.	
¹³ C NMR:	Instrument:	Bruker Avance III-500 Field strength: 126 MHz	
C INIVITY.	Solvent:	CDCl ₃ (77.19 ppm)	
	Spectral data:	δ 14.1, 18.3, 21.4, 23.4, 26.0, 26.8, 27.6, 31.2, 31.7, 31.8, 36.5, 39.1, 40.7, 45.5, 50.4, 54.5, 70.0, 81.8, 127.2, 143.9 ppm	
Melting point:		218-219 °C	
Microanalysis:	Found:	C = 69.9%; H = 9.4%; Cl% = 10.2% (August 2017)	
-	Calculated:	C = 70.1%; H = 9.3%; Cl% = 10.2% (Calculated for $C_{20}H_{31}ClO_2$ with 2.4% methanol)	