

National Measurement Institute






# CERTIFIED REFERENCE MATERIAL CERTIFICATE OF ANALYSIS

# NMIA D979: (±)-3-Fluoroamphetamine hydrochloride

Report ID: D979.2025.01 (Bottled 160811)

Chemical Formula: C9H12FN.HCI

Molecular Weight: 189.7 g/mol (HCI), 153.2 g/mol (base)



## **Certified value**

Batch No.	CAS No.	Purity (mass fraction)
12-D-05	1716-59-2 (HCI) 1626-71-7 (base)	99.4 ± 0.9%

The uncertainty has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit (k = 2).

**IUPAC name:** 1-(3-Fluorophenyl)-2-propanamine hydrochloride.

**Expiration of certification:** The property values are valid till 21 August 2035, ten years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

**Description:** White powder prepared by synthesis and certified for identity and purity by NMI Australia. Packaged in amber glass bottles with a septum and crimped aluminium cap or screw top cap.

Intended use: This certified reference material is suitable for use as a primary calibrator.

**Instructions for use:** Equilibrate the bottled material to room temperature before opening.

Recommended storage: When not in use this material should be stored at or below 25 °C in a closed container in a dry, dark area.


**Metrological traceability:** The certified purity value is traceable to the SI unit for mass (kg) through Australian national standards via balance calibration. In the mass balance approach all impurities are quantified as a mass fraction and subtracted from 100%. Quantitative NMR provides an independent direct measure of the mass fraction of the analyte of interest, calibrated with an internal standard certified for purity (mass fraction).

**Stability:** This material has demonstrated stability over a minimum period of ten years. The measurement uncertainty at the 95% confidence interval includes a stability component which has been estimated from annual stability trials.

The long-term stability of the compound in solution has not been examined.

**Homogeneity assessment:** The homogeneity of the material was assessed using purity assay by GC-FID on ten randomly selected 1-2 mg sub samples of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

**Safety:** Treat as a hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.



Dr Stephen R. Davies, Team Leader, Chemical Reference Materials, NMI. 21 October 2025

This report supersedes any issued prior to 21 October 2025.

NATA Accreditation No. 198 / Corporate Site No. 14214.

CIPM MRA notice: This certificate is consistent with the capabilities that are included in Appendix C of the CIPM MRA drawn up by the CIPM. Under the CIPM MRA, all participating institutes recognize the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C. The "CIPM MRA Logo" and this statement attest only to the measurement(s) applied for determining the certified values on the certificate (for details see <a href="http://www.bipm.org">http://www.bipm.org</a>).

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

#### **Characterisation Report:**

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The certified purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and <sup>1</sup>H NMR spectroscopy. The purity value is calculated as per Equation 1.

Purity =  $(100 \% - I_{ORG}) x (100 \% - I_{VOL} - I_{NVR})$ 

Equation 1

I<sub>ORG</sub> = Organic impurities of related structure, I<sub>VOL</sub> = volatile impurities, I<sub>NVR</sub> = non-volatile residue.

Supporting evidence is provided by qualitative headspace GC-MS analysis of occluded solvents, elemental microanalysis and quantitative nuclear magnetic resonance (qNMR). The purity estimated by qNMR was obtained using the three-proton doublet at 1.9 ppm against a certified internal standard of maleic acid.

GC-FID: Instrument:

Varian CP-3800 or Agilent 7890

Column: Program: VF-1MS or HP-1, 30 m  $\times$  0.32 mm I.D.  $\times$  0.25  $\mu$ m 60 °C (1 min), 7 °C/min to 100 °C (5 min), 25 °C/min to 300 °C (3 min)

Injector: 250 °C Detector Temp: 320 °C

Carrier: Helium
Split ratio: 20/1

Relative mass fraction of the main component as the free base:

Initial analysis: Mean = 99.7%, s = 0.04% (10 sub samples in duplicate, August 2012) Re-analysis: Mean = 99.5%, s = 0.01% (5 sub samples in duplicate, April 2021) Re-analysis: Mean = 99.5%, s = 0.05% (5 sub samples in duplicate, August 2025)

GC-FID:

Instrument: Varian CP-3800

Column: HP-5, 30 m  $\times$  0.32 mm I.D.  $\times$  0.25  $\mu$ m

Program: 60 °C (1 min), 7 °C/min to 100 °C (5 min), 25 °C/min to 300 °C (3 min)

Injector: 250 °C

Detector Temp: 320 °C

Carrier: Helium

Split ratio: 20/1

Relative mass fraction of the main component as the free base:

Initial analysis: Mean = 99.7%, s = 0.03% (10 sub samples in duplicate, August 2012) Re-analysis: Mean = 99.6%, s = 0.01% (7 sub samples in duplicate, August 2013) Re-analysis: Mean = 99.6%, s = 0.04% (7 sub samples in duplicate, July 2014) Re-analysis: Mean = 99.6%, s = 0.01% (5 sub samples in duplicate, July 2015) Re-analysis: Mean = 99.5%, s = 0.03% (5 sub samples in duplicate, July 2018)

Karl Fischer analysis:

Moisture content ≤ 0.1% mass fraction (August 2012, July 2013, 2014, 2015, 2018,

March 2021, and August 2025).

Thermogravimetric analysis:

Non volatile residue < 0.2% mass fraction (August 2012). The volatile content (e.g. organic solvents and/or water) could not be determined because of the inherent

volatility of the material.

QNMR:

Instrument: Bruker Avance-III-400

Field strength: 400 MHz Solvent: D<sub>2</sub>O (4.79 ppm)

Internal standard: Maleic acid (98.7% mass fraction)

Initial analysis: Mean (1.3 ppm) = 99.8%, s = 0.1% (5 sub samples, September 2012)

### Spectroscopic and other characterisation data

GC-MS: Instrument: Agilent 6890/5973

Column: TG-1MS, 30 m x 0.25 mm l.D. x 0.25 μm

Program: 60 °C (1 min), 10 °C/min to 100 °C, 15 °C/min to 300 °C (3 min)

Injector: 250 °C Transfer line temp: 300 °C

Carrier: Helium, 1.0 mL/min

Split ratio: 20/1

The retention time of the free base is reported along with the major peaks in the mass spectrum. The latter are

reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.

Free base (7.1 min): 152 (1), 138 (3), 109 (18), 83 (8), 44 (100) m/z

LC/ESI -MS: Instrument: Waters Acquity UPLC/TQD

 $\begin{array}{ll} \mbox{Operation:} & \mbox{Positive ion mode, direct infusion at 5 $\mu$L/min} \\ \mbox{Ionisation:} & \mbox{ESI spray voltage at 3.0 kV positive ion} \\ \end{array}$ 

Capillary voltage: 3 kV Cone voltage: 20 V

Peak: 154.1 (M+H+) m/z

HS-GC-MS: Instrument: Agilent 6890/5973/G1888

Column: DB-624, 30 m x 0.25 mm l.D. x 1.4 μm

Program: 50 °C (5 min), 7 °C/min to 120 °C, 15 °C/min to 220 °C (8.3 min)

Injector: 150  $^{\circ}$ C Transfer line temp: 280  $^{\circ}$ C

Carrier: Helium, 1.2 mL/min

Split ratio: 50/1

Solvents detected: None detected

TLC: Conditions: Kieselgel 60F<sub>254</sub>. Methanol/Conc. NH<sub>3</sub> (200/3)

Single spot observed,  $R_f = 0.5$ . Visualisation with UV at 254 nm

IR: Instrument: Biorad FTS3000MX FT-IR

Range: 4000-400 cm<sup>-1</sup>, KBr powder

Peaks: 3083, 2887, 2690, 2596, 2499, 2048, 2019, 1616, 1592, 1484, 1452, 1392, 1247, 1208,

1141, 952, 873, 795, 755, 692, 521, 432 cm<sup>-1</sup>

<sup>1</sup>H NMR: Instrument: Bruker Avance DMX-600

Field strength: 600 MHz Solvent: D<sub>2</sub>O (4.79 ppm)

Spectral data:  $\delta$  1.32 (3H, d, J = 6.7 Hz), 2.97 (2H, d, J = 7.2 Hz), 3.66 (1H, sextet, J = 6.8 Hz), 7.07 –

7.15 (3H, m), 7.42 (1H, m) ppm

<sup>13</sup>C NMR: Instrument: Bruker Avance DMX-600

Field strength: 150 MHz Solvent: D<sub>2</sub>O

Spectral data:  $\delta$  17.5, 39.7 (d,  $J_{C-F}$  = 1.7 Hz), 48.9, 114.1 (d,  $J_{C-F}$  = 19.6 Hz), 116.1 (d,  $J_{C-F}$  = 21.1 Hz),

125.3 (d,  $J_{C-F} = 2.8 \text{ Hz}$ ), 130.6 (d,  $J_{C-F} = 8.5 \text{ Hz}$ ), 138.5 (d,  $J_{C-F} = 7.3 \text{ Hz}$ ), 162.8 (d,  $J_{C-F} = 7.3 \text{ Hz}$ ), 162.8 (d,  $J_{C-F} = 7.3 \text{ Hz}$ )

242.9 Hz) ppm

<sup>19</sup>F NMR: Instrument: Bruker Avance-400

Field strength: 376 MHz Solvent: D<sub>2</sub>O

Spectral data:  $\delta$  -113.52 ppm

Melting point: 145-148 °C

Microanalysis: Found: C = 57.1%; H = 7.0%; N = 7.3%; Cl = 18.7%; F = 9.8% (August, 2012)

Calculated: C = 57.0%; H = 6.9%; N = 7.4%; CI = 18.7%; F = 10.0% (Calculated for  $C_9H_{12}IFN.HCI$ )

Page 4 of 4