




## CERTIFIED REFERENCE MATERIAL CERTIFICATE OF ANALYSIS

### NMIA D560: 17 $\alpha$ -Methyl-5 $\alpha$ -androstane-3 $\alpha$ , 17 $\beta$ -diol

Report ID: D560.2026.01 (Ampouled 210513)

Chemical Formula: C<sub>20</sub>H<sub>34</sub>O<sub>2</sub>

Molecular Weight: 306.5 g/mol



#### Certified value

Batch No.	CAS No.	Mass per ampoule
<b>99-000005</b>	<b>641-82-7</b>	<b>1002 ± 8 µg</b>

The uncertainty has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit (k = 2).

**IUPAC name:** (3 $\alpha$ ,5 $\alpha$ ,17 $\beta$ )-17-Methandrostan-3,17-diol.

**Expiration of certification:** The property values are valid till 16 January 2031, five years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to ampoules that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

**Description:** The compound is supplied as a dried aliquot in a sealed ampoule under an atmosphere of argon. This CRM is intended for a single use to prepare a standard solution containing 17 $\alpha$ -methyl-5 $\alpha$ -androstane-3 $\alpha$ , 17 $\beta$ -diol. This material was sourced from an external supplier, and certified for identity and purity by NMI Australia.

**Intended use:** This certified reference material is suitable for use as a primary calibrator.

**Instructions for use:** Open the ampoule and carefully rinse the interior at least three times with a suitable organic solvent (e.g. methanol). This will transfer 1002 ± 8 µg of anhydrous 17 $\alpha$ -methyl-5 $\alpha$ -androstane-3 $\alpha$ , 17 $\beta$ -diol. The mass of analyte in each ampoule is calculated from the assigned purity of the bulk and the concentration of bulk material in a stock solution used to prepare the ampoules.

**Recommended storage:** When not in use, this material should be stored at or below 4 °C in a closed container in a dry, dark area.

**Metrological traceability:** The certified purity value is traceable to the SI unit for mass (kg) through Australian national standards via balance calibration. In the mass balance all impurities are quantified as a mass fraction and subtracted from 100%.

**Stability:** This material has demonstrated stability over a minimum period of three years. The measurement uncertainty at the 95% confidence interval includes a stability component which has been estimated from annual stability trials. The long-term stability of the compound in solution has not been examined.

**Homogeneity assessment:** The homogeneity of the material was assessed using purity assay by GC-FID on seven randomly selected ampoules of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

**Safety:** Treat as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

*S. R. Davies*

Dr Stephen R. Davies,  
Team Leader,  
Chemical Reference Materials, NMI.  
28 January 2026

This report supersedes any issued prior to 28 January 2026.

NATA Accreditation No. 198 / Corporate Site No. 14214.

**Legal notice:** Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

---

### Characterisation Report:

GC-FID:	Instrument:	Agilent 7890 or Agilent 8890
	Column:	HP-1, 30 m $\times$ 0.32 mm I.D. $\times$ 0.25 $\mu$ m
	Program:	180 °C (1 min), 20 °C/min to 260 °C (5 min), 10 °C/min to 300 °C (3 min)
	Injector:	250 °C or 200°C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative mass fraction of the main component:	
	Initial analysis:	Mean = 99.8%, s = 0.01% (7 ampoules in duplicate, May 2021)
	Re-analysis:	Mean = 99.8%, s = 0.01% (5 ampoules in duplicate, April 2022)
	Re-analysis:	Mean = 99.8%, s = 0.01% (5 ampoules in duplicate, April 2024)
	Re-analysis:	Mean = 99.8%, s = 0.04% (5 ampoules in duplicate, January 2026)

### The following analytical data was obtained on the bulk material subsequently used in the preparation of the ampoules.

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The certified purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and  $^1$ H NMR spectroscopy. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

$I_{\text{ORG}}$  = Organic impurities of related structure,  $I_{\text{VOL}}$  = volatile impurities,  $I_{\text{NVR}}$  = non-volatile residue.

Supporting evidence is provided by elemental microanalysis.

GC-FID:	Instrument:	HP 5890
	Column:	ZB-1, 30 m $\times$ 0.32 mm I.D. $\times$ 0.25 $\mu$ m
	Program:	180 °C (1 min), 10 °C/min to 220 °C, 20 °C/min to 300 °C (6 min)
	Injector:	250 °C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative mass fraction of the main component:	
	Initial analysis:	Mean = 99.8%, s = 0.004% (7 sub samples in duplicate, December 2004)
GC-FID:	Instrument:	Agilent 6890 and 7890
	Column:	HP-1, 30 m $\times$ 0.32 mm I.D. $\times$ 0.25 $\mu$ m
	Program:	180 °C (1 min), 20 °C/min to 260 °C (5min), 10 °C/min to 300 °C (3 min)
	Injector:	250 °C
	Detector Temp:	320 °C
	Carrier:	Helium
	Split ratio:	20/1
	Relative mass fraction of the main component:	
	Initial analysis:	Mean = 99.9%, s = 0.003% (5 sub samples in duplicate, March 2008)
	Re-analysis:	Mean = 99.9%, s = 0.014% (7 sub samples in duplicate, November 2010)
	Re-analysis:	Mean = 99.8%, s = 0.008% (7 sub samples in duplicate, May 2021)

Karl Fischer analysis:

Moisture content  $\leq$  0.1% mass fraction (May 2021)

Thermogravimetric analysis:

Volatiles content  $<$  0.1% and non-volatile residue  $<$  0.2% mass fraction (June 1999 and May 2006)

## Spectroscopic and other characterisation data

GC-MS:	Parent compound: Instrument: Agilent 6890/5973 Column: HP Ultra 2, 17 m x 0.20 mm I.D. x 0.10 $\mu$ m Program: 180 °C (1 min), 10 °C/min to 220 °C, 20 °C/min to 300 °C (3 min) Injector: 280 °C Split inj: (20/1) Transfer line temp: 300 °C Carrier: Helium, 1.0 mL/min Scan range: 50-550 m/z
	<i>Bis</i> -TMS derivative: Instrument: Agilent 6890/5973 Column: HP Ultra 1, 17 m x 0.22 mm I.D. x 0.11 $\mu$ m Program: 170 °C (0.5 min), 3 °C/min to 234 °C, 10 °C/min to 265 °C (3 min) Injector: 280 °C Split inj: (20/1) Transfer line temp: 300 °C Carrier: Helium Scan range: 50-550 m/z
The retention times of the parent compound and <i>bis</i> -TMS derivative are reported along with the major peaks in the mass spectra. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.	
Parent (5.85 min): 306 (M <sup>+</sup> , 32), 291 (95), 230 (100), 215 (88), 165 (81), 107 (78) m/z <i>Bis</i> -TMS (10.50 min): 450 (M <sup>+</sup> , 2), 435 (17), 255 (7), 143 (100), 73 (34) m/z	
The <i>bis</i> -silylated derivative of the synthetic material co-elutes with a derivatised comparison sample of 17 $\alpha$ -methyl-5 $\alpha$ -androstane-3 $\alpha$ ,17 $\beta$ -diol and the two materials produce matching mass spectra.	
TLC:	Conditions: Kieselgel 60F <sub>254</sub> . Ethyl acetate/dichloromethane (3:7) Single spot observed, R <sub>f</sub> = 0.4.
IR:	Instrument: FT-IR, Biorad WIN FTS40 Range: 4000 - 400 cm <sup>-1</sup> , KBr powder Peaks: 3357, 1448, 1370, 1266, 1170, 1008 cm <sup>-1</sup>
<sup>1</sup> H NMR:	Instrument: Bruker Avance III-400 Field strength: 400 MHz Solvent: CDCl <sub>3</sub> (7.26 ppm) Key spectral data: $\delta$ 0.79 (3H, s), 0.84 (3H, s), 1.20 (3H, s), 4.03 (1H, d, J = 2.1 Hz) ppm
<sup>13</sup> C NMR:	Instrument: Bruker Avance III-400 Field strength: 100 MHz Solvent: CDCl <sub>3</sub> (77.16 ppm) Spectral data: $\delta$ 11.4, 14.1, 20.5, 23.4, 26.0, 28.6, 29.2, 31.8, 31.9, 32.4, 36.0, 36.3, 36.5, 39.1, 39.4, 45.7, 50.9, 54.5, 66.7, 81.9 ppm
Melting point:	183-186 °C
Microanalysis:	Found: C = 78.4%; H = 11.4% (June 2007) Calculated: C = 78.4%; H = 11.2% (Calculated for C <sub>20</sub> H <sub>34</sub> O <sub>2</sub> )