




CERTIFIED REFERENCE MATERIAL CERTIFICATE OF ANALYSIS

NMIA D283b: Mefenorex hydrochloride

Report ID: D238b.2019.01

Chemical Formula: C₁₂H₁₈ClN.HCl

Molecular Weight: 248.2 g/mol (HCl)



Certified value

Batch No.	CAS No.	Purity (mass fraction)
05-D-12	5586-87-8	99.3 ± 0.7 %

The uncertainty has been calculated according to ISO Guide 35 and is stated at the 95% confidence limit ($k = 2$).

IUPAC: 3-Chloro-N-(1-phenyl-2-propanyl)-1-propanamine hydrochloride

Expiration of certification: The property values are valid till 29 July 2024, i.e. five years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to sample bottles that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

Description: White solid prepared by synthesis, and certified for identity and purity by NMIA. Packaged in amber glass bottles with a screw top cap.

Intended use: This certified reference material is suitable for use as a primary calibrator.

Instructions for use: Equilibrate the bottled material to room temperature before opening.

Recommended storage: When not in use, this material should be stored at or below 25 °C in a closed container in a dry, dark area.

Metrological traceability: The certified purity value is traceable to the SI unit for mass (kg) through Australian national standards via balance calibration. In the mass balance approach all impurities are quantified as a mass fraction and subtracted from 100%.

Stability: This material has demonstrated stability over a minimum period of five years. The measurement uncertainty at the 95% coverage interval includes a stability component which has been estimated from annual stability trials. The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by GC-FID on five randomly selected 1-2 mg sub samples of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Safety: Treat as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,
Team Leader,
Chemical Reference Materials, NMI.
9 August 2019

This report supersedes any issued prior to 9 August 2019

NATA logo notice: Accredited for compliance with ISO 17034. Accreditation No. 198 / Corporate Site No. 20844. The results of the tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

Legal notice: Neither NMIA as a representative of the Commonwealth of Australia, nor any person acting on NMIA's behalf, assumes any liability with respect to the use of, or for damages resulting from the use of, this reference material or the information contained in this document.

Characterisation Report:

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The certified purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID and Karl Fischer analysis. The purity value is calculated as per Equation 1.

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

I_{ORG} = Organic impurities of related structure, I_{VOL} = volatile impurities, I_{NVR} = non-volatile residue

Supporting evidence is provided by qualitative elemental microanalysis.

GC-FID: Instrument: Agilent 6890N or 7890
 Column: HP-1 Capillary, 30 m × 0.32 mm I.D. × 0.25 μm
 Program: 60 °C (1 min), 20 °C/min to 300 °C (3 min)
 Injector: 150 °C
 Detector Temp: 300 °C
 Carrier: Helium
 Split ratio: 20/1
 Relative mass fraction of the main component:
 Initial analysis: Mean = 99.2%, s = 0.06% (3 sub samples in duplicate, September 2005)
 Re-analysis: Mean = 99.3%, s = 0.03% (5 sub samples in duplicate, October 2008)
 Re-analysis: Mean = 99.3%, s = 0.03% (5 sub samples in duplicate, October 2011)

GC-FID: Instrument: Agilent 6890N, 7890 or Varian CP-3800
 Column: HP-1 or VF-1MS Capillary, 30 m × 0.32 mm I.D. × 0.25 μm
 Program: 60 °C (1 min), 20 °C/min to 300 °C (3 min)
 Injector: 180 °C
 Detector Temp: 320 °C
 Carrier: Helium
 Split ratio: 20/1
 Relative mass fraction of the main component:
 Re-analysis: Mean = 99.3%, s = 0.01% (5 sub samples in duplicate, August 2014)
 Re-analysis: Mean = 99.5%, s = 0.01% (5 sub samples in duplicate, July 2019)

Thermogravimetric analysis: Volatile content was not determined due to decomposition of the material between 110-120 °C

Karl Fischer analysis: Moisture content < 0.1% mass fraction (August 2006)
 Moisture content 0.4% mass fraction (September 2007)
 Moisture content 0.4% mass fraction (October 2008)
 Moisture content < 0.1% mass fraction (October 2011)
 Moisture content < 0.1% mass fraction (August 2014)
 Moisture content < 0.1% mass fraction (July 2019)

Spectroscopic and other characterisation data

GC-MS:	Instrument: HP5890/5971A Column: ZB-5, 30 m × 0.25 mm I.D. × 0.20 μm Program: 100 °C (1 min), 10 °C/min to 250 °C, (5 min) Injector: 220 °C Transfer line temp: 280 °C Carrier: Helium, 1.0 mL/min Split ratio: 20/1
	The retention time of the free base is reported along with the major peaks in the mass spectrum. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak 9.70 min: 196 (M ⁺ , 1), 120 (100), 91 (22), 84 (4), 76 (4), 65 (6), 56 (5) <i>m/z</i>
TLC:	Conditions: Kieselgel 60F ₂₅₄ . Methanol/Conc. aqueous NH ₃ (200/3) Single spot observed, R _f = 0.63. Visualisation with ninhydrin dip
IR:	Instrument: Biorad FTS300MX FT-IR Range: 4000-400 cm ⁻¹ , KBr pellet Peaks: 2962, 2806, 2499, 2468, 2422, 1588, 1499, 1454, 1389, 1316, 738, 698 cm ⁻¹
¹ H NMR:	Instrument: Bruker DPX2-300 Field strength: 300 MHz Solvent: MeOH- <i>d</i> ₄ (3.31 ppm) Spectral data: δ 1.25 (3H, d, <i>J</i> = 6.5 Hz), 2.22 (2H, m), 2.75 (1H, dd, <i>J</i> = 10.1, 13.2 Hz), 3.22-3.32 (3H, m), 3.54 (1H, m), 3.73 (2H, t, <i>J</i> = 6.2 Hz), 7.25-7.40 (5H, m) ppm
¹³ C NMR:	Instrument: Bruker DPX2-300 Field strength: 75 MHz Solvent: MeOH- <i>d</i> ₄ (49.0 ppm) Spectral data: δ 15.9, 30.4, 40.2, 42.3, 43.9, 57.2, 128.4, 130.0, 130.4, 137.2 ppm
Melting point:	124-128 °C
Microanalysis:	Found: C = 58.3 %; H = 7.9 %; N = 5.6% (September 2005) Calculated: C = 58.1 %; H = 7.7 %; N = 5.6% (Calculated for C ₁₂ H ₁₈ ClN.HCl)