




REFERENCE MATERIAL PRODUCT INFORMATION SHEET

NMIA D624: 7 β ,17 α -Dimethyl-5 β -androstane-3 α ,17 β -diol

Report ID: D624.2020.01 (Ampouled 091006)

Chemical Formula: C₂₁H₃₆O₂

Molecular Weight: 320.5 g/mol



Property value

Batch No.	CAS No.	Mass per ampoule
99-S-31	153546-23-7	987 μ g

IUPAC name: (3 α , 5 β , 7 β , 17 β)-7,17-Dimethylandrostane-3,17-diol

Expiration of certification: The property values are valid till 1 July 2030, i.e. ten years from the date of re-certification provided the **unopened** material is handled and stored in accordance with the recommendations below. The material as issued in the unopened container and stored as recommended below should be suitable for use beyond this date, subject to confirmation of batch stability from the issuing body. The expiry date/shelf life does not apply to ampoules that have been opened. In such cases it is recommended that the end-user conduct their own in-house stability trials.

Description: The compound is supplied as a dried aliquot in a sealed ampoule under an atmosphere of argon. The reference material is intended for a single use to prepare a standard solution containing D624. Material was sourced from an external supplier, and certified for identity and purity by NMIA.

Intended use: This reference material is recommended for qualitative analysis only.

Instructions for use: Open the ampoule and carefully rinse the interior at least three times with a suitable organic solvent (e.g. chloroform). This will transfer an estimated 987 μ g of anhydrous 7 β ,17 α -dimethyl-5 β -androstane-3 α ,17 β -diol. The mass of analyte in each ampoule is calculated from the assigned purity of the bulk and the concentration of bulk material in a stock solution used to prepare the ampoules.

Recommended storage: When not in use, this material should be stored at or below 4 $^{\circ}$ C in a closed container in a dry, dark area.

Stability: The long-term stability of the compound in solution has not been examined.

Homogeneity assessment: The homogeneity of the material was assessed using purity assay by GC-FID on seven randomly selected ampoules of the material. The material was judged to be sufficiently homogeneous at this level of sampling as the variation in analysis results between samples was not significantly different at a 95% confidence level from that observed on repeat analysis of the same sample.

Safety: Treat as hazardous substance. Use appropriate work practices when handling to avoid skin or eye contact, ingestion or inhalation of dust. Refer to the provided safety data sheet.

S. R. Davies

Dr Stephen R. Davies,
Team Leader,
Chemical Reference Materials, NMI.
13 July 2020

This report supersedes any issued prior to 13 July 2020.

NATA logo notice: Accredited for compliance with ISO 17034. Accreditation No. 198 / Corporate Site No. 20844. The results of the tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

Legal notice: Terms and Conditions associated with the provision of this reference material can be found on the NMIA website.

Characterisation Report:

GC-FID: Instrument: Agilent 6890N
 Column: HP-1 Capillary, 30 m × 0.32 mm I.D. × 0.25 μm
 Program: 180 °C (1 min), 5 °C/min to 250 °C, 20 °C/min to 300 °C (3 min)
 Injector: 250 °C
 Detector Temp: 320 °C
 Carrier: Helium
 Split ratio: 20/1

Relative peak area of the main component:
 Initial analysis: Mean = 98.9%, s = 0.01% (7 ampoules in duplicate, October 2009)
 Re-analysis: Mean = 99.0%, s = 0.01% (5 ampoules in duplicate, September 2012)
 Re-analysis: Mean = 99.1%, s = 0.01% (5 ampoules in duplicate, July 2015)
 Re-analysis: Mean = 99.1%, s = 0.01% (5 ampoules in duplicate, July 2020)

The following analytical data was obtained on the bulk material subsequently used in the preparation of the ampoules.

The identity was confirmed by a range of spectroscopic techniques, NMR, IR and MS. The purity value was obtained by mass balance from a combination of traditional analytical techniques, including GC-FID, thermogravimetric analysis, Karl Fischer analysis and ¹H NMR spectroscopy. The purity value is calculated as per Equation 1

$$\text{Purity} = (100 \% - I_{\text{ORG}}) \times (100 \% - I_{\text{VOL}} - I_{\text{NVR}}) \quad \text{Equation 1}$$

I_{ORG} = Organic impurities of related structure, I_{VOL} = volatile impurities, I_{NVR} = non-volatile residue.

Supporting evidence is provided by elemental microanalysis.

GC-FID: Instrument: Agilent 6890N
 Column: HP-1 Capillary, 30 m × 0.32 mm I.D. × 0.25 μm
 Program: 180 °C (1 min), 5 °C/min to 250 °C, 20 °C/min to 300 °C (3 min)
 Injector: 250 °C
 Detector Temp: 320 °C
 Carrier: Helium
 Split ratio: 20/1

Relative peak area of the main component:
 Initial analysis: Mean = 99.5%, s = 0.03% (10 sub samples in duplicate, October 1999)
 Re-analysis: Mean = 99.3%, s = 0.02% (5 sub samples in duplicate, June 2005)
 Re-analysis: Mean = 98.9%, s = 0.02% (5 sub samples in duplicate, October 2009)

Karl Fischer analysis: Moisture content 0.43% mass fraction (June 2009)

Thermogravimetric analysis: Volatiles content and non-volatile residue < 0.2% mass fraction

Spectroscopic and other characterisation data

GC-MS: Parent compound:
Instrument: HP6890/5973
Column: HP Ultra 2, 17 m x 0.22 mm I.D. x 0.11 μm
Program: 150 °C (1 min), 15 °C/min to 300 °C (3 min)
Injector: 280 °C
Split ratio: Splitless
Transfer line temp: 300 °C
Carrier: Helium
Scan range: 50-550 *m/z*

Bis-TMS derivative:
Instrument: HP 6890/5973
Column: HP Ultra, 17 m x 0.22 mm I.D. x 0.11 μm
Program: 170 °C, 3 °C/min to 234 °C, 10 °C/min to 265 °C (2 min)
Injector: 280 °C
Split ratio: 15/1
Transfer line temp: 300 °C
Carrier: Helium
Scan range: 50-550 *m/z*

The retention times of the parent compound and *bis*-TMS derivative are reported with the major peaks in the mass spectra. The latter are reported as mass/charge ratios and (in brackets) as a percentage relative to the base peak.

Parent (7.5 min): 320 (*M*⁺, 4), 302 (46), 287 (58), 269 (72), 244 (86), 231 (100) *m/z*
Bis-TMS (11.7 min): 449 (*M*⁺-CH₃, 1), 374 (3), 284 (5), 269 (3), 242 (4), 229 (4), 143 (100) *m/z*

TLC: Conditions: Kieselgel 60F₂₅₄. Hexane/ethyl acetate (1:1)
Single spot observed, *R*_f = 0.24 (3 sub samples)

IR: Instrument: FT-IR, Biorad WIN FTS40
Range: 4000-400 cm⁻¹, KBr pellet
Peaks: 3384, 1456, 1406, 1377, 1227, 1102, 1067, 1042, 945 cm⁻¹

¹H NMR: Instrument: Bruker Advance-300
Field strength: 300 MHz
Solvent: MeOH-*d*₄ (3.31 ppm)
Key Spectral data: δ 0.86 (3H, s), 0.95 (3H, s), 0.97 (3H, d), 1.20 (3H, s), 3.56 (1H, m) ppm
Ethyl acetate at <0.1% mass fraction was observed in the ¹H NMR

¹³C NMR: Instrument: Bruker Advance-300
Field strength: 75 MHz
Solvent: MeOH-*d*₄ (49 ppm)
Spectral data: δ 15.5, 22.5, 24.5, 24.6, 26.7, 28.8, 31.5, 33.6, 33.9, 36.0, 37.4, 38.2, 39.8, 40.2, 42.6, 44.2, 46.0, 48.5, 53.0, 72.7, 81.7 ppm

Melting point: 174-176 °C

Microanalysis: Found: C = 78.5%; H = 11.4%
Calculated: C = 78.7%; H = 11.3% (Calculated for C₂₁H₃₆O₂)